

Article

## Evaluation of a Centrifuged Double Y-Shape Microfluidic Platform for Simple Continuous Cell Environment Exchange

Akihiro Hattori <sup>1</sup> and Kenji Yasuda <sup>1,2,\*</sup>

<sup>1</sup> Kanagawa Academy of Science and Technology, KSP East 310, 3-2-1 Sakado, Kawasaki 213-0012, Japan; E-Mail: hattori.bmi@gmail.com

<sup>2</sup> Department of Biomedical Information, Institute of Biomaterials and Bioengineering, Tokyo Medical and Dental University, 2-3-10 Kanda-Surugadai, Chiyoda, Tokyo 101-0062, Japan

\* Author to whom correspondence should be addressed; E-Mail: yasuda.bmi@tmd.ac.jp; Tel.: +81-3-5280-8046; Fax: +81-3-5280-8049.

Received: 7 December 2011; in revised form: 6 January 2012 / Accepted: 9 January 2012 /

Published: 13 January 2012

---

**Abstract:** We have demonstrated the efficacy of a microfluidic medium exchange method for single cells using passive centrifugal force of a rotating microfluidic-chip based platform. At the boundary of two laminar flows at the gathering area of two microfluidic pathways in a Y-shape, the cells were successfully transported from one laminar flow to the other, without mixing the two microfluidic mediums of the two laminar flows during cell transportation, within 5 s with 1 g (150 rpm) to 36.3 g (900 rpm) acceleration, with 93.5% efficiency. The results indicate that this is one of the most simple and precise tools for exchanging medium in the shortest amount of time.

**Keywords:** microfluidic chip; Y-shape platform; medium exchange; environmental control; single cell; centrifugal force; on-chip cellomics

---

### 1. Introduction

In many areas of cell-based screenings such as blood monitoring, process control in pharmaceutical fermentations, continuous real-time monitoring of biological agents is of the utmost importance. The key requirement to perform such monitoring tasks is the fast and precise exchange processing of cell environments such as buffer exchange. Recently, microfluidic platforms [1,2] for the continuous separation of biomolecules and particles have been proposed. Some of these use passive working

principles [3–5], others apply dielectrophoresis [6,7], acoustic standing waves [8–11], segmented flow [12–14], or magnetophoresis [15,16]. Applying centrifugal force is one of the most powerful candidates for the passive steady-non pulsatile-flow in microfluidic systems [17] for continuous fully automated processing such as cell arrangement [18], DNA concentration [19], biomolecular separation [20], and also a compact disc (CD) microfluidic chip based enzyme-linked immunosorbent assay (ELISA) [21], whole blood glucose analysis [22], and hybridization assay for phenylketonuria (PKU) screening [23]. Especially, extraction of plasma from whole blood exploiting microfluidic pathways with centrifugal pumping [24] and cell trapping exploiting reagent-impregnated agarose-made micro holes with centrifugal liquid pumping [25] were examined for practical applications. Yet the majority of the proposed platforms are at the proof-of-concept stage of the system integration, and none of them have examined the ability of microfluidic platforms for quick exchange of cell environments in single cell level.

Therefore, in this paper, we have evaluated the efficacy of the exchange of environments of single cells within the simple Y-shape microfluidic chip with passive centrifugal force with charge coupled device (CCD) camera image analysis.


## 2. Results and Discussion

### 2.1. System Design

The continuous exchange of a medium condition surrounding single cells has been performed by the following three processing steps: cell introduction into the Y-shape microfluidic pathway, a medium exchange step that removes the environmental medium buffer by cells' shift from one laminar flow to the other through the boundary of two laminar flows, and cell collection by another reverse-Y-shape microfluidic pathway. On our microfluidic platform the above three steps are implemented in a microfluidic structure which is contained in a monolithic chip. A rotating platform provides the required passive centrifugal force field for cell motion within the chip. For the constant centrifugal force generation, a rotating stage with a stepping motor (TAMAGAWA SEIKI, 5PHASE STEPPING MOTOR 02K-S523W-Σ) and a control unit (SIGMA KOKI, Stage controller SC-101G) was mounted in the optical microscope system (OLYMPUS IX70) [Figure 1 (a)].

The microfluidic chip comprises five parts [Figure 1 (b)]: (1) two inlets (inner one is for sample cell solution and the other outer one is for the washing buffer); (2) Y-shape microfluidic pathway (flow channels) guiding two buffers into the channel junction; (3) the channel junction where two laminar flows are brought into contact via a laminar-flow interface (boundary) in order to transfer the cells from one buffer to the other; (4) reverse Y-shape pathway to split into two laminar flows once again; and (5) two outlets (inner one is for sample cell buffer and the other outer one is for the cells with washing buffer). The 2D-layout of the microfluidic structure was fabricated in the polydimethylsiloxane (PDMS) [26,27] attached to a 1-mm-thick glass slide.

**Figure 1.** Schematic view of the system setup and microfluidic structure of microfluidic chip. (a) System setup; (b) microfluidic chip design; (c) and (d) gathering part of double Y-shape microfluidic pathway.



## 2.2. Rotation of Chip for Centrifugal Force Generation


The rotating stage was set at the position of the focal plane of the microscope, and the chip was set on the stage (4 cm from the center of rotation), providing sufficient centrifugal force for cells transportation inside the microfluidic chip. The role of centrifugal force in this system was for two purposes: to generate the driving force for injecting buffers into the microfluidic flow with the same pressure and velocity; and secondly to transport cells from one laminar flow to the other.

In order to maintain the two laminar flows at the junction area, the velocity must be synchronized precisely at this area. In this system, each of the two inlets and two outlets were designed at the same distance from the center of rotation (Figure 2). The acceleration in microfluidic chip was determined by  $r\omega^2$ , where  $r$  is the distance of the chip from the center of rotation (4 cm),  $\omega$  is angular velocity of rotator. In this case, the accelerations were 1 g for 2.5 Hz rotation, and 36.3 g for 15 Hz.


## 2.3. Cell Motion in Microfluidic Chip Under Centrifugation

Under the influence of the centrifugal force provided by the rotation of the microfluidic chip, the cells gathered on the outer wall of the microfluidic pathway. As shown in [Figure 3 (a,b)], before centrifugal force was applied, whole erythrocyte cells were dispersed in the pathway. Whereas, once the centrifugal force was applied [Figure 3 (c,d)], whole cells shifted to the outer wall of microfluidic pathway under 1 g acceleration condition with a  $100 \mu\text{m}\cdot\text{s}^{-1}$  velocity of buffer.

**Figure 2.** Schematic images explaining spatial arrangements of chip inlets/outlets arrangements and chip rotation for centrifugal force generation.



**Figure 3.** Effect of centrifugal force on erythrocytes in the microfluidic chip. (a) micrograph of erythrocyte distribution before centrifugation started; (b) spatial distribution of erythrocytes in (a); (c) micrograph after centrifugation started; and (d) spatial distribution in (c).




#### 2.4. Cell Transportation Between Two Laminar Flows

We checked two topics regarding this centrifugation platform: firstly, the evaluation of cell transportation from one pathway to the other, and, secondly, the evaluation of mixture of two laminar flows during the cell transportation process.

Figure 4 (a) shows the results of the above two subjects, *i.e.*, cell transportation and medium mixture. As shown in this figure, a HeLa cell with a  $100 \mu\text{m}\cdot\text{s}^{-1}$  flow velocity, shifted successfully from the first laminar flow to the other along the  $400 \mu\text{m}$  length of two laminar flow boundaries at the gathering area. In contrast, the medium buffers maintained their boundaries, and no mixture was observed either visibly or in fluorescence detection of collected buffer (we added fluorescent dye (Rhodamine B) into one of two buffers for quantitative evaluation of mixture).

**Figure 4.** Time course tracking of cell transportation. (a) movement of single HeLa cell from one laminar flow to the other under  $1 g$  acceleration with  $2.5 \text{ Hz}$  rotation; (b) movement under  $36.3 g$  with  $15 \text{ Hz}$  rotation.




We have also checked the cell transportation under stronger centrifugal force conditions. As shown in Figure 4 (b), we added  $36.3 g$  ( $15 \text{ Hz}$ ) acceleration into the microfluidic chip and confirmed the cells having  $200 \mu\text{m}\cdot\text{s}^{-1}$  velocity successfully shifted from one laminar flow to the other, and no fluorescent medium mixture was observed using fluorescence detector.

As this exchange was performed successfully even under  $1 g$  centrifugal force condition, that means we can exchange medium buffer continuously using gravity without using centrifugal force. As expected, when we set this microfluidic chip perpendicular to the gravity, the cells were separated effectively with no centrifugal force applied.

Figure 5 shows the images of inner outlet (a) and outer outlet (b) after the cell transportation experiment was completed. As shown in the micrographs, most of the HeLa cells were transported to the outer outlet (b), whereas a small amount of cells remained in the inner outlet (a), which might be transported before centrifugal force was generated. Figure 5(c,d) are the binary images of Figure 5 (a,b), and indicate the rough population of cells in those outlets, *i.e.*, the cross sectional areas of (c) and (d) are 211 pixels and 3022 pixels, respectively. Hence, the estimated transportation efficiency in this experiment was 93.5% [=3022/(3022 + 211)].

**Figure 5.** Images of sample outlets after the cell transportation. (a) Inner outlet; (b) Outer outlet; (c) and (d) are binary images of cells in (a) and (b), respectively; (e) The positions of outlets.



### 3. Experimental Section

#### 3.1. Optical Setup

For the optical observation, a charge-coupled device (CCD) camera (TOSHIBA JK-TU53H) equipped with  $\times 4$  objective lens (OLYMPUS, Uplan Fl 4 $\times$ ) was used to acquire images and video clips of the cells inside the chip.

#### 3.2. Samples

HeLa cell line was obtained commercially (Dainippon Sumitomo Pharma) and were suspended in PBS at a concentration of about  $10^4$ – $10^5$  cells per milliliter. Erythrocytes were obtained from horse blood and were suspended in physiological saline at a hematocrit of 3.3% for experiments. The cells

were suspended in the standard phosphate buffer and applied to the chip. For the washing buffer, the same standard phosphate buffer containing Rhodamine B was applied.

### 3.3. Experiment Procedure

Just before rotation starts, 50  $\mu\text{L}$  of sample cell buffer was applied into the inner inlet, and 50  $\mu\text{L}$  of sample cell buffer was applied into the outer inlet. Then the rotation was started and CCD observation was recorded. After whole buffer at the inlets were transported to the outlets, rotation was stopped and the chip was removed from the rotating stage, and observed the outlets by optical microscope.

### 3.4. Flow Velocity Estimation

Flow velocity in the microfluidic flowchannels were estimated by the measurement of movement of cells in the flowchannels by CCD camera images (frame rate, 1/30 s intervals).

### 3.5. Superimposed Image Acquisition and Image Analysis

For superimposed, quantitative evaluation and comparison of images, we used Image J software. The acquired analog CCD camera images were recorded in Digital Tape Recorder (SONY WV-D9000), and the recorded data was digitized and stored using image converter of PC (SONY RX-71). The digitized bit map images were then imported Image J software for the further processing.

## 4. Conclusions

In this paper, we introduced the experimental results of a fast and simple cell medium exchange using a double Y-shape microfluidic chip and centrifugal force. Based on the simple rotating platform, we demonstrated that cells can shift effectively from one pathway to the other without any mixture of two laminar flows of different buffers even using 1  $g$  acceleration. Using appropriate buffer solutions, the reported platform can be used for continuous automation of quick and single-step minimum buffer exchange in many other well-known processes of cell studies, such as quick exchange of enzymatical treatment buffer without dilution steps or inhibitor additions, quick staining of cells with minimum volume of expensive fluorescent dye, or precise time course measurement of effect of compounds.

## Acknowledgements

We sincerely thank the deceased K. Orita for his experimental help and assistance. This work was financially supported by the Kanagawa Academy of Science of Technology.

## References

1. Haeberle, S.; Brenner, T.; Zengerle, R.; Ducree, J. Centrifugal extraction of plasma from whole blood on a rotating disk. *Lab Chip* **2006**, *6*, 776–781.
2. Haeberle, S.; Zengerle, R. Microfluidic platforms for lab-on-a-chip applications. *Lab Chip* **2007**, *7*, 1094–1110.

3. Bhagat, A.A.; Kuntaegowdanahalli, S.S.; Papautsky, I. Continuous particle separation in spiral microchannels using Dean flows and differential migration. *Lab Chip* **2008**, *8*, 1906–1914.
4. Hsu, C.H.; Di Carlo, D.; Chen, C.; Irimia, D.; Toner, M. Microvortex for focusing, guiding and sorting of particles. *Lab Chip* **2008**, *8*, 2128–2134.
5. Meagher, R.J.; Light, Y.K.; Singh, A.K. Rapid, continuous purification of proteins in a microfluidic device using genetically-engineered partition tags. *Lab Chip* **2008**, *8*, 527–532.
6. Tai, C.H.; Hsiung, S.K.; Chen, C.Y.; Tsai, M.L.; Lee, G.B. Automatic microfluidic platform for cell separation and nucleus collection. *Biomed. Microdevices* **2007**, *9*, 533–543.
7. Chang, S.; Cho, Y.H. A continuous size-dependent particle separator using a negative dielectrophoretic virtual pillar array. *Lab Chip* **2008**, *8*, 1930–1936.
8. Yasuda, K.; Umemura, S.-i.; Takeda, K. Concentration and fractionation of small particles in liquid by ultrasound. *Jpn. J. Appl. Phys.* **1995**, *34*, 2715.
9. Yasuda, K.; Kamakura, T. Acoustic radiation force on micrometer-size particles. *Appl. Phys. Lett.* **1997**, *71*, 1771–1773.
10. Yasuda, K.; Haupt, S.S.; Umemura, S.-i.; Yagi, T.; Nishida, M.; Shibata, Y. Using acoustic radiation force as a concentration method for erythrocytes. *J. Acoust. Soc. Am.* **1997**, *102*, 642–645.
11. Yasuda, K. Non-destructive, non-contact handling method for biomaterials in micro-chamber by ultrasound. *Sens. Actuators B: Chem.* **2000**, *64*, 128–135.
12. Castell, O.K.; Allender, C.J.; Barrow, D.A. Novel biphasic separations utilising highly selective molecularly imprinted polymers as biorecognition solvent extraction agents. *Biosens. Bioelectron.* **2006**, *22*, 526–533.
13. Castell, O.K.; Allender, C.J.; Barrow, D.A. Continuous molecular enrichment in microfluidic systems. *Lab Chip* **2008**, *8*, 1031–1033.
14. Castell, O.K.; Allender, C.J.; Barrow, D.A. Liquid-liquid phase separation: Characterisation of a novel device capable of separating particle carrying multiphase flows. *Lab Chip* **2009**, *9*, 388–396.
15. Pamme, N.; Manz, A. On-chip free-flow magnetophoresis: Continuous flow separation of magnetic particles and agglomerates. *Anal. Chem.* **2004**, *76*, 7250–7256.
16. Yung, C.W.; Fiering, J.; Mueller, A.J.; Ingber, D.E. Micromagnetic-microfluidic blood cleansing device. *Lab Chip* **2009**, *9*, 1171–1177.
17. Atencia, J.; Beebe, D.J. Steady flow generation in microcirculatory systems. *Lab Chip* **2006**, *6*, 567–574.
18. Rhee, S.W.; Taylor, A.M.; Cribbs, D.H.; Cotman, C.W.; Jeon, N.L. External force-assisted cell positioning inside microfluidic devices. *Biomed. Microdevices* **2007**, *9*, 15–23.
19. Min, J.; Kim, J.H.; Lee, Y.; Namkoong, K.; Im, H.C.; Kim, H.N.; Kim, H.Y.; Huh, N.; Kim, Y.R. Functional integration of DNA purification and concentration into a real time micro-PCR chip. *Lab Chip* **2011**, *11*, 259–265.
20. Gong, M.; Bohn, P.W.; Sweedler, J.V. Centrifugal sedimentation for selectively packing channels with silica microbeads in three-dimensional micro/nanofluidic devices. *Anal. Chem.* **2009**, *81*, 2022–2026.
21. Lai, S.; Wang, S.; Luo, J.; Lee, L.J.; Yang, S.T.; Madou, M.J. Design of a compact disk-like microfluidic platform for enzyme-linked immunosorbent assay. *Anal. Chem.* **2004**, *76*, 1832–1837.



22. Grumann, M.; Steigert, J.; Riegger, L.; Moser, I.; Enderle, B.; Riebeseel, K.; Urban, G.; Zengerle, R.; Ducree, J. Sensitivity enhancement for colorimetric glucose assays on whole blood by on-chip beam-guidance. *Biomed. Microdevices* **2006**, *8*, 209–214.
23. Chen, B.; Zhou, X.; Li, C.; Wang, Q.; Liu, D.; Lin, B. Rapid screening of phenylketonuria using a CD microfluidic device. *J. Chromatogr. A* **2011**, *1218*, 1907–1912.
24. Haerberle, S.; Brenner, T.; Zengerle, R.; Ducree, J. Centrifugal extraction of plasma from whole blood on a rotating disk. *Lab Chip* **2006**, *6*, 776–781.
25. Chen, H.; Li, X.; Wang, L.; Li, P.C. A rotating microfluidic array chip for staining assays. *Talanta* **2010**, *81*, 1203–1208.
26. Qin, D.; Xia, Y.; Whitesides, G.M. Soft lithography for micro- and nanoscale patterning. *Nat. Protoc.* **2010**, *5*, 491–502.
27. Hattori, A.; Yasuda, K. Comprehensive study of microgel electrode for on-chip electrophoretic cell sorting. *Jpn. J. Appl. Phys.* **2010**, *49*, 06GM04:1–06GM04:4.

© 2012 by the authors; licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution license (<http://creativecommons.org/licenses/by/3.0/>).