

SUPPORTING INFORMATION

Method validation against DEQAS




Figure S1 Comparison between DEQAS data for 25(OH)D scheme and our lab results. Solid dots denote blind results submitted to DEQAS as lab 2388 prior to publication of the report. Red area indicates DEQAS acceptable range ($\pm 25\%$ from the target value)




Figure S2 Comparison between DEQAS data for 1,25(OH)₂D scheme and our lab results. Solid dots denote blind results submitted to DEQAS as lab 2388 prior to publication of the report. Red area indicates DEQAS acceptable range ($\pm 30\%$ from the target value)