



## Supplementary Tables

**Table S1.** T2\* mapping scan parameters

Scanner	Institution	Field strength	Manufacturer	Model name	Flip angle (degree)	FOV	Slice Thickness (mm)	Pixel spacing (mm <sup>2</sup> )	Matrix size
MR-A1	A	1.5T	Philips	Achieva dStream	20	300*300	10	1.34*1.34	224*224
MR-A2	A	3.0T	Philips	Ingenia CX	25	300*300	8	1.34*1.34	224*224
MR-A3	A	3.0T	Siemens	Prisma Fit	15	300*196	8	1.17*1.17	256*168
MR-B	B	3.0T	Siemens	Verio	12	380*332	10	1.17*1.17	256*176

FOV: field of view.

---

**Table S2.** Repetition time (TR) and echo time (TE)

Scanner	TR (ms)	TE (1)	TE (2)	TE (3)	TE (4)	TE (5)	TE (6)	TE (7)	TE (8)
MR-A1	14.6	1.2	3.2	5.2	7.2	9.3	11.3	13.3	-
MR-A2	14.4	1.1	2.8	4.6	6.3	8	9.7	11.5	13.2
MR-A3	20	1.9	4.4	6.9	9.5	12	14.6	-	-
MR-B	20	2.5	4.8	7.2	9.5	11.8	14.1	16.4	18.8

TE: echo time; TR: repetition time.