


Mass Spectrum



Sample Information

Month-Day Processed : 12/10/20
 Time Processed : 07:39:09 PM
 Injection Volume : 0.3
 Sample Name : CeHS-1 GP
 Sample ID : U461AFJ270-3
 Theoretical MW : 4016.78
 Observed MW : 4016.0

Interface

Nebulizing Gas Flow : 1.5L/min
 CDL Temp : 250
 Block Temp : 200

:ESI

Drying Gas Flow : 5 L/min
 T.Flow : 0.2 ml/min
 B.conc : 50% H₂O/50% MeOH

Equipment

Interface Bias : +4.5 kV
 Equipment : GK11010007