

Supplementary Table 1. The HPLC method for fingerprint profile of *Nepenthes thorellii* x (*ventricosa* x *maxima*).

HPLC	Shimadzu SIL-10AD <i>VP</i> auto injector Shimadzu LC-20AD prominence liquid chromatograph Shimadzu SPD-M10A <i>VP</i> diode array detector																																			
Solid phase	Phenomenex Luna 5 μ C18(2) 100A 250*4.60 mm, 5 μ P/NO 00G-4252-E0																																			
Mobile phase	A: 0.1% TFA _(aq) B: ACN <table><tr><td></td><td>Time</td><td>Module</td><td>Action</td><td>Value</td></tr><tr><td>1</td><td>0.10</td><td>Pumps</td><td>B.Conc</td><td>20</td></tr><tr><td>2</td><td>60.00</td><td>Pumps</td><td>B.Conc</td><td>100</td></tr><tr><td>3</td><td>70.00</td><td>Pumps</td><td>B.Conc</td><td>100</td></tr><tr><td>4</td><td>80.00</td><td>Pumps</td><td>B.Conc</td><td>20</td></tr><tr><td>5</td><td>90.00</td><td>Pumps</td><td>B.Conc</td><td>20</td></tr><tr><td>6</td><td>90.00</td><td>Controller</td><td>Stop</td><td></td></tr></table>		Time	Module	Action	Value	1	0.10	Pumps	B.Conc	20	2	60.00	Pumps	B.Conc	100	3	70.00	Pumps	B.Conc	100	4	80.00	Pumps	B.Conc	20	5	90.00	Pumps	B.Conc	20	6	90.00	Controller	Stop	
	Time	Module	Action	Value																																
1	0.10	Pumps	B.Conc	20																																
2	60.00	Pumps	B.Conc	100																																
3	70.00	Pumps	B.Conc	100																																
4	80.00	Pumps	B.Conc	20																																
5	90.00	Pumps	B.Conc	20																																
6	90.00	Controller	Stop																																	
Flow rate	1 mL/min																																			
Analysis time	60 mins																																			
Pressure	112 kgf (20% ACN _(aq))																																			
Injection volume	50 μL																																			
Sample	EANT (1 mg/mL, in MeOH) plumbagin (1 mg/mL, in MeOH) <i>cis</i> -Isoshinanolone (1 mg/mL, in MeOH) Quercetin 3-O-(6"-n-butyl β-D-glucuronide) (1 mg/mL, in MeOH)																																			
Wavelength	210 nm, 254 nm																																			