

**Table S1.** Capacitance parameters and change over 23 hours with albumin solution (3 g/L).

	With silicone oil (n=472)					Without silicone oil (n=477)					
<b>Hou r</b>	Min cap	Max cap	Medi an	Mean min cap	SD	Mi n cap	Max cap	Medi an	Mean min cap	SD	P-value (mean min cap)
1	381	433	405	406	12	382	415	406	401	10	0.176
2	413	458	433	433	11	406	464	438	434	17	0.56
3	417	471	439	439	15	407	482	437	441	18	0.839
4	417	479	447	444	17	418	488	443	449	19	0.756
5	406	510	454	451	21	426	506	446	456	26	0.914
6	422	497	452	453	19	416	554	450	459	36	0.903
7	426	506	457	457	19	420	566	457	468	39	0.675
8	426	508	455	458	19	418	583	460	477	45	0.386
9	423	499	454	457	17	423	590	479	488	48	0.062
10	425	500	458	458	18	419	614	485	496	51	0.011
11	419	507	460	461	19	436	614	503	511	52	<0.001
12	425	508	461	462	19	437	627	522	523	57	<0.001
13	428	505	473	468	20	454	649	524	534	58	<0.001
14	436	487	469	467	15	467	670	524	541	66	<0.001
15	441	500	480	476	17	456	716	560	561	71	<0.001
16	430	503	478	476	19	447	727	545	567	76	<0.001
17	440	513	485	481	19	454	760	572	585	81	<0.001
18	441	520	488	483	22	452	754	596	600	79	<0.001
19	440	544	490	493	24	452	759	645	615	81	<0.001
20	437	535	497	496	24	479	729	659	639	65	<0.001
21	439	554	502	504	32	470	788	659	652	71	<0.001
22	440	556	502	501	27	494	790	670	665	63	<0.001
23	400	660	503	503	56	472	785	677	674	67	<0.001