



Article Microsphere Coupled Off-Core Fiber Sensor for Ultrasound Sensing

Gerard Tatel and Xiaoyi Bao *

Physics Department, University of Ottawa, Ottawa, ON K1N 6N5, Canada; gtate010@uottawa.ca * Correspondence: xbao@uottawa.ca

Abstract: A compact fiber ultrasound-sensing device comprising a commercially available Barium Titanate (BaTiO₃) glass microsphere coupled to an open cavity off-core Fabry–Perot interferometer (FPI) fiber sensor is proposed and demonstrated. The open cavity is fabricated through splicing two segments of a single mode fiber (SMF-28) at lateral offsets. The lateral offset is matched to the radius of the microsphere to maximize their coupling and allow for an increased sensing response. Furthermore, the microsphere can be moved along the open-air cavity to allow for tuning of the reflection spectrum. The multiple passes of the FPI enabled by the high refractive index microsphere results in a 40 dB enhancement of finesse and achieves broadband ultrasound sensing from 0.1–45.6 MHz driven via a piezoelectric transducer (PZT) centered at 3.7 MHz. The goal is to achieve frequency detection in the MHz range using a repeatable, cost effective, and easy to fabricate FPI sensor design.

Keywords: Fabry-Perot interferometer; microsphere; ultrasound; off-core; finesse

1. Introduction

Fiber optic interferometric sensors have gained interest in recent times due to their distinctive characteristics, such as compact size, repeatability, and multi-parameter sensing capabilities. Due to their diverse sensing applications and inexpensive designs, these sensors have applications in strain monitoring [1-3], refractive index measurements [4-6], and temperature [7–9]. More specifically, Fabry–Perot interferometers (FPI) have been investigated thoroughly due to desirable characteristics, such as high finesse-associated narrow spectrum peaks and high sensitivity, compact size, and immunity to electromagnetic interference. FPI are typically formed by cascading reflective surfaces or reflectors along the light's propagation path and their interference. Fiber optic FPI sensing devices can be divided into two subcategories: intrinsic FPIs, where light interactions are inside of the fiber, which can be formed through techniques, such as thin film deposition [10]; Bragg gratings [11]; and micro machining [12], and extrinsic FPIs, where light interacts with external cavities, such as air or other polymers [13]. These open-air cavity FPIs are typically fabricated through expensive and complex methods, such as fs laser micromachining [14]. The sensing capabilities of these type of fiber optic sensors have been exploited and enhanced through new and old fiber optic technology innovations, such as fiber grating [15-17], surface plasmon resonance [18], and specialty fibers [19].

To mitigate the cost for these open-air cavity sensing devices, splicing two sections of single mode fiber segments at a lateral offset to form a simple and compact sensor has been explored [20–22]. Splicing several offset segments beyond two sections has also been explored [23]. At each silica–air interface there will be a reflection, essentially turning the device into multiple cascading FPIs. The large offset of this FPI allows light to spread from the main incoming fiber to both the offset cladding and the surrounding open-air cavity, this will yield two definitive interferometer arms with a large refractive index difference. These types of open cavity FPIs have been demonstrated to have a diverse range of sensing capabilities, such as temperature [24], refractive index [25], stress and strain [26,27], and ultrasound detection. For ultrasound detection, Fan et al. detail that the importance of high reflectivity, relatively large contrast, and linewidth would benefit the high frequency



Citation: Tatel, G.; Bao, X. Microsphere Coupled Off-Core Fiber Sensor for Ultrasound Sensing. *Sensors* 2022, 22, 5328. https:// doi.org/10.3390/s22145328

Academic Editor: Mario Iodice

Received: 31 May 2022 Accepted: 15 July 2022 Published: 16 July 2022

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Copyright: © 2022 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (https:// creativecommons.org/licenses/by/ 4.0/). ultrasound sensor response [28]. Because the displacement of ultrasound sensing at MHz can be as small as sub-um [29], which is impossible to be detected with interferometers due to the diffraction limit of the laser wavelength, to increase the detection sensitivity for intensity detection, the sensor must operate at a quadrature point where relative intensity change versus wavelength shifts due to the displacement of the ultrasound signal, defined as intensity slope. To increase the sensitivity of the ultrasound sensing at the highest frequency, the intensity slope should be at its maximum. To improve this slope for FPI, the finesse is proportional to intensity slope change. For a fixed reflectivity and cavity length, the spectrum of the FPI exhibits narrow linewidth with a high finesse \mathcal{F} . Light propagates in more FP cavities representing a higher finesse \mathcal{F}^N contributed to the product of the finesse of N individual FPIs with the same resonant frequency, these individual FPIs are represented by curved arrows in Figure 1b. To reach this target, we added a microsphere, as shown in Figure 1a, with a high refractive index between two sections of the off-core fiber (open cavity); through the front and back surfaces of the microsphere, the first FPI reaches the double pass, and the same principle is applied to the second FPI, the front and back surface of the microsphere introduced the 2nd double pass FPI, which leads to a total of \mathcal{F}^6 , which can be seen in Figure 1b, circled in red. In addition to the FPI finesse enhancement, the microsphere itself is a resonator device, which enables light to be stored and confined at a resonant frequency. This confined light will circulate within the device through total internal reflection.



Figure 1. (a) 3D representation of the proposed sensor. (b) Multiple FPI paths depicted by curved arrows; the blue arrow depicts the path through air exclusively.

In this paper, an extrinsic FPI fiber sensor is proposed and fabricated by the lateral offset splicing of two single mode fiber segments (Corning SMF-28) acting as the base structure, which is then modified and enhanced by utilizing the benefits of increased finesse by adding a Barium Titanate (BaTiO₃) glass microsphere (refractive index of 1.94, commercially available from Cospheric LLC) coupled in the open-air cavity. The lateral offset and microsphere diameter are carefully matched to achieve the maximum contrast; this will allow for an increase in sensitivity. Light is reflected at each air–silica boundary. From the reflection spectrum, quadrature points are selected due to the wavelength shift-induced maximum intensity change and are used for ultrasound detection. The fabricated sensor is tested with an ultrasound source generated from a piezoelectric transducer (PZT) centered at 3.7 MHz attached to a thin steel plate with the excitation of its high-order harmonics to increase the ultrasound frequency range.

2. Materials and Methods

A visual representation of the proposed sensing device is illustrated in Figure 1a. To begin, after prepping two SMF segments for splicing and loading it into the fiber splicer (Ericsson Cables, Sundbyberg, Sweden, FSU 995 FA), manual mode is selected. Using the two views in the splicer (side and top view), the top view is aligned (via lining up the SMF core) while the other is set to an offset distance h (the distance from the incoming SMF center point of the core to the 1st segment SMF's edge), circled in red in Figure 1a. The offset is imperative as it will affect the beam path through the air and microsphere interfaces, which is vital for characteristics, such as contrast and reflectivity. Once the leading fiber

and 1st segment is fused, it is carefully cleaved under a confocal microscope to the desired length L_1 . This length will be the air cavity enclosed by two SMF segments (incoming SMF and end segment SMF). Similarly, the end segment is spliced and fused to the previous SMF segment, making sure all the cores align in the top view. Once the end segment is matched to the same offset it is spliced and fused; it is brought under the confocal microscope again for the final cleave at the desired length L_2 . Under the microscope and shielded from the environmental effects, a single BaTiO₃ microsphere is selected and picked using fabricated fiber half tapers with a waist diameter of ~5 µm. Combining with translational stages at different positions, the microsphere is deposited onto the open-air cavity, and its position can be moved using the same fiber tapers. This process is illustrated in Figure 2.



Figure 2. Fabrication process of the proposed sensor.

In the most basic case (just the off-core), we can analyze the sensor as cascading FPIs and derived with 2-beam approximation with three mirrors. The total reflected electrical fields is approximated as E_r given by [30]:

$$E_r = A + Be^{-j\varphi_1} + Ce^{-j(\varphi_1 + \varphi_2)},$$
(1)

where the input field (E_0), transmission loss in the cavity (α), round trip propagation phase shifts ($\phi_{1,2}$) are given by:

$$\varphi_1 = \frac{4\pi n_{air} L_1}{\lambda}, \varphi_2 = \frac{4\pi n_{smf} L_2}{\lambda}, \tag{2}$$

$$A = E_0 R_1^{1/2}, B = E_0 (1 - \alpha) (1 - R_1) R_2^{1/2}, C = E_0 (1 - \alpha) (1 - R_1) (1 - R_2) R_3^{1/2},$$
(3)

$$R_1 = R_2 = R_3 = \left(\frac{n_{smf} - n_{air}}{n_{smf} + n_{air}}\right)^2,$$
(4)

Using Equations (1)–(4), the total reflection spectrum can be simplified to:

$$I_r = |A|^2 + |B|^2 + |C|^2 + 2AB\cos(\varphi_1) + 2BC\cos(\varphi_2) + 2AC\cos(\varphi_1 + \varphi_2)$$
(5)

From the above equations: a change in refractive index *n* between n_{smf} and n_{air} would change the phase ϕ and adding an extra path length along the air for the light to travel would change the spectrum. The reflection spectrum of an FPI is wavelength dependent intensity modulation, generally caused by the optical phase difference between the two beams. The optical path difference caused by the microsphere is given by [31]:

$$OPD_{ms} = n\pi d, \tag{6}$$

where *n* is the microsphere refractive index and *d* is the diameter of the microsphere.

The initial FPI structure (without the microsphere) has been shown previously to be capable of enhancing high-order mode interference, which increases multi-mode interference and improves the intensity change with wavelength shifts (i.e., spectrum slope). The slope of the reflection spectrum given by $S = dR/d\lambda$, where *S* is the slope and *R* is the reflectivity; this details that a greater slope indicates a stronger multimode interference, which leads to better sensitivity. This corresponds to the spectrum's quadrature points. The off-core FPI has the capability to add extra modes via the air and silica cladding to enhance the quality factor with an additional one or two round trips, and in turn leads to an improved ultrasound response. The light propagating from the fiber core will also be coupled to the microsphere. The barium in the glass microsphere composition allows for an increase in refractive index to 1.94 (previously mentioned). The higher refractive index allows for the confined light within the microsphere to slow down through this medium allowing for more bends total internal reflection, all of this will lead to more efficient refraction.

3. Results and Discussion

An experimental setup can be seen in Figure 3. The reflection spectrum is obtained through an optical circulator using an erbium-doped fiber amplifier source (INO, Quebec, QC, Canada, FAF-50) and optical spectrum analyzer (Yokogawa, Tokyo, Japan, AQ6375). The reflection spectrum of the microsphere coupled off-core sensor is analyzed under different microsphere diameters, illustrated in Figure 4, which have an offset distance of ~11.4 µm. Different microsphere diameters yield different reflection spectrum for a specific offset. The microsphere diameter and core offset distance play a direct contribution to the reflection contrast. This enhanced contrast and reflectivity determine the dynamic range and signal strength of this device. When the input light from the main fiber core travels through the first air-silica boundary into the open cavity it will meet the microsphere; this leads to a corresponding large number of reflections within the microsphere, which allows for high-order modes and a large contrast increase. This is heavily dependent on the alignment and core offset. When the microsphere diameter is too large compared to the core offset it leads to large propagation loss and a diminished contrast resulting in weak multi-mode interference. The offset distance h should be equal to the radius of the microsphere. This alignment is crucial because the light path from the incoming core will be able to directly interact with the apex point of the microsphere (facing the first silica/air interface) allowing for an increase in reflection and coupling. Since the offset distance is ~11.4 μ m, the 5 μ m, 10 μ m, 16 μ m, and 50 μ m diameter microspheres are not ideal. The ideal microsphere for this specific offset distance would be with a 22.8 μ m diameter.



Figure 3. Experimental setup. The ultrasound generation and detection components are indicated within the red dashed box.

Figure 5a depicts the change in the reflection spectrum as the microsphere distance from the first silica–air interface increases (along the center) for a constant open cavity segment length L_1 . As the distance between the beginning of the cavity and microsphere increases, the free spectral range of the reflection decreases. It is determined that the microsphere location with the highest contrast is located between the first 1/3 to $\frac{1}{2}$ distance of the off-core open-air segment. Figure 5b shows the test repeated with different samples (with relatively the same h), changing L_1 to ensure its consistency where the distance is normalized with respect to the open-air cavity length. Once the microsphere is positioned longitudinally (along the fiber direction), it can be moved by a few micrometers (much smaller than the diameter of the microsphere, which is 22 µm, to ensure low loss from leaky modes) laterally (perpendicular to the fiber direction) to display a full FSR shift, which can be seen in Figure 5c. This shift will generally maintain the same shape and FSR of the spectrum due to the FP cavity between the off-core fiber and microsphere, allowing further tunability.



Figure 4. Reflection Spectrum of the same off-core configuration with different microsphere diameters (i) 5 μ m (ii) 10 μ m (iii) 16 μ m (iv) 50 μ m.



Figure 5. (a) Microsphere location with respect to the first silica–air interface (start of the open cavity) (i) No microsphere (ii) 50 μ m (iii) 100 μ m (iv) 150 μ m (v) 200 μ m (vi) 250 μ m (b) Normalized microsphere location for samples with different open cavity lengths with an offset of ~11–12 μ m and an optimized microsphere diameter of ~22 μ m. (c) a full FSR spectrum shift via moving the microsphere laterally at a fixed longitudinal position.

Analyzing the data given in both Figures 4 and 5, a sample is fabricated seen in Figure 6a, ensuring the offset distance and the microsphere radius is comparable, and is located between 1/3 and $\frac{1}{2}$ of the total open cavity distance. A green light is sent through the fiber to confirm the offset and microsphere are aligned (Figure 6b). The specific sensor parameters for the device under the test can be seen in Table 1. As a result of the multibeam interference, enhanced finesse via an increased number of FPIs and strong dips in the reflection spectrum allow for a broad tuning range, typically associated with higher sensitivity. The reflection spectrum and its initial spectrum before adding the microsphere are shown in Figure 7. A maximum contrast of ~44 dB can be achieved, which is a drastic increase in contrast when comparing it to its initial off-core case with no microsphere of ~4.2 dB, an increase of over 10 times. If we suppose a single pass FPI (without the microsphere), the finesse is 4, corresponding to the blue spectra in Figure 7, with 6 passages, the new finesse is 4^6 which is equivalent to 4096, 30 dB. Although, with the multiple surface reflections due to the high reflective index within the microsphere, the new finesse is 4^8 , 45 dB, which has an enhancement factor close to the red spectra.



Figure 6. (a) Microscope image of the fabricated sample described in Table 1. (b) Microscope image of green light passing through the sensor to help with alignment.

Table 1. Sensor device specifications.

| Specification | |
|-----------------------------|---------|
| Offset (h) | 11.4 μm |
| 1st Segment Length (L_1) | 381 µm |
| End Segment Length (L_2) | 155 μm |
| BaTiO ₃ Diameter | 22 µm |

To test the device for its ultrasound sensing response, a piezoelectric transducer (PZT) fixed to a thin steel plate (0.25 mm thickness) is used as the ultrasound source driven by a function generator (Agilent, Santa Clara, CA, USA, 33250A) (seen previously in Figure 1 boxed in red). A tunable laser source (Agilent 81940A) is set to the specific wavelength 1545.7 nm, which corresponds to a quadrature point in the reflection spectrum for its steep slope and large contrast. It should be noted, although a large contrast is desirable for improved sensitivity, a steep slope plays a more crucial role. The locked probe wavelength can respond to the periodic modulation from the propagating ultrasound waves, which can be analyzed through an electronic spectrum analyzer (ESA) (Rohde & Shwarz, Columbia, MD, USA, F&W Signal Spectrum Analyzer) via a photodetector (Thorlabs, Newton, NJ, USA, PDB45QC-AC). The device under test is encapsulated inside of an acrylic case for added protection from external disturbances, and the ESA is placed in a separate room separated by a concrete wall to eliminate any sort of antenna effect from the ultrasound generation. Figure 8a shows the sample's frequency response from 0.1–45.6 MHz. The importance of selecting a proper quadrature point is crucial to the device's ultrasound detection sensitivity. Figure 8b depicts the ultrasound response of an unideal offset-microsphere pairing tested to highlight its importance. Different quadrature points correspond to different maximum

frequency responses. This is due to the different resonant modes in the microcavity and its multimode interference. This device can be compared to Fan et al. [26,28], where the microsphere is glued to the off-core fiber and without a microsphere at all, which limited the tuning range; numerous samples were required due to the fixed nature of the off-core FPI cavity to optimize the spectral contrast, which limited the specific contrast, reflectivity, and linewidth. With the flexible lay sphere in our approach, one sample is adequate to optimize the spectrum for maximum contrast in the reflection spectra. If the microsphere radius equals the offset distance, the spectrum can be aligned at quadrature points for ultrasound sensing measurements with the highest sensitivity.



Figure 7. Optimized offset and microsphere diameter reflection spectrum.



Figure 8. (a) Ultrasound response (b) Ultrasound response with mismatched offsets and microsphere size.

4. Conclusions

This paper gave an overview and background of a proposed extrinsic FPI sensor enhanced by multiple pass FPI due to a BaTiO₃ glass microsphere coupled to the open cavity of a two segment SMF off-core device. The light coupled within the microsphere allows for scattering and internal reflection, which enhances the finesse. Offset alignment of the main incoming SMF and the center of the microsphere is imperative for maximum slope and contrast in the reflection spectrum, which allows for ideal quadrature points to be determined that are best suited for ultrasound detection. This simple and inexpensive sensor offers more applications beyond ultrasound sensing, including refractive index sensing, temperature monitoring, and biological sensing.

Author Contributions: Conceptualization, methodology, formal analysis, investigation, resources, data curation, writing—original draft preparation, writing—review and editing and visualization, G.T. and X.B. (All authors); supervision, X.B. All authors have read and agreed to the published version of the manuscript.

Funding: Canada Research Chairs (950231352); Natural Sciences and Engineering Research Council of Canada (RGPIN/06302-2020, STPGP 506628). New Frontiers in Research Fund of Canada: NFRFE-2019-00504.

Informed Consent Statement: Not applicable.

Data Availability Statement: The data presented in this study are available upon request from the corresponding author.

Acknowledgments: We thank the members of Xiaoyi Bao's Fiber Optics Group in The University of Ottawa, for use of the lab and equipment.

Conflicts of Interest: The authors declare no conflict of interest.

References

- Kim, G.; Cho, T.; Hwang, K.; Lee, K.; Lee, K.S.; Han, Y.G.; Lee, S.B. Strain and temperature sensitivities of an elliptical hollow-core photonic bandgap fiber based on Sagnac interferometer. *Opt. Express* 2009, *17*, 2481–2486. [CrossRef] [PubMed]
- Sirkis, J.S.; Brennan, D.D.; Putman, M.A.; Berkoff, T.A.; Kersey, A.D.; Friebele, E.J. In-line fiber etalon for strain measurement. *Opt. Lett.* 1973, 18, 1973–1975.
- Shi, Q.; Li, F.; Wang, Z.; Jin, L.; Hu, J.J.; Liu, J.; Kai, G.; Dong, X. Environmentally stable Fabry-Perot-type strain sensor based on hollow-core photonic bandgap fiber. *IEEE Photon. Technol. Lett.* 2008, 20, 237–239. [CrossRef]
- 4. Kim, D.W.; Zhang, Y.; Cooper, K.L.; Wang, A. In-fiber reflection mode interferometer based on a long-period grating for external refractive-index measurement. *Appl. Opt.* **2005**, *44*, 5368–5373. [CrossRef]
- Choi, H.Y.; Mudhana, G.; Park, K.S.; Paek, U.C.; Lee, B.H. Cross-talk free and ultra-compact fiber optic sensor for simultaneous measurement of temperature and refractive index. *Opt. Express* 2009, 18, 141–149. [CrossRef]
- 6. Chen, D.; Liu, W.; Jiang, M.; He, S. High-resolution strain/temperature sensing system based on a high-finesse fiber cavity and time-domain wavelength demodulation. *J. Lightw. Techol.* **2009**, *27*, 2477–2481. [CrossRef]
- Frazao, O.; Aref, S.H.; Baptista, J.M.; Santos, J.L.; Latifi, H.; Farahi, F.; Kobelke, J.; Schuster, K. Fabry-Perot cavity based on a suspended-core fiber for strain and temperature measurement. *IEEE Photon. Technol. Lett.* 2009, 21, 1229–1231. [CrossRef]
- Deng, M.; Tang, C.P.; Zhu, T.; Rao, Y.J. PCF-based Fabry-Perot interferometric sensor for strain measurement at high temperatures. *IEEE Photon. Technol. Lett.* 2011, 23, 700–702. [CrossRef]
- Choi, H.Y.; Park, K.S.; Park, S.J.; Paek, U.C.; Lee, B.H.; Choi, E.S. Miniature fiber-optic high temperature sensor based on a hybrid structured Fabry-Perot interferometer. *Opt. Lett.* 2008, 33, 2455–2457. [CrossRef]
- Rao, Y.J.; Deng, M.; Duan, D.W.; Yang, X.C.; Zhu, T.; Cheng, G.H. Micro Fabry-Perot interferometers in silica fibers machined by femtosecond laser. *Opt. Express* 2007, 15, 14123–14128. [CrossRef]
- 11. Wang, Z.; Shen, F.; Song, L.; Wang, X.; Wang, A. Multiplexed fiber Fabry-Pérot interferometer sensors based on ultrashort Bragg gratings. *IEEE Photon. Technol. Lett.* 2007, *19*, 622–624. [CrossRef]
- 12. Morris, P.; Hurrell, A.; Shaw, A.; Zhang, E.; Beard, P. A Fabry-Perot fiber-optic ultrasonic hydrophone for the simultaneous measurement of temperature and acoustic pressure. *J. Acoust. Soc. Am.* **2009**, *125*, 3611–3622. [CrossRef] [PubMed]
- 13. Rao, Y.J. Recent progress in fiber-optic extrinsic Fabry Perot interferometric sensors. Opt. Fiber Technol. 2006, 12, 227–237. [CrossRef]
- 14. Lin, C.; Jiang, L.; Xiao, H.; Chai, Y.; Chen, H.; Tsai, H. Fabry–Perot interferometer embedded in a glass chip fabricated by femtosecond laser. *Opt. Lett.* 2009, 34, 2408–2410. [CrossRef] [PubMed]
- 15. Lim, J.H.; Jang, H.S.; Lee, K.S.; Kim, J.C.; Lee, B.H. Mach-Zehnder interferometer formed in a photonic crystal fiber based on a pair of long-period fiber gratings. *Opt. Lett.* **2004**, *29*, 346–348. [CrossRef]
- Kim, Y.J.; Paek, U.C.; Lee, B.H. Measurement of refractive-index variation with temperature by use of long-period fiber gratings. Opt. Lett. 2002, 27, 1297–1299. [CrossRef]
- 17. Ding, J.-F.; Zhang, A.P.; Shao, L.-Y.; Yan, J.-H.; He, S. Fiber-taper seeded long-period grating pair as a highly sensitive refractiveindex sensor. *IEEE Photon. Technol. Lett.* **2005**, *17*, 1247–1249. [CrossRef]
- Sharma, A.K.; Jha, R.; Gupta, B.D. Fiber-optic sensors based on surface plasmon resonance: A comprehensive review. *IEEE Sens.* J. 2007, 7, 1118–1129. [CrossRef]

- 19. Lee, B.H.; Eom, J.B.; Park, K.S.; Park, S.J.; Ju, M.J. Specialty fiber coupler; Fabrications and applications. J. Opt. Soc. Korea 2010, 14, 326–332. [CrossRef]
- 20. Baharin, N.; Azmi, A.; Abdullah, A.; Noor, M. Refractive index sensor based on lateral-offset of coreless silica interferometer. *Opt. Laser Technol.* **2018**, 99, 396–401. [CrossRef]
- Zhou, M.; Shenglai, Z.; Lui, F.; Peng, J.; Li, L. Single-mode fiber sensor based on core-offset splicing. In Proceedings of the International Conference on Optical Instruments and Technology: Optical Sensors and Applications, Beijing, China, 6–9 November 2011; p. 8199. [CrossRef]
- Gao, H.; Wang, J.; Shen, J.; Zhang, S.; Xu, D.; Zhang, Y.; Li, C. Study of the Vernier Effect Based on the Fabry–Perot Interferometer: Methodology and Application. *Photonics* 2021, *8*, 304. [CrossRef]
- 23. Fan, H.; Chen, L.; Bao, X. Fiber-Optic Sensor Based on Core-Offset Fused Unequal-Length Fiber Segments to Improve Ultrasound Detection Sensitivity. *IEEE Sens. J.* 2020, 20, 9148–9154. [CrossRef]
- 24. Hao, X.; Tong, Z.; Zhang, W.; Cao, Y. A fiber laser temperature sensor based on SMF core-offset structure. *Opt. Commun.* 2015, 335, 78–81. [CrossRef]
- Zhao, Y.; Jiang, J.; Yang, Y.; Wang, Y.; Lu, J.; Jiacheng, L. High-Sensitive All-Fiber Fabry-Perot Interferometer Gas Refractive Index Sensor Based on Lateral offset Splicing and Vernier Effect. In Proceedings of the 18th International Conference on Optical Communications and Networks (ICOCN), Huangshan, China, 5–8 August 2019; pp. 1–3. [CrossRef]
- Fan, H.; Chen, L.; Bao, X. Temperature-Insensitive Strain Sensor Based on Microsphere-Embedded Core-Offset Fiber with High Sensitivity. J. Lightwave Technol. 2021, 39, 2547–2551. [CrossRef]
- Yongfeng, W.; Zhang, Y.; Wu, J.; Yuan, P. Fiber-Optic Hybrid-Structured Fabry–Perot Interferometer Based on Large Lateral Offset Splicing for Simultaneous Measurement of Strain and Temperature. J. Lightwave Technol. 2017, 35, 4311–4315.
- Fan, H.; Zhang, L.; Gao, S.; Chen, L.; Bao, X. Ultrasound sensing based on an in-fiber dual-cavity Fabry–Perot interferometer. Optic Letters 2019, 44, 3606–3609. [CrossRef]
- 29. Poiana, D.A.; Garcia-Souto, J.A.; Bao, X. Acousto-optic comb interrogation system for random fiber grating sensors with sub-nm resolution. *Sensors* **2021**, *21*, 3967. [CrossRef]
- Yang, Y.; Yongguang, W.; Juixing, J.; Zhao, Y.; He, X.; Li, L. High-sensitive all-fiber Fabry-Perot interferometer gas refractive index sensor based on lateral offset splicing and Vernier effect. *Optik* 2019, 196, 163181. [CrossRef]
- 31. Bai, X.; Wand, D. An In-Fiber Coupler for Whispering-Gallery-Mode Excitation in a Microsphere Resonator. *IEEE Photonics Technol. Lett.* **2020**, *32*, 188–191. [CrossRef]