

Table S1. Imprecision (CV%) and inaccuracy (%) values of analytical methods evaluated on the highest and lower calibration points.

	Imprecision (CV%)	Inaccuracy (%)
ANCA (ng/mL)		
20	0.07	−0.10
0.62	0.29	14.5
MPO (ng/mL)		
29.7	3.74	0.28
0.04	11.0	−10.7
PR3 (ng/mL)		
10	3.6	0.05
0.16	1.1	−0.11