

Supporting Information

Synthesis and Evaluation of Small Molecule Drug Conjugates Harnessing Thioester-linked Maytansinoids

Chen-Fu Lo^{1#}, Tai-Yu Chiu^{1#}, Yu-Tzu Liu^{2#}, Li-Rung Huang^{2#}, Teng-Kuang Yeh^{1#}, Kuan-Hsun Huang¹, Kuan-Liang Liu¹, Chia-Yu Hsu¹, Ming-Yu Fang¹, Yu-Chen Huang¹, Tsu-An Hsu¹, Chiung-Tong Chen¹, Lun Kelvin Tsou*¹

¹ Institute of Biotechnology and Pharmaceutical Research, National Health Research Institutes, Miaoli 35053, Taiwan, ROC


² Institute of Molecular and Genomic Medicine, National Health Research Institutes, Miaoli 35053, Taiwan, ROC

Correspondence and requests for materials should be addressed to L.K.T.
(kelvintsou@nhri.edu.tw)


Table of Contents

Supporting Figure 1	2-3
Copies of ¹ H NMR, ¹³ C NMR and mass spectra	4 -27

Zn8_DM1: 0H VS. 24H in plasma.



Zn10_DM4: 0H VS. 24H in plasma.




Supporting Figure S1. Incubation of Linker-drug 3 and 4, conjugates Zn8_DM1 and Zn10_DM4, DM1 and DM4 with plasma for 24 hrs. HPLC condition is listed below.


HPLC Information

Agilent 1100 HPLC was utilized in this project. Detailed chromatographic conditions used for purity analysis are listed in Table S1.

Table S1. Chromatographic conditions and parameters



Parameters	Value
Column	Thermo hypersil GOLD 250*4.6 particle: 5 μ m
Mobile Phase	A: 0.1% TFA in H ₂ O ; B: MeOH
Gradient table	Time(min) % B
	0 10
	2 10
	6 50
	16 90
	17 90
Flow rate	0.5 ml/min
Injection volume	10 μ L
Detector wavelength	UV at 254 nm
Run time	25 min
Post time	6 min
Column temperature	RT
Sample temperature	RT







Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	1.0 Bar
Focus	Active	Set Capillary	4500 V	Set Dry Heater	200 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	6.0 l/min
Scan End	1500 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Waste
		Set Corona	0 nA	Set APCI Heater	0 °C











Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	1.0 Bar
Focus	Active	Set Capillary	4500 V	Set Dry Heater	200 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	6.0 l/min
Scan End	1500 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Waste
		Set Corona	0 nA	Set APCI Heater	0 °C










Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	1.0 Bar
Focus	Active	Set Capillary	4500 V	Set Dry Heater	200 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	6.0 l/min
Scan End	1500 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Waste
		Set Corona	0 nA	Set APCI Heater	0 °C









Charge number:1


Tolerance:150.00[ppm], 100.00 .. 100....

Unsaturation Number:-200.5 .. 300.0 (...)


Element:¹²C:84 .. 84, ¹H:0 .. 94, ³⁵Cl:2 .. 2, ¹⁴N:9 .. 9, ²³Na:0 .. 2, ¹⁶O:8 .. 8

Mass	Intensity	Calc. Mass	Mass Difference [ppm]	Possible Formula
1426.66139	28257.29	1426.66024	0.81	¹² C ₈₄ ¹ H ₉₄ ³⁵ Cl ₂ ¹⁴ N ₉ ¹⁶ O ₈
1448.64210	106156.39	1448.64218	-0.06	¹² C ₈₄ ¹ H ₉₃ ³⁵ Cl ₂ ¹⁴ N ₉ ²³ Na ₁ ¹⁶ O ₈







Charge number:1 Tolerance:300.00[ppm], 250.00 .. 250.... Unsaturation Number:-100.0 .. 200.0 (...
Element:¹²C:119 .. 119, ¹H:0 .. 140, ³⁵Cl:3 .. 3, ¹⁴N:12 .. 12, ²³Na:0 .. 1, ¹⁶O:17 .. 17, ³²S:1 .. 1




Mass	Intensity	Calc. Mass	Mass Difference [mDa]	Mass Difference [ppm]
2167.90420	54324.13	2167.90651	-2.31	-1.06







Acquisition Parameter


Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	1.0 Bar
Focus	Active	Set Capillary	4500 V	Set Dry Heater	200 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	6.0 l/min
Scan End	3000 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Waste
		Set Corona	0 nA	Set APCI Heater	0 °C









Charge number:1 Tolerance:300.00[ppm], 250.00 .. 250.... Unsaturation Number:-100.0 .. 200.0 (...
Element:¹²C:122 .. 122, ¹H:0 .. 146, ³⁵Cl:3 .. 3, ¹⁴N:12 .. 12, ²³Na:0 .. 1, ¹⁶O:17 .. 17, ³²S:1 .. 1



Mass	Intensity	Calc. Mass	Mass Difference [mDa]	Mass Difference [ppm]
2209.95189	25511.96	2209.95346	-1.57	-0.71





Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	1.0 Bar
Focus	Active	Set Capillary	4500 V	Set Dry Heater	200 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	6.0 l/min
Scan End	3000 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Waste
		Set Corona	0 nA	Set APCI Heater	0 °C

