Article

Modeling and Reconstruction Strategy for Compton Scattering Tomography with Scintillation Crystals

Lorenz Kuger¹ and Gael Rigaud^{2,*} ¹ Department of Applied Mathematics and Theoretical Physics, University of Cambridge, Cambridge CB3 0WA, UK; mail@lorenzkuger.de² Center for Industrial Mathematics, University of Bremen, 28344 Bremen, Germany

* Correspondence: gael.rigaud@uni-bremen.de

Abstract: The recent development of energy-resolved scintillation crystals opens the way to build novel imaging concepts based on the variable energy. Among them, Compton scattering tomography (CST) is one of the most ambitious concepts. Akin to Computerized Tomography (CT), it consists in probing the attenuation map of an object of interest using external ionizing sources but strives to exploit the scattered radiation as an imaging agent. For medical applications, the scattered radiation represents 70 to 80% when the energy of the source is larger than 100 keV and results from the Compton effect. This phenomenon stands for the collision of a photon with an electron and rules the change of course and loss of energy undergone by the photon. In this article, we propose a modeling for the scattered radiation assuming polychromatic sources such as ⁶⁰Co and scintillation crystals such as LBC:Ce. Further, we design a general strategy for reconstructing the electron density of the target specimen. Our results are illustrated for toy objects.

Citation: Kuger, L.; Rigaud, G. Modeling and Reconstruction Strategy for Compton Scattering Tomography with Scintillation Crystals. *Crystals* **2021**, *11*, 641. <https://doi.org/10.3390/cryst11060641>

Academic Editors: Hiroshi Sakurai and Kosuke Suzuki

Received: 7 May 2021

Accepted: 31 May 2021

Published: 4 June 2021

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Copyright: © 2021 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

Keywords: Compton spectrum; scintillations crystals; modeling

1. Introduction

Invented and theorized by the Nobel medal award-winners G. Hounsfield and A.M. Cormack between 1963 and 1979, the concept of Computerized Tomography (CT) has become an essential way to investigate the inside of a human body or of any type of medium. Its principle relies on the phenomenon of attenuation, characterized by the map μ_E , suffered by a photon-beam which travels through matter. This phenomenon is ruled by the Beer–Lambert law

$$I(\mathbf{x} + T\theta, \theta, E) = I(\mathbf{x}, \theta, E) A_E(\mathbf{x}, \mathbf{x} + T\theta) = g_0(E, \mathbf{x} + T\theta, \mathbf{x}) \quad (1)$$

with $A_E(\mathbf{x}, \mathbf{x} + T\theta) = \exp\left(-\int_0^T \mu_E(\mathbf{x} + t\theta) dt\right)$

which quantifies the loss of intensity $I(\mathbf{x}, \theta, E)$ of the photon beam by the exponential of the total attenuation along the travel path between $\mathbf{x}$ and $\mathbf{x} + T\theta$.

For instance, in a fan-beam CT-scan, an X-ray tube illuminates a target with a given intensity. A set of detectors located outside the target, typically on an annulus, will collect the incoming photon flux and characterize the attenuation map of the medium according to the Beer–Lambert law Equation (1). Afterwards, the general (inverse) problem consists in reconstructing the attenuation map μ_E from the measured beam intensities for different angular views. We refer to [1,2] for a general review on reconstruction algorithms for imaging techniques. Since the advent of CT, many imaging concepts have emerged and the need in imaging has grown. One can mention Single Photon Emission CT (SPECT), Positron Emission Tomography (PET) and also Cone-Beam CT for the imaging systems based on an ionizing source.

Due to technological limitations, the factor energy has first been ignored as a potential parameter for an imaging system. However, the development of crystals and detector fabrication technologies enabling the collection of photons and separating them by range of energies has opened the way to many more imaging architectures capable of enhancing the image quality, optimizing the acquisition process or compensating for some limitations (such as limited angle issues), see [3–10]. Furthermore, the use of high-energy X-rays has increased over the years and shows interesting properties for industrial applications, see [11–13]. In order to provide a reliable information for an imaging system based on Compton scattering, the crystals should have:

- Good energy resolution;
- High density, high atomic number which enhances the probability for photoelectric absorption;
- No intrinsic activity in the measurement range (355 keV, 1332 keV) which would create a disturbance in the measured distribution;
- Small size and compactness.

In this communication, we will consider the example of LBC:Ce scintillators, which provides a better energy resolution than CeBr₃ scintillators and estimates their impact on our approach. We refer to [14,15] for an exhaustive study on CeBr₃. LBC:Ce scintillation crystals can now be produced with 5 cm diameter, 5 cm thickness and achieve about 2.8% FWHM at 662 keV [16,17]. Furthermore, one can note that LBC:Ce is self-radioactive but only in the range from 1550 keV to 2250 keV [18] which is over the measurement range (355 keV, 1332 keV) considered here. Table 1 delivers the characteristics of LBC:Ce and CeBr₃ crystals which motivates the use of LBC:Ce crystals for our application, see [17]. Regarding the density of the crystals, we should properly distinguish between crystals and scanning objects. Indeed, both object and crystal will scatter the radiation to some extent. Since this study focuses on the scattering induced by the object, the crystals are assumed to perfectly absorb the scattered radiation.

Table 1. Characteristics of LBC:Ce and CeBr₃ crystals.

	LBC:Ce	CeBr ₃
Emission Wavelength [nm]	380	380
Energy resolution at 662 keV [% FWHM]	2.8	3.8
Light Yield [photons/MeV]	75,000	60,000
Decay Time [ns]	<25	19
Density [g/cm ³]	≈5	5.1
Atomic Number (effective)	44 ... 45.2	45.9
Intrinsic Activity [Bq/cm ³] in range [1550 keV, 2250 keV]	≈1.5	<0.002
hygroscopic	yes	yes

The physical interactions between photons and matter can be distinguished into: Thomson–Rayleigh scattering, photoelectric absorption, Compton scattering and pair production. In the classic range of applications of the X-rays or γ -rays, i.e., >60 keV, the photoelectric absorption and the Compton scattering are the dominant phenomena, see [19]. While the photoelectric absorption plays an important role in the attenuation of the photon beam, a measured photon either suffers no interaction (primary radiation) or is scattered (scattered radiation); this is the reason why the Compton scattering is more natural to exploit than the photoelectric effect as an imaging agent.

The Compton effect stands for the collision of a photon with an electron. The photon transfers a part of its energy E_0 to the electron which experiences a recoil while the photon

is scattered by an (scattering) angle ω with the axis of propagation, see Figure 1. The energy of the photon after scattering is expressed by the Compton formula [20],

$$E_1 = \frac{E_0}{1 + \frac{E_0}{mc^2}(1 - \cos \omega)} = \mathcal{E}(E_0, \omega), \quad (2)$$

where $mc^2 = 511$ keV represents the energy of an electron at rest.

Figure 1. Geometry of the Compton effect: the incident photon energy E_0 yields a part of its energy to an electron and is scattered with an angle ω .

Usually energy resolved cameras are collimated and combined with polychromatic sources, leading to multi-channel CT-data and corresponding estimations of the attenuation map. Here, we intend to focus instead on the phenomenon of Compton scattering which enables a modeling of the energetic data in terms of electron density and for corresponding emission energies, see [21–36]. The purpose of this communication is to validate the possibility to use scintillation crystals and their energetic sensitivity in a fan-beam CT scan from a bichromatic ionizing source (here the Cobalt-60). To achieve this, the modeling and the reconstruction strategy are presented in Section 2. Section 3 presents the architecture and properties of the designed scanner with synthetic data as well as reconstructions. The validation of our approach with the properties of the LBC:Ce scintillators motivates the potential of the imaging system in particular with better resolved crystals or semiconductor detectors such as CZT (CdZnTe) [37].

2. Modeling and Processing of the Compton Scattered Data

We first assume the source to be monochromatic, i.e., it emits photons with same energy E_0 , for the sake of clarity. For sufficiently large E_0 , larger than 80 keV in medical applications for instance, the Compton effect represents a substantial part of the radiation as more than 70% of the emitted radiation is scattered within the whole body. As depicted by Figure 2, the Compton effect ruled by Equation (2) brings an interesting diversity into the measured spectrum as a monochromatic source leads to a polychromatic radiation measured by scintillation crystals. The development of scintillation crystals such as LBC:Ce and the improvement of their energy resolution delivers a new dimension to explore.

Neglecting Thomson–Rayleigh scattering and the pair production, we focus on the Compton scattering to interpret our data and decompose the spectrum $\text{Spec}(E, \mathbf{s}, \mathbf{d})$ measured at a detector $\mathbf{d}$ with energy E as follows

$$\text{Spec}(E, \mathbf{s}, \mathbf{d}) = \sum_{n=0}^{\infty} g_n(E_0, E, \mathbf{s}, \mathbf{d}). \quad (3)$$

In this equation, g_0 represents the primary radiation which crossed the object without being subject to the Compton effect. It corresponds to the signal measured in conventional

CT, Equation (1). The functions $g_n(E, \mathbf{s}, \mathbf{d})$ correspond to the photons that were measured at $\mathbf{d}$ with incoming energy E after n scattering events. The schematic curve with marks $\circ$ depicted in Figure 2 (right), often called Compton scattered spectrum, is typical from what one observes when measuring the spectrum of a monochromatic source, see for instance [38]. We refer to [39] for a detailed review on the technology of detection.

Figure 2. (Left) Illustration of the multiple scattering—the detector $\mathbf{d}$ measures photons of energy E that have not been scattered (primary) as well as scattered at different orders (here 1, 2 and 3). (Right) a schematic spectrum of the detector ($\circ$) is illustrated via its decomposition in terms of type of scattered data and depicts the response to a monochromatic source due to the Compton effect.

In the following, we study how the scattered radiation g_n , $n = 1, 2, \dots$ behaves and explain how to deal with it. To simplify the notations, we denote $E_{n+1} = \mathcal{E}(E_n, \omega_{n+1})$ with E_n the scattered energy after n scattering events and ω_{n+1} the $(n + 1)$ th scattering angle.

2.1. The Scattered Flux

We first focus on the first-order scattering as depicted in Figures 1 and 3 for the notations. The travel of a scattered photon beam follows the same scheme:

- The photon beam is emitted by the point source $\mathbf{s}$ in a differential solid angle $d\Omega_c$ with an energy E_0 ;
- Photons may be absorbed or scattered along the path following the Beer–Lambert law (Equation (1)) leading to the attenuation of the beam by the weight $A_{E_0}(\mathbf{s}, \mathbf{x})$;
- A part of the beam may collide with an electron at position $\mathbf{x}_1$ which belongs to a differential volume $d\mathbf{x}$;
- Due to Compton scattering, the photon is scattered by an angle ω_1 within a solid angle $d\Omega_c$ which follows the Klein–Nishina probability $\left(\frac{d\sigma}{d\Omega}\right)_{\omega_1}$ ([40]);
- Again the scattered beam is attenuated by a weight $A_{E_1}(\mathbf{x}, \mathbf{d})$ due to absorption or scattering of higher order;
- The attenuated and scattered beam finally reaches the point detector at $\mathbf{d}$.

This scheme leads to model the variation of the number of photons g_1 scattered at $\mathbf{x}$ and detected at $\mathbf{d}$ with energy E_1 by

$$\frac{d^3 g_1(\mathbf{x}, \mathbf{d}, \mathbf{s})}{d\mathbf{x} d\Omega_c^2} = I(\mathbf{s}, \mathbf{x} - \mathbf{s}; E_0) \left(\frac{d\sigma}{d\Omega} \right)_{\omega_1} A_{E_0}(\mathbf{s}, \mathbf{x}) A_{E_1}(\mathbf{x}, \mathbf{d}) n_e(\mathbf{x}). \quad (4)$$

The second variable in the intensity I allows the source to be anisotropic. This formula describes the evolution of the first scattered radiation which is detected at a given energy and at a given detector position.

Figure 3. Geometry of the Compton effect for two successive scattering events.

Regarding the first-order scattering, the Compton Formula (2) relates the energy to the scattering angle and thus delivers a unique geometry for each energy measured. Indeed, all scattering points $\mathbf{x}$ responsible for a detected scattered photon at energy E_1 belong to

$$\mathfrak{T}(\omega_1, \mathbf{s}, \mathbf{d}) = \left\{ \mathbf{x} \in \mathbb{R}^3 : \frac{\mathbf{x} - \mathbf{s}}{\|\mathbf{x} - \mathbf{s}\|} \cdot \frac{\mathbf{d} - \mathbf{x}}{\|\mathbf{d} - \mathbf{x}\|} = \cos \omega_1 \right\}. \quad (5)$$

$\mathfrak{T}(\omega_1, \mathbf{s}, \mathbf{d})$ defines in 3D the lemon part of a spindle torus ($\omega_1 < \pi/2$) and the apple part of a spindle torus ($\omega_1 > \pi/2$). In 2D, this reduces to two circular-arcs as depicted in Figure 4a. Figure 4b shows the scanning of a specimen by paired circular-arcs for different energy levels or equivalently different scattering angles ω_1 .

Figure 4. (a) Two scattering events for a triplet $(\omega, \mathbf{s}, \mathbf{d})$ belongs to the same spindle torus (3D) and pair of circular-arcs (2D). (b) Illustration of the scanning by circular-arcs for various energy levels and for $(\mathbf{s}, \mathbf{d})$ fixed.

An integral modeling of g_1 can then be obtained by integrating Equation (4) over the whole domain of the specimen and leads to a generalized Radon transform along $\mathfrak{T}(\omega_1, \mathbf{s}, \mathbf{d})$, noted $\mathcal{C}_1(n_e)$,

$$\mathcal{C}_1(n_e)(E_0, E_1, \mathbf{s}, \mathbf{d}) = \int_{\mathfrak{T}(\omega_1, \mathbf{s}, \mathbf{d})} I(\mathbf{s}, \mathbf{x} - \mathbf{s}; E_0) \left(\frac{d\sigma}{d\Omega} \right)_{\omega_1} A_{E_0}(\mathbf{s}, \mathbf{x}) A_{E_1}(\mathbf{x}, \mathbf{d}) n_e(\mathbf{x}) d\mathbf{x}. \quad (6)$$

We refer to [41,42] for more details.

As explained above, the spectrum cannot be reduced to the first-order scattering and multiple scattering must be taken into account. However, the question of its modeling arises. Physically, the same reasoning as for the first-order scattering can be iterated. The second-order scattering is illustrated in Figure 3. The computation of the successive scattering events leads thus to

$$\frac{d^{2n+1}g_n(\mathbf{x}_0, \dots, \mathbf{x}_{n+1})}{d\mathbf{x}_1 \dots d\mathbf{x}_n d\Omega_c^{n+1}} = I(\mathbf{x}_0, \mathbf{x}_1 - \mathbf{x}_0; E_0) \prod_{i=1}^n \left(\frac{d\sigma}{d\Omega} \right)_{\omega_i} A_{E_{i-1}}(\mathbf{x}_{i-1}, \mathbf{x}_i) A_{E_i}(\mathbf{x}_i, \mathbf{x}_{i+1}) n_e(\mathbf{x}_i), \quad (7)$$

with $\mathbf{x}_0 = \mathbf{s}$ and $\mathbf{x}_{n+1} = \mathbf{d}$. The components of the spectral data are then the functions g_n obtained by integrating Equation (7) over the support of the object. The difficulty here is the interpretation of this variation as an operator $\mathcal{C}_n(n_e)$ due to the complicated relations between successive scattering angles and the measured energy. This work was done for $n = 2$ in [41] for the 3D case and in [42] for the 2D case. The model is similar to $\mathcal{C}_1$ but with a trickier integration support and a quadratic dependency on the electron density.

2.2. Extension to Polychromatic γ -ray Sources

Monochromatic sources are of course difficult to produce and extremely expensive. While X-ray tubes produce a wide spectrum with few characteristic peaks, γ -ray sources emit essentially photons with energies at their characteristic peaks. In both cases, the spectrum of an ionizing source can be decomposed into

$$T(E) = T_{flat}(E) + \sum_{i=1}^N T_{peak}^{(i)}(E) \quad (8)$$

with T_{flat} the smooth and wide part of the spectrum and $T_{peak}^{(i)}$ the characteristic peaks. Including this information into the scattered spectrum in Equation (3) leads to

$$\text{Spec}_T(E, \mathbf{s}, \mathbf{d}) = \sum_{n=0}^{\infty} \int_0^{\infty} T(E_0) g_n(E_0, E, \mathbf{s}, \mathbf{d}) dE_0. \quad (9)$$

In this work, we consider the widely used Cobalt-60 which produces two characteristic peaks at $E_0^{(1)} = 1.173$ MeV and $E_0^{(2)} = 1.333$ MeV. In that case, one can interpret the measured spectral data after scattering by

$$\text{Spec}_{Co60}(E, \mathbf{s}, \mathbf{d}) = \sum_{n=0}^{\infty} \mathcal{C}_n(n_e)(E_0^{(1)}, E, \mathbf{s}, \mathbf{d}) + \mathcal{C}_n(n_e)(E_0^{(2)}, E, \mathbf{s}, \mathbf{d}) \quad (10)$$

using the integral modeling for g_n .

2.3. Modeling the Impact of the Scintillation Crystals

For ^{60}Co sources, LBC:Ce crystals demonstrates a 2.8% FWHM at 662 keV, 2.1% FWHM at 1.173 MeV and 2% FWHM at 1.332 MeV. This resolution as well as its diameter 5cm has a huge impact on our modeling and on the quality of the scattered data. Since an exact estimation of the point spread function in our model is extremely difficult due to the complexity of scattering inside the crystals, it can be relevant to simplify this model to a convolution, noted $*$, with a Gaussian normal distribution

$$G_{\sigma}(E) = \frac{1}{\sqrt{2\pi\sigma^2}} \exp\left(\frac{-|E|^2}{2\sigma^2}\right) \quad (11)$$

with standard deviation satisfying for LBC:Ce crystals

$$\frac{\sigma}{\text{keV}} = 0.306 \sqrt{\frac{E}{\text{keV}}}. \quad (12)$$

In addition, the size, shape and potential capsule of the crystal, noted here $\mathbf{Cr}(\mathbf{d})$, can have an important impact on the resolution and the measurement, see Figure 5. Indeed, our models $\mathcal{C}_n$ assume the crystal to be a point while $\mathbf{Cr}(\mathbf{d})$ is in practice a cylinder. Assuming that each element of a crystal behaves as a detector point, our models can thus be adapted by considering the transformation

$$\mathcal{D}g(E, \mathbf{s}, \mathbf{d}) = \int_{\mathbf{d}' \in \mathbf{Cr}(\mathbf{d})} g(E, \mathbf{d}', \mathbf{s}) d\mathbf{d}'. \quad (13)$$

Figure 5. The size and shape of the crystal implies the integration along a beam of circular-arcs and not only one pair (when the crystal is assumed to be a point) anymore.

We note that depending on the shape of the specimen under study and on the desired resolution of the reconstructed image, the operator $\mathcal{D}$ can be negligible in comparison with the limitations of the energy resolution and the convolution with G_σ . We note also that the *thickness* of scanning depicted in Figure 5, which affects the final resolution of the reconstructed image, corresponds to between a half and a third of the crystal depth. This is the reason why its impact may be negligible if the achieved image resolution is slightly smaller than the crystal size. This consideration shall be studied more in detail in further studies. Taking into account the stochastic nature of the emission of photons, we obtain a final model

$$\text{Spec}_{C_{060}}^{LBC}(E, \mathbf{s}, \mathbf{d}) = \text{Pois}(\mathcal{D}\text{Spec}_{C_{060}}(\cdot, \mathbf{s}, \mathbf{d}) * G_\sigma)(E, \mathbf{s}, \mathbf{d}) \quad (14)$$

in which $\text{Pois}(\cdot)$ stands for the Poisson distribution and with $\text{Spec}_{C_{060}}$ given in Equation (10).

2.4. A General Approach for the Reconstruction of the Electron Density

In order to exploit the scattered spectrum as an imaging agent, we need to find n_e from $\text{Spec}_{C_{060}}^{LBC}$ in Equation (14).

The main issue for solving this inverse problem is the complexity of $\mathcal{C}_n$, $n > 1$, in terms of modelling and computation time. The use of neural networks could circumvent this complexity but the lack of database for such scattered data prevents this approach. The use of standard optimization techniques such as the ROF model for total-variation regularization [43] requires however an accurate knowledge and an efficient computation of the forward model. The only suitable model for such standard approaches in our case is the first-order scattered radiation modelling, $\mathcal{C}_1(n_e)$.

Following on from [41,42], the first-order scattered intensity $\mathcal{C}_1(n_e)$ is less smooth than the multiple-order scattered intensity $\mathcal{C}_n(n_e)$, $n \geq 2$. Indeed, the complexity of multiple

scattering tends to smooth the information about the electron density n_e and thus the first-order scattering constitutes the most reliable information in the Compton spectrum. Relying upon theoretical results, this observation implies enhancing the variations of the data in order to reduce the part of multiple scattering. This can be done by considering the following problem:

$$\text{find } n_e \text{ from } \partial_E \text{Spec}_{C_{60}}^{LBC} \approx \partial_E \mathcal{C}_1(n_e) \quad (15)$$

in which ∂_E denotes the derivative with respect to the energy. Following the ROF-model [43], a solution can be constructed by solving the minimization problem

$$\min_{n_e} \frac{1}{2} \left\| \partial_E \text{Spec}_{C_{60}}^{LBC} - \partial_E \mathcal{C}_1(n_e) \right\|^2 + \lambda TV(n_e) \quad (16)$$

where $TV(n_e)$ stands for the total-variation of n_e . We consider this reconstruction method in the following section.

3. A CT-CST Scanner

We propose combining the standard fan-beam CT scanner with CST, see Figure 6. The architecture is identical with the exception that an X-ray tube and the CCD cameras are replaced respectively by a γ -ray source (60Co) and scintillation crystals (LBC:Ce). Taking advantage of the energetical variable brought by scintillators, we measure only 12 projections. Each projection is obtained by a 30° rotation of the scanner or of the specimen.

In our setting, 40 crystals are placed on the half circle opposite the source position. For simplification of the mathematical and numerical model, every detector as well as the source is modeled as a point of no space expansion. The imaged area is a square centered in the architecture of size $150 \times 150 \text{ cm}^2$, see Figure 6a. The complete scanning process is depicted in Figure 6b. Furthermore, the source emits 10^6 photons per projection.

Figure 6. (a) Scanning architecture for a single source, the radius of the circle passing through detectors and source is set to 1.5 m. (b) A depiction of the sampling scheme with fan-beam geometry, the source is rotated around the object and acquires data from different sides.

We assume that the energy of the photons is measured in the energy range [355 keV, 1332 keV]. The lower bound is hereby chosen as the energy of a photon that has initial energy 1173 keV and is scattered once, changing its trajectory angle by $\pi/2$, see Equation (2). The energy range is sampled equidistantly with 128 measured levels, leading to a necessary sampling of 7.68 keV. In practice, the energy resolution of scintillation crystals is not constant; however, this information can be simply incorporated in our approach and thus we consider a constant energy resolution for the sake of simplicity.

3.1. Ballistic Data

For the simulations, we consider a simple toy object made of an annulus of aluminium with a rectangle of polyethylene inside, see Figure 7a. The details are given in the following Table 2:

Table 2. Material and dimensions of the specimen and of the scintillator.

	Annulus	Rectangle	Crystal
Surface	Aluminium	Polyethylene	LBC:Ce
Thickness	53 cm outer/45 cm inner radius 2 cm	$45 \times 38 \text{ cm}^2$ 2 cm	$5 \times 5 \text{ cm}^2$ 5 cm

The ballistic part of the spectral data, g_0 defined in Equation (1), corresponds to the standard measure in CT. Due to the scintillation crystals, the primary measurement suffers also the energy resolution of the LBC:Ce and thus can be modeled by $G_\sigma * g_0$.

Obtained by the standard TV regularization method [43], the average of the reconstructions of the attenuation map μ_E at energies 1173 and 1333 keV is depicted in Figure 7b. The reconstruction suffers many artifacts which can be explained by: the energy resolution of the crystals and the very limited number of data (12 views and 40 detectors). However, this provides very important information for us. Indeed, at the energy range considered here, [355 keV, 1332 keV], the attenuation map is essentially proportional to the electron density, see [19],

$$\mu_E(\mathbf{x}) \approx \sigma_c(E)n_e(\mathbf{x}) \quad (17)$$

in which $\sigma_c(E)$ stands for the Compton total-cross section at energy E . Thereby, the ballistic data g_0 delivers us a first approximation for the sought-for electron density and can help us to compute a more accurate model $\mathcal{C}_1(n_e)$.

Figure 7. (a) Original phantom: (b) Reconstruction of the attenuation map from the ballistic data.

3.2. Results Based on the Compton Scattered Spectrum

Due to the nature of scattering, the complete architecture has to be three dimensional. As an illustration for our general approach, we considered the much simpler fan-beam geometry CT-scan with a 2 cm thickness. This way the geometry of 3D-scattering can be approximated by a 2D-geometry. Therefore, the torus $\mathfrak{T}$, see Equation (5), simplifies to two circular-arcs in this setting.

The scattered data for one projection view is depicted in Figure 8c. The spectrum is here essentially composed of the first-order and second-order scattered radiation, Figure 8a,b, respectively.

Figure 8. (a) First-order scattering for one view; (b) Second-order scattering for one view; (c) Compton spectrum for one view. (d–g) depict the profiles for detector position $\mathbf{d}_1$, $\mathbf{d}_2$, $\mathbf{d}_3$, $\mathbf{d}_4$ respectively.

In order to emphasize the general reconstruction strategy described in Section 2.4, we consider the profiles of the measured spectrum in Figure 8 for four different detector positions $\mathbf{d}_1$, $\mathbf{d}_2$, $\mathbf{d}_3$, $\mathbf{d}_4$ making an angle between source and center of the circle of -20 , -10 , 0 and 10 degrees respectively. We observe the two characteristic peaks of ^{60}Co and the Compton spectrum composed of first-order and multiple scattering. As demonstrated in [41,42], it is possible to exploit the smoothness properties of the different components of the spectrum. This is done simply by a discrete derivative applied on the spectral data. The derivative of the spectrum for one view and one detector is given in Figure 9. We observe that the differentiation step leads to reducing the part of multiple scattering and highlights the first-order scattering, which has the most suited structure for standard reconstruction methods. Using standard TV-regularization as for the ballistic data, we

propose solving the problem (15). The result is depicted in Figure 10b. The reconstruction is satisfactory but has a relative poor accuracy. While the modeling of the crystals has a huge impact on this reconstruction, a more theoretical aspect has to be considered. Indeed, the attenuation factors A_{E_0} and A_{E_1} are the most harmful physical factors of the inverse problem as they increase dramatically the ill-posedness or ill-conditioning of the method, i.e., its instability. While this instability is controlled by the TV-regularization, it leads to a reduced quality of reconstruction. To illustrate this aspect, we consider also reducing the complete system scanner/specimen by a factor of 66%, 50% and 40%. The results are depicted in Figure 10b–d. By downsizing the object, we shorten the length of the photon travel and thus diminish the attenuation weights. As measured by the errors in Table 3, the reconstruction quality increases significantly when we reduce the size of the object. However, it also implies that the crystal size is reduced by the same factor.

Figure 9. (a) Figure 8f without the ballistic contribution; (b) Derivatives of (a).

Figure 10. Reconstruction of the electron density for a scanner of size (a) 150 cm; (b) 100 cm; (c) 75 cm and (d) 60 cm.

Table 3. Peak signal-to-noise ratio and mean square errors of the reconstructions in Figure 10.

	150 cm	100 cm	75 cm	60 cm
PSNR	13.69 dB	18.02 dB	19.16 dB	19.62 dB
MSE	4.28%	1.58%	1.21%	1.09%

4. Conclusions and Discussion

This work illustrates the potential of scintillation crystals in the novel Compton scattering tomography. The Compton spectrum is decomposed into multiple scattering events of order n and modeled by integral transforms. The proposed reconstruction method takes advantage of the smoothness properties of the n th-order scattered radiation by applying standard techniques combined with a differentiation step. As revealed by the simulations, the ill-posedness induced by the attenuation factors and the energy resolution of the crystals has a substantial impact on the quality and accuracy of reconstructions. To address this issue, the development of more flexible and suited reconstruction methods is the core of our future research.

A limitation of this study is the 2D aspect of the simulations. Indeed, the nature of Compton scattering is three dimensional; this is why the simulations made here will be extended to the 3D case in order to reveal the full potential of the approach. In addition, we considered that $\approx 10^7$ photons were emitted by the γ -ray source for the total of views and were held into the slice of thickness 2 cm. In a real system, many more photons would be required in order to obtain a signal-to-noise ratio satisfactory in the data.

Furthermore, while the reconstruction of the electron density is shown possible, the resolution of reconstruction depends proportionally on the crystal size. Therefore, based on this preliminary work, the relevance and reliability of CST scanners lies upon many technological challenges: (i) the production of scintillation crystals smaller than the millimeter in order to compete with conventional CT, (ii) with the smallest FWHM possible and (iii) the production of γ -ray sources able to emit a large number of photons (10^7 photons per second) would help the method produce fast and accurate images.

Author Contributions: Software, L.K.; Writing—original draft, L.K. and G.R. All authors have read and agreed to the published version of the manuscript.

Funding: This research was funded by the Deutsche Forschungsgemeinschaft (DFG) under the grant RI 2772/2-1.

Conflicts of Interest: The authors declare no conflict of interest.

Abbreviations

The following abbreviations are used in this manuscript:

CeBr ₃	Cerium Bromide
CCD	Charge-coupled device
CST	Compton scattering tomography
CT	Computerized tomography
LBC:Ce	Lanthanum BromoChloride (La(Br _x Cl _{1-x}) ₃ :Ce)
MSE	Mean square error
PSNR	Peak signal-to-noise ratio

References

1. Natterer, F. *The Mathematics of Computerized Tomography*; Classics in Mathematics; Society for Industrial and Applied Mathematics: New York, NY, USA, 2001.
2. Natterer, F.; Wübbeling, F. *Mathematical Methods in Image Reconstruction*; Classics in Mathematics; Society for Industrial and Applied Mathematics: New York, NY, USA, 2001.
3. Hounsfield, G. Computerized transverse axial scanning (tomography). I. Description of system. *Br. J. Radiol.* **1973**, *46*, 1016–1022. [[CrossRef](#)] [[PubMed](#)]

4. Alvarez, R.; Macovski, A. Energy-selective reconstructions in X-ray computerized tomography. *Phys. Med. Biol.* **1976**, *21*, 733–744. [CrossRef]
5. Primak, A.; Giraldo, J.R.; Liu, X.; Yu, L.; McCollough, C. Improved dual-energy material discrimination for dual-source CT by means of additional spectral filtration. *Med. Phys.* **2009**, *36*, 1359–1369. [CrossRef]
6. Shefer, E.; Altman, A.; Behling, R.; Goshen, R.; Gregorian, L.; Roterman, Y.; Uman, I.; Wainer, N.; Yagil, Y.; Zarchin, O. State of the Art of CT Detectors and Sources: A Literature Review. *Curr. Radiol. Rep.* **2013**, *1*, 76–91. [CrossRef]
7. Tracey, B.; Miller, E. Stabilizing dual-energy X-ray computed tomography reconstructions using patch-based regularization. *Inverse Probl.* **2015**, *31*, 05004. [CrossRef]
8. McCollough, C.; Leng, S.; Lifeng, Y.; Fletcher, J. Dual- and Multi-Energy CT: Principles, Technical Approaches, and Clinical Applications. *Radiology* **2015**, *276*, 637–653. [CrossRef] [PubMed]
9. Goo, H.; Goo, J. Dual-Energy CT: New Horizon in Medical Imaging. *Korean J. Radiol.* **2017**, *18*, 555–569. [CrossRef]
10. Fredenberg, E. Spectral and dual-energy X-ray imaging for medical applications. *Nucl. Instrum. Methods Phys. Res. Sect. A Accel. Spectrometers Detect. Assoc. Equip.* **2018**, *878*, 74–87. [CrossRef]
11. Tian, Y.; Shimazoe, K.; Yan, X.; Ueda, O.; Ishikura, T.; Fujiwara, T.; Uesaka, M.; Ohno, M.; Tomita, H.; Yoshihara, Y.; et al. High energy X-ray photon counting imaging using linear accelerator and silicon strip detectors. *Nucl. Instrum. Methods Phys. Res. Sect. A Accel. Spectrometers Detect. Assoc. Equip.* **2016**, *830*, 251–255. [CrossRef]
12. Izumi, S.; Kamata, S.; Satoh, K.; Miyai, H. High energy X-ray computed tomography for industrial applications. *IEEE Trans. Nucl. Sci.* **1993**, *40*, 158–161. [CrossRef]
13. Toyokawa, H.; Kanada, H.; Kaihori, T.; Koike, M.; Yamada, K. Application of high-energy photon CT system with laser-Compton scattering to non-destructive test. *IEEE Trans. Nucl. Sci.* **2008**, *55*, 3571–3578. [CrossRef]
14. Fraile, L.M.; Mach, H.; Vedia, V.; Olaizola, B.; Pazy, V.; Picado, E.; Udías, J.M. Fast timing study of a CeBr₃ crystal: Time resolution below 120 ps at 60Co energies. *Nucl. Instrum. Methods Phys. Res. Sect. A Accel. Spectrometers Detect. Assoc. Equip.* **2013**, *701*, 235–242. [CrossRef]
15. Otaka, Y.; Shimazoe, K.; Mitsuya, Y.; Uenomachi, M.; Seng, F.W.; Kamada, K.; Yoshikawa, A.; Sakuragi, S.; Binder, T.; Takahashi, H. Performance Evaluation of Liquid-Processed CeBr₃ Crystals Coupled With a Multipixel Photon Counter. *IEEE Trans. Nucl. Sci.* **2020**, *67*, 988–993. [CrossRef]
16. Quarati, F.; Dorenbos, P.; van der Biezen, J.; Owens, A.; Selle, M.; Parthier, L.; Schotanus, P. Scintillation and detection characteristics of high-sensitivity CeBr₃ gamma-ray spectrometers. *Nucl. Instrum. Methods Phys. Res. A* **2013**, *729*, 596–604. [CrossRef]
17. Hellma Materials GmbH. Radiation Detection Crystals. 2014. Available online: <https://www.hellma.com/en/crystalline-materials/radiation-detection-crystals/> (accessed on 26 May 2021).
18. Hellma Materials GmbH, Scionix Holland BV and Gonitec. Scintillation Properties of High-Resolution La(Br_xCl_{1-x})₃:Ce and High-Sensitivity CeBr₃. 2017. Available online: https://indico.cern.ch/event/388511/contributions/2612029/attachments/1522152/2378455/SCINT2017_211_Poster.pdf (accessed on 26 May 2021).
19. Stonestrom, J.P.; Alvarez, R.E.; Macovski, A. A framework for spectral artifact corrections in X-ray CT. *IEEE Trans. Biomed. Eng.* **1981**, *28*, 128–141. [CrossRef]
20. Compton, A.H. A quantum theory of the scattering of X-rays by light elements. *Phys. Rev.* **1923**, *21*, 483–502. [CrossRef]
21. Lale, P.G. The Examination of Internal Tissues, using Gamma-ray Scatter with a Possible Extension to Megavoltage Radiography. *Phys. Med. Biol.* **1959**, *4*, 159–167. [CrossRef]
22. Norton, S.J. Compton scattering tomography. *J. Appl. Phys.* **1994**, *76*, 2007–2015. [CrossRef]
23. Cesareo, R.; Borlino, C.C.; Brunetti, A.; Golosio, B.; Castellano, A. A simple scanner for Compton tomography. *Nucl. Instruments Methods Phys. Res. A* **2002**, *487*, 188–192. [CrossRef]
24. Adejumo, O.O.; Balogun, F.A.; Egbedokun, G.G.O. Developing a Compton scattering tomography system for soil studies: Theory. *J. Sustain. Dev. Environ. Prot.* **2011**, *1*, 73–81.
25. Anghaie, S.; Humphries, L.L.; Diaz, N.J. Material characterization and flaw detection, sizing, and location by the differential gamma scattering spectroscopy technique. Part1: Development of theoretical basis. *Nucl. Technol.* **1990**, *91*, 361–375. [CrossRef]
26. Balogun, F.A.; Cruvinel, P.E. Compton scattering tomography in soil compaction study. *Nucl. Instrum. Methods Phys. Res. A* **2003**, *505*, 502–507. [CrossRef]
27. Brunetti, A.; Cesareo, R.; Golosio, B.; Luciano, P.; Ruggero, A. Cork quality estimation by using Compton tomography. *Nucl. Instrum. Methods Phys. Res. B* **2002**, *196*, 161–168. [CrossRef]
28. Clarke, R.L.; Dyk, G.V. A new method for measurement of bone mineral content using both transmitted and scattered beams of gamma-rays. *Phys. Med. Biol.* **1973**, *18*, 532–539. [CrossRef] [PubMed]
29. Evans, B.L.; Martin, J.B.; Burggraf, L.W.; Roggemann, M.C. Nondestructive inspection using Compton scatter tomography. *IEEE Trans. Nucl. Sci.* **1998**, *45*, 950–956. [CrossRef]
30. Farmer, F.T.; Collins, M.P. A new approach to the determination of anatomical cross-sections of the body by Compton scattering of gamma-rays. *Phys. Med. Biol.* **1971**, *16*, 577–586. [CrossRef]
31. Gorshkov, V.A.; Kroening, M.; Anosov, Y.V.; Dorjgochoo, O. X-ray scattering tomography. *Nondestruct. Test. Eval.* **2005**, *20*, 147–157. [CrossRef]

-
32. Harding, G.; Harding, E. Compton scatter imaging: A tool for historical exploitation. *Appl. Radiat. Isot.* **2010**, *68*, 993–1005. [[CrossRef](#)] [[PubMed](#)]
 33. Meneley, D.A.; Hussein, E.M.A.; Banerjee, S. On the solution of the inverse problem of radiation scattering imaging. *Nucl. Sci. Eng.* **1986**, *92*, 341–349.
 34. Arendtsz, N.V.; Hussein, E.M.A. Energy-spectral Compton scatter Imaging—Part 1: Theory and mathematics. *IEEE Trans. Nucl. Sci.* **1995**, *42*, 2155–2165. [[CrossRef](#)]
 35. Guzzardi, R.; Licitra, G. A critical review of Compton imaging. *CRC Crit. Rev. Biomed. Imaging* **1988**, *15*, 237–268.
 36. Rigaud, G. Compton Scattering Tomography: Feature Reconstruction and Rotation-Free Modality. *SIAM J. Imaging Sci.* **2017**, *10*, 2217–2249. [[CrossRef](#)]
 37. Gurov, Y.B.; Egorov, N.Y.; Ponomarev, D.V.; Rozov, S.V.; Sandukovsky, V.G.; Shakhov, K.V.; Yakushev, E.A.; Zhivun, V.M. Study of characteristics of CdZnTe detector. *J. Instrum.* **2019**, *14*, P11002. [[CrossRef](#)]
 38. Negm, H.; Ohgaki, H.; Izuru, D.; Hayakawa, T.; Zen, H.; Kii, T.; Masuda, K.; Hori, T.; Hajima, R.; Shizuma, T.; Kikuzawa, N. Reaction-yield dependence of the (γ , γ') reaction of ^{238}U on the target thickness. *J. Nucl. Sci. Technol.* **2014**, 1–10. [[CrossRef](#)]
 39. Knoll, G.F. *Radiation Detection and Measurement*; John Wiley & Sons, Inc.: New York, NY, USA, 1989.
 40. Klein, O.; Nishina, Y. Über die Streuung von Strahlung durch freie Elektronen nach der neuen relativistischen Quantendynamik von Dirac. *Z. Phys.* **1929**, *52*, 853–869. [[CrossRef](#)]
 41. Rigaud, G. 3D Compton scattering imaging with multiple scattering: Analysis by FIO and contour reconstruction. *Inverse Probl.* **2021**. [[CrossRef](#)]
 42. Kuger, L.; Rigaud, G. Joint fan-beam CT and Compton scattering tomography: Analysis and image reconstruction. *arXiv* **2020**, arXiv:2008.06699.
 43. Rudin, L.; Osher, S.; Fatemi, E. Nonlinear total variation based noise removal algorithms. *Physica D* **1992**, *46*, 259–268. [[CrossRef](#)]