

Supplementary material

Dispersed Conducting Polymer Nanocomposites with Glucose Oxidase and Gold Nanoparticles for the Design of Enzymatic Glucose Biosensors

Natalija German^{1,2}, **Almira Ramanaviciene**^{1,2}, **Arunas Ramanavicius**^{3,4*}

¹ Department of Immunology, State Research Institute Centre for Innovative Medicine, Santariskiu 5, LT-08406 Vilnius, Lithuania; natalija.german@imcentras.lt (N.G.); almira.ramanaviciene@chf.vu.lt (AR)

² NanoTechnas—Center of Nanotechnology and Materials Science, Faculty of Chemistry and Geosciences, Vilnius University, LT-03225 Vilnius, Lithuania

³ Department of Physical Chemistry, Faculty of Chemistry and Geosciences, Vilnius University, Naugarduko 24, LT-03225 Vilnius, Lithuania

⁴ Division of Materials Science and Electronics, State Scientific Research Institute Center for Physical Sciences and Technology, Savanorių ave. 231, LT-02300 Vilnius, Lithuania

* Correspondence: arunas.ramanavicius@chf.vu.lt




Figure S1. The reproducibility of enzymatic biosensors based on GR electrode modified by PANI-AuNPs_(6 nm)-GOx/GOx for 19.7 mmol L⁻¹ of glucose concentration. Amperometric responses were registered in 0.05 mol L⁻¹ SA buffer, pH 6.0, with 0.01 mol L⁻¹ KCl and 6.0 mmol L⁻¹ PMS.