

Article

Methods Used for Endotracheal Tube Cuff Inflation and Pressure Verification in Veterinary Medicine: A Questionnaire on Current Practice

Iris Veen * and Janny C. de Grauw

Department of Clinical Sciences, Faculty of Veterinary Medicine, Utrecht University,
3584 CM Utrecht, The Netherlands

* Correspondence: i.veen@uu.nl

Simple Summary: Endotracheal intubation is the process whereby a tube is placed in the trachea during anaesthesia to assist breathing, administer anaesthetic gases and prevent fluids from entering the trachea. Endotracheal intubation is a routine procedure in veterinary anaesthesia, yet no guidelines exist for establishment of a sealed airway. Through an online survey of veterinary professionals who administer anaesthesia, we aimed to assess specific aspects of current practice of endotracheal intubation in veterinary medicine. The pressure in the cuff (a balloon at the end of the tube that, when inflated, contacts the trachea to seal off the airway) was measured by almost one-third of respondents in cats and dogs but by less than one-tenth of respondents in farm animals and horses. Respondents seemed to target a similar cuff inflation pressure, regardless of species, although a higher pressure was more often selected in horses compared to dogs, cats and farm animals. The preferred technique to verify cuff seal was the same in dogs, cats and farm animals, whereas in horses, a different technique was preferred. Cuff pressure measurement remains uncommon in veterinary anaesthesia. The development of recommendations for cuff inflation, including cuff pressure ranges for various species, can help to improve practice.

Citation: Veen, I.; Grauw, J.C.d. Methods Used for Endotracheal Tube Cuff Inflation and Pressure Verification in Veterinary Medicine: A Questionnaire on Current Practice. *Animals* **2022**, *12*, 3076. <https://doi.org/10.3390/ani12223076>

Academic Editors: Cinzia Benazzi and Gabrielle C. Musk

Received: 10 September 2022

Accepted: 4 November 2022

Published: 8 November 2022

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Copyright: © 2022 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

Abstract: Endotracheal intubation is a routine procedure in veterinary anaesthesia, yet no consensus guidelines exist for endotracheal tube (ETT) cuff inflation and pressure measurement. The aim of this study was to assess current practice of ETT cuff inflation and seal verification in veterinary medicine. An online questionnaire was distributed among veterinary professionals who administer anaesthesia, comprising six demographic and twelve ETT cuff-related questions per species. N = 348 questionnaires were completed. Cuff pressure was measured by 30% of respondents in cats, 32% in dogs and 9% in both farm animals and horses. Anaesthesia diplomates were not more likely to measure cuff pressure than others, except in cats (OR: 1.8; 95% CI: 1.1–2.9). The most frequently selected recommended range of cuff pressure was 20–30 cm H₂O, regardless of species, although >30 cm H₂O was selected significantly more often in horses compared to dogs, cats and farm animals. The preferred technique to verify cuff seal was minimal occlusive volume in dogs, cats and farm animals, whereas in horses, the preferred method was verification of normal capnogram waveform. ETT cuff pressure measurement remains uncommon in veterinary anaesthesia. The development of consensus recommendations for cuff inflation, including evidence-based target cuff pressure ranges for various species and different ETT models or materials, can help to improve practice.

Keywords: cuff pressure; endotracheal tube cuff; cuff pressure measurement; cuff inflation

1. Introduction

Endotracheal intubation is routinely performed in veterinary medicine; however, no consensus guidelines exist for endotracheal tube (ETT) cuff inflation and optimal target

range for cuff pressure. There may also be relevant differences between species and/or between ETT make, model or material.

In dogs, digital palpation of the pilot balloon and inflation to minimum (audible) occlusive volume proved ineffective to ensure inflation of the ETT cuff within the optimal pressure range [1]. In a similar study in cats, pilot balloon palpation, minimum occlusive volume and loss of resistance techniques all performed poorly at achieving predetermined ETT cuff pressures [2]. To the best of our knowledge, there is no available literature on cuff inflation techniques and their relative performance in horses and farm animals.

The recommended ETT cuff pressure range of 20–30 cm H₂O most commonly cited in the medical literature is based on decreases in tracheal capillary perfusion when ETT cuff pressures over 30 cm H₂O were used in rabbits [3] and humans [4]. However, this recommendation pertains to high-volume–low-pressure cuff on polyvinyl chloride (PVC) ETTs, which are frequently used for endotracheal intubation in humans and companion animals. In large animals, silicone ETTs fitted with low-volume–high-pressure cuffs are regularly used. These cuffs are elastic and known to have greater compliance than PVC cuffs [1,5]. As a result, no simple linear relation between intracuff pressure and tracheal transmural pressure exists for these tubes [6–8]. In horses, an ETT cuff pressure of more than 80 cm H₂O was required to provide a seal sufficient to prevent liquid leakage around the low-volume–high-pressure cuff silicone 30 mm ETT [9], which is much higher than the 20–30 cm H₂O reported for PVC ETTs.

As there are no veterinary guidelines and literature data are contradictory or incomplete, in this study, we sought to assess current practice for ETT cuff inflation and pressure verification in veterinary anaesthesia.

The first aims of the present study were to evaluate (I) common ETT material properties and reuse of ETTs in various species, (II) techniques used for ETT cuff inflation and cuff pressure measurement, (III) the target range of cuff pressure used, (IV) preferred techniques to establish ETT cuff seal and (V) the frequency and timing of rechecking ETT cuff seals. A secondary aim was to investigate the influence of training level (diplomate vs. first-opinion veterinarian) and years of experience on cuff inflation practice.

2. Materials and Methods

2.1. Questionnaire

An Internet-based questionnaire cross-sectional study was conducted via Facebook and through e-mail list servers. The questionnaire (Appendix A) was developed based on recent literature [10], expert opinion and the checklist for reporting results of Internet e-surveys [11]. Deviation from CHERRIES guidelines is mentioned where applicable. The questionnaire was piloted for usability and technical functionality among Utrecht University senior veterinary anaesthesia providers.

The final survey included six demographic questions about gender, country of work, job title, place of work (first-opinion practice, referral practice, university teaching hospital or other), years of veterinary experience and animal species routinely intubated (dogs, cats, farm animals and/or horses). Eight additional questions per species category covered the material of ETTs and reuse of ETTs, techniques used for ETT cuff inflation, preferred techniques used to establish an adequate tracheal seal and, if the ETT cuff seal is routinely rechecked, what the participant thinks the recommended range of cuff pressure is; finally, the participants were asked to prioritise items associated with ETT cuffing.

Adaptive questioning was used in two specific questions. If participants indicated a cuff pressure measurement device was used for ETT cuff inflation, follow-up questions asked which cuff pressure measurement device was used and what cuff pressure was targeted. If participants answered that the ETT cuff was rechecked, two further questions asked with respect to when and how. Therefore, in addition to the six demographic questions, respondents were asked eight to twelve questions per species category, depending on their answers.

Institutional ethical approval for this study was sought but ruled not to be required under Dutch legislation (documentation on file).

2.2. Distribution of the Questionnaire and Informed Consent

Study participants were recruited via a closed Facebook community for Dutch veterinarians (“Het Dierenartsengilde”), which is a private community for veterinarians only. Furthermore, study participants were recruited through the e-mail list servers of the European College of Veterinary Anaesthesia and Analgesia (ECVAA and the ACVA-L server).

The questionnaire was drafted and provided to all study participants in English. The survey remained open from the 9th of December to the 31st of December 2021 (22 days). An e-mail reminder to participate was sent 2 weeks after the opening invitation of the survey. The questionnaire was not password-protected and was developed and made available using Qualtrics XM online survey software.

Informed consent was obtained before the questionnaire could be started. Study participants were informed about the purpose of the study, that that data collection was anonymous and that they could opt out at any moment. Participation in the questionnaire was voluntary, and no incentives were offered for completion of the survey.

2.3. Target Population

The target population consisted of veterinarians (first-opinion practitioners, as well as residency-trained anaesthesiologists, including diplomates of the ECVAA and ACVAA) and veterinary anaesthesiology nurses/technicians who routinely perform endotracheal intubation for anaesthesia. Study participants who indicated that they do not regularly perform endotracheal intubation in dogs, cats, farm animals and/or horses were immediately redirected to the end of the survey.

The closed Facebook community for Dutch veterinarians (“Het Dierenartsengilde”) has a membership count of 3500 individual veterinarians, and the ECVAA and ACVA-L e-mail list servers comprise of an estimated 415 and 1000 members, respectively.

2.4. Data Handling

Survey completeness was checked for every respondent, and incomplete questionnaires were excluded from analysis. Software (IBM SPSS 27) was used to search for duplicate entries; none were detected. The final question on prioritising items associated with ETT cuffing (see Appendix A) was excluded from analysis post hoc, as participant feedback indicated it was not sufficiently intuitive how respondents could drag and drop the items to rank them from highest to lowest importance; therefore, when the original order of items appeared as the answer, we could not be sure whether or not this was intentional.

2.5. Statistical Analysis

Descriptive analysis was performed on answers to the demographic questions. The Shapiro–Wilk test was used to assess normality. A Wilcoxon signed-rank test was performed for comparison of differences between species for the same questionnaire item. Odds ratios were calculated to compare the odds for dichotomized respondent demographic categories diplomate/non-diplomate, first-opinion practice/non-first-opinion practice and >10 years of experience/≤10 years of experience. All statistical analyses were performed in IBM SPSS 27.

3. Results

A total of 425 people participated in the study, with 348 questionnaires fully completed and included for analysis. Owing to the complete anonymization of the questionnaire, it was not possible to calculate a view rate (ratio of unique survey visitors/unique site visitors). The participation rate (ratio of first survey page visitors/visitors who agreed to participate) and completion rate (ratio of users who finished the survey/users who agreed to participate) were 407/423 (96%) and 348/423 (82%), respectively [11]. Results and statistics are presented in Tables 1 and 2. A comprehensive overview of all odds ratios for respondent answers can be found in Appendix B.

Table 1. Response frequency table (percentage of total number of responses) for each species category and statistical significance per questionnaire item between species. Note: For some questionnaire items, study participants could select multiple options; therefore, the total is > 100%.

| Questionnaire Item | Species | | | |
|--|-------------------|-------------------|-------------------|--------------------|
| | Dogs | Cats | Farm Animals | Horses |
| What material is the endotracheal tube made of? | | | | |
| Silicone | 67.3 ‡¶ | 45.2 ‡§¶ | 85.6 ‡¶ | 91.3 ‡‡§ |
| Rubber | 7.7 ‡ | 4.6 ‡ | 13.6 | 12.6 |
| PVC | 55.5 §¶ | 56.0 §¶ | 48.3 ‡‡¶ | 10.2 ‡‡§ |
| Don't know | 16.2 | 16.0 | 0 | 0.8 |
| Non-selected | 2.7 § | 0.6 | 5.1 ⁺ | 2.4 |
| Do you reuse the endotracheal tube? | | | | |
| Yes | 76.4 §¶ | 75.4 §¶ | 89.0 ‡‡ | 96.1 ‡‡ |
| No | 20.4 | 22.8 §¶ | 3.4 ‡ | 1.6 ‡ |
| Other | 3.2 §¶ | 1.5 | 3.4 ‡¶ | 0 ‡§ |
| Non-selected | 0 § | 0.3 § | 4.2 ‡‡ | 2.4 |
| What method or methods do you use for cuff inflation? | | | | |
| Syringe filled with air | 90.6 ‡ | 87.4 ‡ | 90.7 | 91.3 |
| Syringe filled with fluid | 0.9 | 0.9 | 0 | 1.6 |
| Cuff pressure measurement device | 30.4 | 31.7 | 8.5 ‡‡ | 9.4 ‡‡ |
| Other | 8.7 | 0 | 0 | 0 |
| Non-selected | 0 § | 0.9 | 5.1 ⁺ | 2.4 |
| How do you measure cuff pressure? | | | | |
| Manometer | 35.9 §¶ | 30.1 § | 60.0 ‡‡ | 100.0 ⁺ |
| Tru-Cuff [®] | 10.7 §¶ | 11.7 §¶ | 0 ‡‡ | 0 ‡‡ |
| AG Cuffill [®] | 35.0 ‡§¶ | 41.7 ‡§¶ | 10.0 ‡‡ | 0 ‡‡ |
| PressureEasy [®] | 23.3 ‡ | 18.4 ‡ | 0 | 0 |
| Other | 11.7 §¶ | 11.7 § | 10.0 ‡‡ | 0 ⁺ |
| Non-selected | 1.0 | 0 | 20.0 | 0 |
| What is the cuff pressure you use? (if using measurement device) | | | | |
| Don't know | 7.5 | 10.7 | 0 | 0 |
| <20 cm H2O | 3.0 | 12 | 12.5 | 0 |
| 20–30 cm H2O | 71.6 | 62.7 | 62.5 | 18.2 |
| >30 cm H2O | 14.9 | 14.7 | 25.0 | 81.8 |
| Other | 3.0 | 0 | 0 | 0 |
| Non-selected | 0 | 0 | 0 | 0 |
| What do you think the recommended range of cuff pressure is? | | | | |
| No answer | 23.3 §¶ | 24.0 § | 16.1 ‡‡ | 18.1 ⁺ |
| Don't know | 15.6 § | 15.7 § | 5.9 ‡‡ | 3.1 |
| <20 cm H2O | 14.2 [*] | 17.5 [*] | 12.7 [*] | 7.9 [*] |
| 20–30 cm H2O | 31.6 § | 28.6 | 36.4 ⁺ | 38.6 |

| | | | | |
|--|----------|----------|----------|----------|
| >30 cm H2O | 10.3 §¶ | 7.7 §¶ | 23.7 †‡ | 26.0 †‡ |
| Other | 5.0 | 5.5 | 5.1 | 6.3 |
| Non-selected | 0 | 0 | 0 | 0 |
| What techniques do you use to establish a good seal of the cuff? | | | | |
| Minimal Occlusive Volume | 35.8 §¶ | 38.0 §¶ | 27.5 †‡¶ | 17.6 †‡§ |
| Absence smell of inhalational anaesthetic | 6.2 §¶ | 7.4 §¶ | 11.7 †‡¶ | 15.0 †‡§ |
| Palpation pilot balloon | 14.5 | 14.1 | 16.1 | 13.4 |
| Loss of resistance | 1.9 §¶ | 0 † | 0 † | 0 † |
| Syringe pressure | 4.5 ¶ | 3.8 ¶ | 5.1 | 8.2 †‡ |
| Filling of the ventilator bellows | 3.9 §¶ | 3.7 §¶ | 12.4 †‡¶ | 17.6 †‡§ |
| Verification normal capnogram waveform | 20.7 † | 19.2 † | 20.9 | 19.9 |
| Measurement cuff pressure | 11.5 §¶ | 12.9 §¶ | 4.8 †‡ | 4.6 †‡ |
| Auscultation trachea | 1.1 | 1.0 | 1.5 ¶ | 3.6 § |
| Other | 4.4 | 4.0 | 3.4 | 5.5 |
| Non-selected | 0.9 †§ | 1.8 † | 7.6 † | 7.1 |
| Why do you use this technique? | | | | |
| No specific preference | 7.2 | 8.5 | 13.1 | 10.3 |
| By habit | 25.9 | 26.1 | 26.3 | 26.6 |
| I was taught to do it this way | 47.7 † | 47.1 † | 46.7 ¶ | 54.0 § |
| Literature | 13.3 §¶ | 13.3 §¶ | 7.9 †‡ | 5.5 †‡ |
| Other | 5.9 | 5.0 | 5.9 | 3.6 |
| Non-selected | 0 | 0 | 0 | 0 |
| Do you re-check the cuff? | | | | |
| No, never | 8.4 | 9.3 | 1.6 | 1.5 |
| Yes, sometimes | 48.9 | 49.9 | 46.7 | 53.8 |
| Yes, frequently | 19.0 †¶ | 15.5 †¶ | 21.3 | 11.4 †‡ |
| Yes, always | 8.3 | 8.4 | 13.2 | 12.1 |
| Only with specific indications | 15.3 §¶ | 16.9 ¶ | 17.2 † | 21.2 †‡ |
| Non-selected | 0 | 0 | 0 | 0 |
| When do you re-check the cuff after endotracheal intubation? | | | | |
| ≤15 min | 11.7 ¶ | 14.1 | 12.7 | 16.5 † |
| >15 to ≤30 min | 24.4 ¶ | 25.1 ¶ | 28.2 ¶ | 14.9 †‡§ |
| >30 min | 9.1 | 8.2 | 10.9 | 12.4 |
| Only with specific indications | 46.9 | 46.4 | 42.7 | 48.8 |
| Other | 7.8 § | 6.2 | 5.4† | 7.4 |
| Non-selected | 0 | 0 | 0 | 0 |
| What techniques do you use to establish a good seal of the cuff? | | | | |
| Minimal Occlusive Volume | 57.3 §¶ | 94.3 §¶ | 53.4 †‡¶ | 36.0 †‡§ |
| Absence smell of inhalational anaesthetic | 30.1 §¶ | 29.7 §¶ | 38.8 †‡ | 40.0 †‡ |
| Palpation pilot balloon | 33.3 | 33.1 | 36.2 | 31.2 |
| Loss of resistance | 3.6 §¶ | 2.0 §¶ | 14.7 †‡ | 18.4 †‡ |
| Syringe pressure | 6.8 ¶ | 4.8 | 10.3 | 11.2 † |
| Filling of the ventilator bellows | 20.1 * | 16.4 * | 42.2 * | 52.8 * |
| Verification normal capnogram waveform | 52.4 † | 48.8 † | 50.9 | 56.0 |
| Measurement cuff pressure | 20.1 †§¶ | 26.6 †§¶ | 12.1 †‡ | 8.8 †‡ |
| Auscultation trachea | 2.3 | 2.7 | 3.4 | 5.6 |
| Other | 5.5 | 4.4 | 2.6 ¶ | 5.6 § |
| Non-selected | 9.7 | 10.5 | 6.8 | 4.7 |

PVC = polyvinyl chloride; * significant difference between all species within the same questionnaire item ($p < 0.05$); † significant difference from dogs within the same questionnaire item ($p < 0.05$); ‡ significant difference from cats within the same questionnaire item ($p < 0.05$); § significant difference

from farm animals within the same questionnaire item ($p < 0.05$); ¶ significant difference from horses within the same questionnaire item ($p < 0.05$).

3.1. Demographics

Comprehensive demographic details of the study participants can be found in Appendix C. Almost half of the study participants were first-opinion veterinarians, with 95% practicing in the Netherlands; only 2% of respondents were anaesthesiology nurses/technicians. Nearly 10% of the study participants were anaesthesia residents, whereas ACVAA and ECVAAs comprised of 14 and 18% of respondents, respectively. Respondents worked in first-opinion practice (43%), referral practice (22%) and university teaching hospitals (32%).

Study participants indicated regularly performing endotracheal intubation in dogs (97% of participants), cats (93%), farm animals (34%) and horses (37%). Almost one-third (32%) of the study participants routinely performed endotracheal intubation in all four species categories. Diplomates were overrepresented in the group that intubated farm animals and horses (52% and 54%) compared to those intubating small animals (31%). Intubation of farm animals and horses was rarely performed by respondents working in first-opinion practice (2% in both).

3.2. Material and Reuse of Endotracheal Tubes

Silicone was the most commonly used ETT material in dogs, farm animals and horses, whereas in cats, PVC was the most frequently selected material for ETT. Overall, silicone was far more common as an ETT material type in farm animals and horses (91% and 86%) compared to dogs and cats (67% and 45%; see also Table 1).

In line with this result, reuse of ETTs in dogs and cats was significantly less common than in horses and farm animals. Working in first-opinion practice neither increased nor decreased the odds of re-using the ETT (Table 2).

Table 2. Odds ratio for frequency of responses for diplomates, first-opinion practitioners and respondents more than 10 years of experience for selected questionnaire items.

| | Diplomate | | | First-Opinion Practice | | | >10 Years of Experience | | |
|--|-------------------------|-------|-------|-------------------------|-------|--------|-------------------------|-------|-------|
| | 95% Confidence Interval | | | 95% Confidence Interval | | | 95% Confidence Interval | | |
| | Odds Ratio | Lower | Upper | Odds Ratio | Lower | Upper | Odds Ratio | Lower | Upper |
| Species | | | | | | | | | |
| Dogs | 0.897 | 0.220 | 3.657 | 2.648 | 0.542 | 12.933 | 0.364 | 0.075 | 1.780 |
| Cats | 1.031 | 0.411 | 2.583 | 0.959 | 0.409 | 2.252 | 1.460 | 0.626 | 3.406 |
| Farm Animals | 5.010 | 3.072 | 8.170 | 0.010 | 0.002 | 0.041 | 1.463 | 0.929 | 2.304 |
| Horses | 4.611 | 2.842 | 7.481 | 0.008 | 0.002 | 0.034 | 1.364 | 0.875 | 2.128 |
| Reuse of endotracheal tubes | | | | | | | | | |
| Dogs | 1.221 | 0.745 | 2.001 | 0.980 | 0.614 | 1.654 | 1.316 | 0.773 | 2.241 |
| Cats | 1.785 | 1.090 | 2.924 | 0.852 | 0.530 | 1.371 | 1.280 | 0.759 | 2.157 |
| Farm Animals | 1.059 | 0.270 | 4.159 | * | * | * | 0.588 | 0.059 | 5.850 |
| Horses | 0.917 | 0.279 | 3.011 | * | * | * | 1.027 | 0.990 | 1.065 |
| Measuring cuff pressure | | | | | | | | | |
| Dogs | 1.221 | 0.745 | 2.001 | 0.980 | 0.614 | 1.654 | 1.320 | 0.823 | 2.117 |
| Cats | 1.785 | 1.090 | 2.924 | 0.852 | 0.530 | 1.371 | 1.712 | 1.054 | 2.783 |
| Farm Animals | 1.059 | 0.270 | 4.159 | * | * | * | 0.882 | 0.235 | 3.316 |
| Horses | 0.917 | 0.279 | 3.011 | * | * | * | 0.868 | 0.259 | 2.903 |
| ≥3 techniques to verify cuff seal | | | | | | | | | |
| All species | 2.075 | 1.307 | 3.294 | 0.453 | 0.288 | 0.710 | 1.152 | 0.746 | 1.778 |

| Never rechecking the cuff | | | | | | | | | |
|---------------------------|-------|-------|-------|--------|-------|--------|-------|-------|--------|
| Dogs | ** | ** | ** | 10.238 | 3.486 | 30.063 | 0.303 | 0.134 | 0.683 |
| Cats | 0.062 | 0.008 | 0.463 | 8.854 | 3.313 | 23.662 | 0.547 | 0.262 | 1.142 |
| Farm Animals | ** | ** | ** | * | * | * | 0.955 | 0.895 | 1.018 |
| Horses | ** | ** | ** | * | * | * | 0.623 | 0.038 | 10.202 |

A table presenting all odds ratios calculated from the survey data can be found in Appendix B. * It is not possible to calculate a reliable odds ratio for farm animals and horses, owing to a very limited number of non-diplomate respondents in these species categories; ** it is not possible to calculate an odds ratio, as no diplomates answered that they never rechecked the cuff.

3.3. Methods Used for Cuff Inflation, Cuff Pressure Measurement and Target Range of Cuff Pressure Used

The most commonly selected preferred method for cuff inflation across species was a syringe filled with air (Table 1). Respondent answers indicated cuff pressure is measured in 30% of dogs, 32% of cats and 9% of farm animals and horses. Anaesthesia diplomates were not more likely to measure cuff pressure than others, except in cats (see also Table 2). Working in first-opinion practice did not affect the likelihood of measuring cuff pressure in dogs and cats (see also Table 2), whereas more than 10 years of work experience increased the likelihood of measuring cuff pressure in cats (see also Table 2). The most frequently used cuff pressure measurement device was an AG Cuffill® in cats and a handheld sphygmomanometer in the three other species groups. Among study participants indicating the use of a cuff pressure measurement device, 34% did not enter what target cuff pressure was used. After exclusion of these responses, the most commonly targeted cuff pressure range when using a cuff pressure measurement device was 20–30 cm H₂O in cats, dogs and farm animals. In horses, however, the most frequently selected targeted cuff pressure range was >30 cm H₂O. Although there was no significant difference in cuff pressure used between species, respondent answers with respect to their perception of the recommended range of cuff pressure were differed significantly between species.

3.4. Preferred Techniques Used to Establish Endotracheal Tube Cuff Seal

The preferred technique to verify cuff seal was minimal occlusive volume in dogs, cats and farm animals, whereas in horses, the preferred technique was verification of normal capnogram waveform (Table 1). Diplomates were more likely to use ≥3 different techniques to verify cuff seal than other respondents (Table 2).

Preference for a technique was largely dictated by training, as half of all respondents selected “I was taught to do it this way” when asked to explain their preference.

3.5. Frequency and Timing of Rechecking Endotracheal Tube Cuff Seal

The most frequently selected answer option for rechecking cuff seal was “Yes, sometimes” for all species (Table 1). Whereas the frequency of never rechecking cuff seal was not dependent on species, the odds of never rechecking the cuff were higher in first-opinion practice in dogs and cats and lower in cats for respondents who had more than 10 years of work experience (Table 2). Interestingly, no diplomate selected the option of never rechecking the cuff in dogs, and in cats, the odds of diplomates never rechecking cuff seal were significantly lower compared to non-diplomates (Table 2).

Rechecking of ETT cuff seal was most frequently performed only upon specific indication, with the most frequently selected specific indication being suspicion of cuff leakage.

The preferred technique for rechecking cuff seal was similar to the initial method used to establish the cuff seal; minimal occlusive volume was the preferred technique in dogs, cats and horses (57, 94 and 53% of answers, respectively) and verification of normal capnogram waveform in horses (56%).

4. Discussion

It is remarkable that no evidence-based consensus guidelines exist in veterinary medicine for such a common yet crucial procedure as creation of a tracheal seal after endotracheal intubation of various species. The American Animal Hospital Association Anaesthesia and Monitoring guidelines only state that, “A properly inflated cuff on a conventional ETT is necessary to create a seal for adequate positive pressure ventilation and avoid inhalant leakage, being aware that over-inflation may cause tracheal damage” [12]. The Association of Veterinary Anaesthetists recommended requirements for airway management when performing general anaesthesia of dogs, cats and horses merely stipulate that one must “ensure the animal’s airway is patent” [13]. Cuff pressure measurement has been recommended in human [8,14–16] and veterinary medicine [1,9] to guide ETT cuff inflation to sufficient yet not excessive pressure. Importantly, whereas under-inflation may give rise to inhalant leakage and aspiration, cuff over-inflation can lead to tracheal mucosal damage [9,17–19], tracheal necrosis [20]), tracheal perforation [21], airway obstruction [22] and postoperative complications, such as sore throat [23,24].

Despite these risks, in this survey, cuff pressure was measured by less than one-third of respondents in dogs and cats and by less than one in ten in farm animals and horses. Advanced anaesthesia training and more years of work experience did not seem to considerably affect the practice of cuff pressure measurement, as diplomates and respondents with more than 10 years of work experience were more likely to measure cuff pressure only in cats, and working in first-opinion practice in general did not influence the likelihood of measuring cuff pressure. One might expect that individuals with advanced anaesthesia training would be generally more aware of the literature on the risks of ETT cuff under- and over-inflation.

The infrequent use of ETT cuff pressure measurement in veterinary specialist or first-opinion practice, as apparent from this survey, may, in part, also be due to the absence of evidence-based guidelines for recommended cuff pressure ranges. The often generically cited recommended ETT cuff pressure range of 20–30 cm H₂O is derived from studies that showed impaired tracheal capillary perfusion when pressures over 30 cm H₂O were used in rabbits [3] and people [4]. However, venous and lymphatic drainage of the trachea is already impaired at much lower pressures [3], and hypotension can also negatively impact tracheal blood flow [25]. Moreover, the relation between ETT cuff pressure and tracheal transmural pressure is highly dependent on make, material and ETT diameter [1,5] and therefore cannot simply be extrapolated from one species to another. The silicone ETTs that are typically used in large animal anaesthesia feature low-volume–high pressure cuffs, which are more compliant than polyvinyl-chloride (PVC) cuffs [1,5]; therefore, a different cuff pressure target range seems to be necessary to provide an adequate seal in silicone compared to similarly sized PVC ETTs [1,9].

In a human anaesthesia survey, variations in cuff pressure targets were noted, with 11% of respondents targeting 10–20 cm H₂O, 38% 21–25 cm H₂O and 50% 26–30 cm H₂O [26]. In the current study, the range of cuff pressure that was targeted (when cuff pressure was measured) varied likewise, with 7% of respondents targeting <20 cm H₂O, 54% 20–30 cm H₂O and 34% >30 cm H₂O. One-fifth to one-fourth of respondents in our survey answered that they did not know the recommended range of cuff pressure for each species they intubated. This reflects the absence of literature data or consensus guidelines on recommended cuff pressure for various species and different makes and sizes of ETT in veterinary medicine. Alternatively, the lack of a significant difference in targeted cuff pressure between species may also reflect a type II error, given the low number of respondents actually measuring cuff pressure. Notably, the >30 cm H₂O cuff pressure target was selected far more often for horses than for other species. This may indicate that respondents are aware of the higher than 20–30 cm H₂O cuff pressure needed to achieve a seal with typical equine silicone ETTs, owing to their elastic properties and cuff geometry. In fact, 80 cm H₂O cuff pressure proved necessary to obtain a tracheal seal sufficient to prevent liquid leakage in horses when a 30 mm silicone ETT was used [9]. Interestingly, if study

participants were aware of the discrepancy in cuff pressure needed for silicone vs. PVC ETTs, they apparently did not extend this concept to the silicone ETTs used in farm animals and dogs.

Various techniques can be used and are taught to ascertain a tracheal seal after cuff inflation. The most frequently selected technique in cats, dogs and farm animals was manual syringe inflation of the cuff to minimal occlusive volume (MOV), when gas stopped audibly escaping during a positive pressure breath. In human medicine, this appears to be the safest method to achieve a targeted cuff pressure, other than the use of a manometer [8]. However, in large animals, the MOV technique is very impractical, as a lot of air is required, which could lead to multiple syringe detachments for refilling, which is probably why MOV was not the most frequently selected technique in horses. The fact that the use of a manometer was the least selected option for cuff inflation in horses (5%) in the current survey indicates that cuff pressure measurement is still uncommon in equine practice. This may well be partly due to the absence of guidelines on what exact cuff pressure should be targeted in horses.

Our results revealed very limited use of cuff manometers but common use of MOV to establish a cuff seal in dogs, cats and farm animals. This result is not in line with evidence obtained in humans [27–29], as well as in dogs and cats [1,2,30,31] in establishing the inferiority of MOV to the use of a manometer in achieving a pre-specified (i.e., target) cuff pressure. Furthermore, when comparing MOV in cats and dogs to the use of commercially available syringe inflation devices, MOV proved inferior [2,32]. Whereas these studies clearly concluded that a cuff manometer should be used to achieve optimal ETT cuff pressure, the use of MOV was three times more common than cuff pressure measurement in dogs and cats in the current study.

Several studies in human medicine have likewise demonstrated pilot balloon palpation to be an unreliable method to assess ETT cuff pressure [8,33,34], proving inferior to other techniques, such as loss of resistance [35], MOV [27] and the use of a manometer [27,36]. In two studies in cats, pilot balloon palpation performed worse than cuff pressure measurement with a manometer in achieving a target cuff pressure [2,31]. Despite these compelling data proving the inferiority of the technique for the purpose of cuff seal assessment, in our survey, pilot balloon palpation was still selected by 13–16% of respondents. It was the third most selected option for cuff seal verification across species and was selected more frequently than the superior technique of cuff pressure measurement.

Diplomates were more likely than other veterinary anaesthesia providers to use ≥ 3 techniques to establish cuff seal, reflecting that the number of techniques used may be influenced by training level. Whereas the combination of several techniques could prove beneficial in providing an estimate of seal adequacy, particularly in the absence of recommendations for target cuff pressures across species and ETT types, there is no actual evidence to support this notion at the moment. Interestingly, when participants were asked for the reason for their preference of cuff inflation technique, half of them responded that they were taught to do it this way, whereas about one-quarter of respondents indicated their selected techniques were performed by habit. As these answers strongly suggest that veterinary clinical teaching and training have a considerable impact on the technique used to verify tracheal seal after intubation, this provides a direct opportunity for education to help improve practice standards.

Experimental studies have shown that a gradual decrease, as well as fluctuations in cuff pressure, may occur over time in vitro and in intubated humans and animals [37–39], presenting a risk for gas and inhalant leakage and/or aspiration. In our survey, when participants were asked if they rechecked the adequacy of the cuff seal, “Yes, sometimes” was the most selected answer option for all species. Hence, it appears that tracheal seal reassessment is not common and certainly not standard practice in veterinary medicine. Years of work experience did not considerably affect standards of rechecking cuff seal adequacy. However, it appeared that training level and practice setting did affect rechecking standards, as the odds of never rechecking the cuff were considerably increased in first-opinion

practice in dogs and cats, whereas diplomates were significantly less likely to never recheck the cuff in cats. The odds ratios for diplomates never rechecking the cuff in other species could not be calculated because no diplomate selected this option in any other species. This suggests that advanced anaesthesia training improves anaesthetists' attention to ETT cuff seal maintenance. With regard to the frequency and timing of cuff seal verification after intubation, respondents most frequently answered that they do this "only on specific indication", which largely constituted suspicion of anaesthetic gas leakage. Perhaps surprisingly, rechecking cuff seal adequacy within 15 min was more likely to occur in dogs in first-opinion practice, with diplomate status having no effect on the likelihood of rechecking the cuff within 15 min. This may be related to the time taken between induction and theatre transfer (a likely moment for rechecking the cuff after repositioning), as this time frame may be shorter in first-opinion practice than in an university teaching hospital setting. Although guidelines exist for optimal timing and frequency of rechecking the cuff seal, cuff pressure was found to drop significantly in PVC ETTs in Beagle dogs, especially in the first 10 min after the start of anaesthesia [37]. This suggests that it may be sensible to recheck the cuff early and repeatedly. However, every time a cuff inflator is connected to the inflation valve of the ETT pilot balloon, the cuff pressure decreases by an average of 7 cm H₂O [40]. Whereas continuous cuff pressure measurement [41,42] and adjustment [43] are reliably possible and have been shown to reduce the incidence of ventilator-associated pneumonia in humans [44], this remains an experimental procedure due to the high costs and difficulties associated with practical feasibility.

This study represents a first attempt to evaluate current practice for ETT cuff inflation and seal verification in various veterinary species. Given its survey design and study participant recruitment, there are several limitations that should be noted. In terms of survey design, it should be noted that no formal testing for item reliability nor survey interpretation (legibility testing) was performed, partly due to limited ability to recruit a pilot group of representative respondents. Importantly, we used three different online resources to recruit participants. Two of these (ACVA-L and ECVA list servers) have an international membership with a special interest and/or advanced training in anaesthesia, whereas the Dutch closed Facebook community ('Dierenartsengilde') is aimed at first-opinion veterinarians. This led to the vast majority of respondents being Dutch veterinarians working in first-opinion practice. Therefore, for the results of this survey, we cannot distinguish between the effects of being in first-opinion practice and the effect of being based in the Netherlands on respondent answers. In hindsight, this could have been prevented if we had specifically targeted an international group of anaesthetists working in first-opinion practice, which we might have been able to achieve through the Association of Veterinary Anaesthetists (AVA). Our results are more likely to reflect Dutch first-opinion practice than first-opinion practice in general; further studies should be undertaken targeting an international first-opinion practice population to verify or extend our findings. Based on the Dutch situation, anaesthesia nurses and technicians were treated as one subgroup in the questionnaire, although their responsibilities and legal status can differ between countries. As the subgroup of anaesthesia nurses and technicians made up only 2% of study respondents, this is unlikely to have had an appreciable impact on our results. Furthermore, diplomates were slightly overrepresented in the responses for farm animals and horses; they comprised 54 and 52% of respondents for these species, respectively, versus 31% for both dogs and cats. The percentage of respondents working in first-opinion practice that provided answers for farm animals and horses was very low (2%); hence, answers provided in this survey are not a reliable reflection of first-opinion practice for farm animal and horses. In terms of answer frequencies, statistics on cuff pressure measurement may have suffered from type II error due to the small number of participants who actually measured cuff pressure. As a further limitation, we did not fully stratify for years of work experience; therefore, we cannot rule out effects within or beyond the dichotomized

boundaries (e.g., being very recently qualified or having >20 years of experience). Accordingly, the results with regard to the impact of work experience should be interpreted with caution.

Lastly, a question added to the survey about differences in cuff inflation methods for silicone and PVC ETTs specifically would have provided insight with respect to respondent awareness of implications of ETT material for cuff inflation practice.

5. Conclusions

Based on the results of this survey, endotracheal tube cuff pressure is not routinely measured in veterinary medicine, and anaesthesia providers seem insufficiently aware of possible differences in endotracheal tube cuff pressure between species and/or make of ETT. Despite literature evidence suggesting cuff pressure measurement by a manometer to be the superior technique for ETT cuff seal verification, this technique is infrequently used. Endotracheal tube cuff pressure is not routinely rechecked in Dutch first-opinion practice, and suspected leakage is the most common reason to recheck the cuff seal. Persistent preference for minimal occlusive volume, pilot balloon palpation and capnogram verification as cuff pressure verification techniques is partly a result of habits but mostly dictated by teaching. Hence, active education and development of evidence-based guidelines could help to improve practice. It should be noted that results from this questionnaire are based on a limited convenience sample with geographical bias and therefore may not be representative of global veterinary practice.

Author Contributions: Conceptualization, I.V. and J.C.d.G.; methodology, I.V. and J.C.d.G.; software, I.V.; validation, I.V. and J.C.d.G.; formal analysis, I.V.; investigation, I.V.; resources, I.V.; data curation, I.V.; writing—original draft preparation, I.V. and J.C.d.G.; writing—review and editing, I.V. and J.C.d.G.; visualization, I.V.; supervision, J.C.d.G.; project administration, I.V. All authors have read and agreed to the published version of the manuscript.

Funding: This research received no external funding.

Institutional Review Board Statement: Ethical review and approval were ruled not to be required under Dutch legislation. A statement to this effect from the institutional ethics committee officer is on file.

Informed Consent Statement: Informed consent was obtained from all subjects involved in the study.

Data Availability Statement: Not applicable.

Acknowledgments: Not applicable.

Conflicts of Interest: The authors declare no conflict of interest.

Appendix A. Questionnaire

Demographic

1. What is your gender?
 - Female
 - Male
 - Non-binary/third gender
 - Prefer not to say
2. In which country do you currently work?
3. What is your job title?
 - Veterinarian
 - Anaesthesiology nurse/technician
 - Anaesthesiology Intern
 - Anaesthesiology Resident
 - ACVAA diplomate

- ECVAA diplomate
 - ECC nurse/technician
 - ECC veterinarian
 - ACVECC diplomate
 - ECVECC diplomate
 - Other (please specify)
4. What is your place of work?
- First opinion practice
 - Referral practice
 - University hospital
 - Other (please specify)
5. How many years of veterinary experience do you have?
- ≤2 years
 - >2 years to ≤5 years
 - >5 years to ≤10 years
 - >10 years
6. In which animal species do you perform endotracheal intubation? (multiple options possible)
- Dogs
 - Cats
 - Farm Animals
 - Horses
- Cuff related questions per species
1. What material is the endotracheal tube for [species] made of? (multiple options possible)
- Silicone
 - Rubber
 - PVC
 - Don't know
2. Do you reuse the endotracheal tube for [species]?
- Yes
 - No
 - Other (please specify)
3. What method or methods do you use for cuff inflation in [species]? (multiple options possible)
- Syringe filled with air
 - Syringe filled with fluid
 - Cuff pressure measurement device
 - Other (please specify)
4. * How do you measure cuff pressure in [species]? (multiple options possible)
- Manometer
 - Tru-Cuff®
 - AG Cuffill®
 - PressureEasy®
 - Other (please specify)

5. What techniques do you use to establish a good seal of the cuff in [species]? (multiple options possible)
 - Minimal occlusive volume (=inflating cuff until no audible leak is present)
 - Absence of smell (=inflation cuff until absence of the smell of inhalational anaesthetic)
 - Palpation pilot balloon
 - Loss of resistance (=hyperinflation cuff and then allowing passive release of all excess air until the plunger of the syringe stops moving)
 - Syringe pressure (=cuffing to a cuff pressure by feeling the pressure in the syringe)
 - Filling of the ventilator bellows
 - Verification of normal capnogram waveform
 - Measurement of cuff pressure
 - Auscultation of the trachea
 - Other (please specify)
6. Why do you use this technique in [species]? (multiple options possible)
 - No specific preference
 - By habit
 - I was taught to do it this way
 - I have read in the literature that this was the recommended method
 - Other (please specify)
7. * What is the cuff pressure you use in [species]? (please include an unit, e.g., kPa, mmHg or cmH₂O)
8. Do you re-check the cuff in [species] (at a later time point)? (multiple options possible)
 - No, never
 - Yes, sometimes
 - Yes, frequently
 - Yes, always
 - Only with specific indications (please specify)
9. ** When do you re-check the cuff in [species] after endotracheal intubation?
 - ≤15 min
 - >15 to ≤30 min
 - >30 min
 - Only with specific indications (please specify)
 - Other (please specify)
10. ** How do you re-check the cuff in [species]?
 - Minimal occlusive volume (=inflating cuff until no audible leak is present)
 - Absence of smell (=inflation cuff until absence of the smell of inhalational anaesthetic)
 - Palpation pilot balloon
 - Loss of resistance (=hyperinflation cuff and then allowing passive release of all excess air until the plunger of the syringe stops moving)
 - Syringe pressure (=cuffing to a cuff pressure by feeling the pressure in the syringe)
 - Filling of the ventilator bellows
 - Verification of normal capnogram waveform
 - Measurement of cuff pressure
 - Auscultation of the trachea
 - Other (please specify)
11. What do you think the recommended range of cuff pressure is in [species]? (please include an unit, e.g., kPa, mmHg or cmH₂O)

12. Please place the following items associated with endotracheal tube cuffing in order of highest (1) to lowest (4) perceived importance in [species].

1. Decrease the risk of aspiration
2. Avoiding contamination of working environment with anaesthetic gas
3. Ability to provide mechanical ventilation
4. Avoiding tracheal mucosal damage

* questions was only available when in question 3 "cuff pressure measurement device" was selected.

** question was only available when in question 8 “no, never” was not selected.

Appendix B. Odds Ratio for Frequency of Response for Diplomates, First Opinion Practitioners and More than 10 Years of Experience for Selected Questionnaire Items

| | Diplomate | | | First-Opinion Practice | | | >10 Years' Experience | | |
|------------------------------------|-------------------------|-------|-------|-------------------------|-------|--------|-------------------------|-------|--------|
| | 95% Confidence Interval | | | 95% Confidence Interval | | | 95% Confidence Interval | | |
| | Odds Ratio | Lower | Upper | Odds Ratio | Lower | Upper | Odds Ratio | Lower | Upper |
| Species | | | | | | | | | |
| Dogs | 0.897 | 0.220 | 3.657 | 2.648 | 0.542 | 12.933 | 0.364 | 0.075 | 1.780 |
| Cats | 1.031 | 0.411 | 2.583 | 0.959 | 0.409 | 2.252 | 1.460 | 0.626 | 3.406 |
| Farm Animals | 5.010 | 3.072 | 8.170 | 0.010 | 0.002 | 0.041 | 1.463 | 0.929 | 2.304 |
| Horses | 4.611 | 2.842 | 7.481 | 0.008 | 0.002 | 0.034 | 1.364 | 0.875 | 2.128 |
| Reuse of endotracheal tubes | | | | | | | | | |
| Dogs | 1.221 | 0.745 | 2.001 | 0.980 | 0.614 | 1.654 | 1.316 | 0.773 | 2.241 |
| Cats | 1.785 | 1.090 | 2.924 | 0.852 | 0.530 | 1.371 | 1.280 | 0.759 | 2.157 |
| Farm Animals | 1.059 | 0.270 | 4.159 | * | | | 0.588 | 0.059 | 5.850 |
| Horses | 0.917 | 0.279 | 3.011 | * | | | 1.027 | 0.990 | 1.065 |
| Measuring cuff pressure | | | | | | | | | |
| Dogs | 1.221 | 0.745 | 2.001 | 0.980 | 0.614 | 1.654 | 1.320 | 0.823 | 2.117 |
| Cats | 1.785 | 1.090 | 2.924 | 0.852 | 0.530 | 1.371 | 1.712 | 1.054 | 2.783 |
| Farm Animals | 1.059 | 0.270 | 4.159 | * | | | 0.882 | 0.235 | 3.316 |
| Horses | 0.917 | 0.279 | 3.011 | * | | | 0.868 | 0.259 | 2.903 |
| ≥ 3 techniques to verify cuff seal | | | | | | | | | |
| All species | 2.075 | 1.307 | 3.294 | 0.453 | 0.288 | 0.710 | 1.152 | 0.746 | 1.778 |
| Never rechecking the cuff | | | | | | | | | |
| Dogs | ** | | | 10.238 | 3.486 | 30.063 | 0.303 | 0.134 | 0.683 |
| Cats | 0.062 | 0.008 | 0.463 | 8.854 | 3.313 | 23.662 | 0.547 | 0.262 | 1.142 |
| Farm Animals | ** | | | * | | | 0.955 | 0.895 | 1.018 |
| Horses | ** | | | * | | | 0.623 | 0.038 | 10.202 |
| Yes, sometimes | | | | | | | | | |
| Dogs | 1.104 | 0.696 | 1.752 | 0.936 | 0.609 | 1.439 | 1.096 | 0.714 | 1.684 |
| Cats | 1.250 | 0.779 | 2.005 | 0.792 | 0.510 | 1.231 | 1.114 | 0.717 | 1.729 |
| Farm Animals | 1.331 | 0.643 | 2.753 | * | | | 0.499 | 0.234 | 1.065 |
| Horses | 0.891 | 0.442 | 1.798 | * | | | 0.611 | 0.294 | 1.270 |
| Yes, frequently | | | | | | | | | |
| Dogs | 2.063 | 1.193 | 3.568 | 0.334 | 0.182 | 0.614 | 1.570 | 0.904 | 2.725 |
| Cats | 1.429 | 0.774 | 2.639 | 0.478 | 0.251 | 0.910 | 1.168 | 0.641 | 2.128 |
| Farm Animals | 0.804 | 0.336 | 1.922 | * | | | 3.091 | 1.071 | 8.915 |
| Horses | 1.447 | 0.483 | 4.337 | * | | | 1.847 | 0.553 | 6.164 |
| Yes, always | | | | | | | | | |

| | | | | | | | | | |
|--------------------------------|-------|-------|--------|-------|-------|--------|-------|-------|-------|
| Dogs | 1.126 | 0.508 | 2.498 | 0.372 | 0.155 | 0.893 | 2.308 | 0.997 | 5.344 |
| Cats | 1.186 | 0.530 | 2.651 | 0.324 | 0.128 | 0.819 | 2.084 | 0.894 | 4.857 |
| Farm Animals | 0.821 | 0.286 | 2.359 | * | | | 0.990 | 0.333 | 2.940 |
| Horses | 0.685 | 0.239 | 1.970 | * | | | 1.445 | 0.470 | 4.443 |
| Only with specific indications | | | | | | | | | |
| Dogs | 0.646 | 0.331 | 1.262 | 2.727 | 1.499 | 4.964 | 0.656 | 0.368 | 1.171 |
| Cats | 0.998 | 0.540 | 1.843 | 1.577 | 0.892 | 2.789 | 0.701 | 0.396 | 1.238 |
| Farm Animals | 4.521 | 1.418 | 14.418 | * | | | 2.982 | 0.933 | 9.531 |
| Horses | 2.336 | 0.963 | 5.664 | * | | | 2.815 | 1.050 | 7.549 |
| Recheck cuff <15 min | | | | | | | | | |
| Dogs | 0.529 | 0.232 | 1.207 | 2.138 | 1.059 | 4.314 | 0.881 | 0.437 | 1.774 |
| Cats | 0.573 | 0.269 | 1.224 | 1.693 | 0.871 | 3.291 | 1.006 | 0.514 | 1.966 |
| Farm Animals | 0.836 | 0.320 | 2.184 | * | | | 0.667 | 0.213 | 2.085 |
| Horses | 0.684 | 0.222 | 2.105 | 5.263 | 0.315 | 87.848 | 0.707 | 0.268 | 1.863 |

* It is not possible to calculate a reliable odds ratio for farm animals and horses due to a very limited number of non-diplomate respondents in these species categories; ** it is not possible to calculate an odds ratio, as no diplomates answered that they never rechecked the cuff.

Appendix C. Demographics of Study Participants

Table A1. Gender.

| | Frequency | Percentage (%) |
|-------------------------|-----------|----------------|
| Female | 259 | 74.4 |
| Male | 83 | 23.9 |
| Non-binary/third gender | 0 | 0 |
| Prefer not to say | 6 | 1.7 |

Table A2. Country.

| | Frequency | Percentage (%) |
|--------------------|-----------|----------------|
| Australia | 14 | 4.0 |
| Austria | 1 | 0.3 |
| Belgium | 10 | 2.9 |
| Canada | 10 | 2.9 |
| Colombia | 1 | 0.3 |
| Finland | 1 | 0.3 |
| France | 3 | 0.9 |
| Germany | 6 | 1.7 |
| Greece | 1 | 0.3 |
| Hong Kong (S.A.R.) | 1 | 0.3 |
| Ireland | 3 | 0.9 |
| Israel | 1 | 0.3 |
| Italy | 11 | 3.2 |
| Netherlands | 163 | 46.8 |
| New Zealand | 1 | 0.3 |
| Norway | 2 | 0.6 |
| South Africa | 2 | 0.6 |
| South Korea | 1 | 0.3 |
| Spain | 6 | 1.7 |
| Sweden | 3 | 0.9 |
| Switzerland | 9 | 2.6 |

| | | |
|--|----|------|
| United Kingdom of Great Britain and Northern Ireland | 42 | 12.1 |
| United States of America | 56 | 16.1 |

Table A3. Job title.

| | Frequency | Percentage (%) |
|-------------------------------------|-----------|----------------|
| Job title | | |
| Veterinarian | 182 | 52.3 |
| Anaesthesiology nurse/technician | 7 | 2.0 |
| Anaesthesiology intern | 1 | 0.3 |
| Anaesthesiology resident | 34 | 9.8 |
| ACVAA diplomate | 48 | 13.8 |
| ECVAA diplomate | 61 | 17.5 |
| ECC nurse/technician | 0 | 0 |
| ECC veterinarian | 1 | 0.3 |
| ACVECC diplomate | 0 | 0 |
| ECVECC diplomate | 0 | 0 |
| Other | 14 | 4.0 |
| Job title (other). | | |
| Specialist small animal surgery | 1 | 7.1 |
| Anaesthesiology professor | 2 | 14.2 |
| PhD anaesthesia | 1 | 7.1 |
| Residency trained anaesthesiologist | 10 | 71.4 |

Table A4. Place of work.

| | Frequency | Percentage (%) |
|---------------------------------------|-----------|----------------|
| Place of work | | |
| First-opinion practice | 151 | 43.4 |
| Referral practice | 75 | 21.6 |
| University hospital | 110 | 31.6 |
| Other | 12 | 3.4 |
| Place of work (other) | | |
| University and referral practice | 2 | 16.7 |
| University and first-opinion practice | 1 | 8.3 |
| First-opinion and referral practice | 1 | 8.3 |
| Consultant | 2 | 16.7 |
| Executive | 1 | 8.3 |
| Research organisation | 2 | 16.7 |
| Preclinical industry | 1 | 8.3 |
| Retired | 1 | 8.3 |

Table A5. Years of experience.

| | Frequency | Percentage (%) |
|-----------------------|-----------|----------------|
| ≤2 years | 38 | 10.9 |
| >2 years to ≤5 years | 38 | 10.9 |
| >5 years to ≤10 years | 75 | 21.6 |
| >10 years | 197 | 56.6 |

Table A6. Animal intubations performed.

| | Frequency | Percentage per Species | Percentage of Total |
|------------------|-----------|------------------------|---------------------|
| Dogs | 339 | 97.4 | 37.3 |
| Cats | 325 | 93.4 | 35.8 |
| Farm animals | 118 | 33.9 | 13.0 |
| Horses | 127 | 36.5 | 14.0 |
| All four species | 111 | | 31.9 |

Table A7. Frequency and percentage of diplomates per species.

| | Frequency | Percentage per Species | Frequency (Total) |
|--------------|-----------|------------------------|-------------------|
| Dogs | 105 | 31.0 | 339 |
| Cats | 101 | 31.1 | 325 |
| Farm animals | 64 | 54.2 | 118 |
| Horses | 66 | 52.0 | 127 |

Table A8. Frequency and percentage of diplomates per country.

| | Frequency | Percentage (%) |
|--|-----------|----------------|
| Australia | 11 | 10.2 |
| Austria | 0 | 0 |
| Belgium | 2 | 1.9 |
| Canada | 6 | 5.6 |
| Colombia | 0 | 0 |
| Finland | 1 | 0.9 |
| France | 1 | 0.9 |
| Germany | 3 | 2.8 |
| Greece | 1 | 0.9 |
| Hong Kong (S.A.R.) | 1 | 0.9 |
| Ireland | 3 | 2.8 |
| Israel | 1 | 0.9 |
| Italy | 11 | 10.2 |
| Netherlands | 1 | 0.9 |
| New Zealand | 0 | 0 |
| Norway | 1 | 0.9 |
| South Africa | 1 | 0.9 |
| South Korea | 0 | 0 |
| Spain | 2 | 1.9 |
| Sweden | 2 | 1.9 |
| Switzerland | 6 | 5.6 |
| United Kingdom of Great Britain and Northern Ireland | 24 | 22.2 |
| United States of America | 37 | 34.3 |

| | |
|-------|-----|
| Total | 108 |
|-------|-----|

Table A9. Frequency and percentage of first-opinion practice per species.

| | Frequency | Percentage per Species | Frequency (Total) |
|--------------|-----------|------------------------|-------------------|
| Dogs | 146 | 43.1 | 339 |
| Cats | 138 | 42.5 | 325 |
| Farm animals | 2 | 1.7 | 118 |
| Horses | 2 | 1.6 | 127 |

Table A10. Frequency and percentage of job title of Dutch compared to non-Dutch respondents.

| | Frequency Dutch | Percentage Dutch | Frequency Non-Dutch | Percentage Non-Dutch |
|----------------------------------|--------------------|---------------------|------------------------|-------------------------|
| Veterinarian | 157 | 45.1 | 25 | 7.2 |
| Anaesthesiology nurse/technician | 0 | 0 | 7 | 2.0 |
| Anaesthesiology intern | 1 | 0.3 | 0 | 0 |
| Anaesthesiology resident | 2 | 0.6 | 32 | 9.2 |
| ACVAA diplomate | 0 | 0 | 48 | 13.8 |
| ECVAA diplomate | 1 | 0.3 | 60 | 17.2 |
| ECC nurse/technician | 0 | 0 | 0 | 0 |
| ECC veterinarian | 1 | 0.3 | 0 | 0 |
| ACVECC diplomate | 0 | 0 | 0 | 0 |
| ECVECC diplomate | 0 | 0 | 0 | 0 |
| Other | 1 | 0.3 | 13 | 3.7 |
| Total | 163 | 46.8 | 185 | 53.1 |

Table A11. Frequency and percentage of Dutch respondents working in first-opinion practice compared to total respondents working in first-opinion practice.

| | Frequency | Percentage |
|---|-----------|------------|
| Dutch respondents | 163 | 46.8 |
| Total respondents | 348 | 100 |
| Dutch respondents working in first-opinion practice | 143 | 94.7 |
| Total respondents working in first-opinion practice | 151 | 100 |

References

- Briganti, A.; Portela, D.A.; Barsotti, G.; Romano, M.; Breggi, G. Evaluation of the endotracheal tube cuff pressure resulting from four different methods of inflation in dogs. *Vet. Anaesth. Analg.* **2012**, *39*, 488–494. <https://doi.org/10.1111/j.1467-2995.2012.00719.x>.
- White, D.M.; Makara, M.; Martinez-Taboada, F. Comparison of four inflation techniques on endotracheal tube cuff pressure using a feline airway simulator. *J. Feline Med. Surg.* **2019**, *22*, 641–647. <https://doi.org/10.1177/1098612x19871701>.
- Nordin, U.; Lindholm, C.-E.; Wolgast, M. Blood Flow in the Rabbit Tracheal Mucosa under Normal Conditions and under the Influence of Tracheal Intubation. *Acta Anaesthesiol. Scand.* **1977**, *21*, 81–94. <https://doi.org/10.1111/j.1399-6576.1977.tb01198.x>.
- Seegobin, R.D.; Van Hasselt, G.L. Endotracheal cuff pressure and tracheal mucosal blood flow: endoscopic study of effects of four large volume cuffs. *Br J Med* **1984**, *288*, 965–968. <https://doi.org/10.1136/bmj.288.6422.965>.
- Dobrin, P.; Canfield, T. Cuffed endotracheal tubes: Mucosal pressures and tracheal wall blood flow. *Am. J. Surg.* **1977**, *133*, 562–568.
- Black, A.; Seegobin, R. Pressures on endotracheal tube cuffs. *Anaesthesia* **1981**, *36*, 498–511. <https://doi.org/10.1111/j.1365-2044.1981.tb10286.x>.
- Brimacombe, J.; Keller, C.; Giampalmo, M.; Sparr, H.J.; Berry, A. Direct measurement of mucosal pressures exerted by cuff and non-cuff portions of tracheal tubes with different cuff volumes and head and neck positions. *Br. J. Anaesth.* **1999**, *82*, 708–711. <https://doi.org/10.1093/bja/82.5.708>.

8. Sultan, P.; Carvalho, B.; Rose, B.O.; Cregg, R. Endotracheal Tube Cuff Pressure Monitoring: A Review of the Evidence. *J. Perioper. Pract.* **2011**, *21*, 379–386. <https://doi.org/10.1177/175045891102101103>.
9. Touzot-Jourde, G.; Stedman, N.L.; Trim, C.M. The effects of two endotracheal tube cuff inflation pressures on liquid aspiration and tracheal wall damage in horses. *Vet. Anaesth. Analg.* **2005**, *32*, 23–29. <https://doi.org/10.1111/j.1467-2995.2004.00170.x>.
10. Bennett, R.C. Web based surveys—Recommendations for their design and interpretation. *Vet. Anaesth. Analg.* **2020**, *47*, 1–2. <https://doi.org/10.1016/j.vaa.2019.12.002>.
11. Eysenbach, G. Improving the Quality of Web Surveys: The Checklist for Reporting Results of Internet E-Surveys (CHERRIES). *J. Med. Internet Res.* **2004**, *6*, e34. <https://doi.org/10.2196/jmir.6.3.e34>.
12. Grubb, T.; Sager, J.; Gaynor, J.S.; Montgomery, E.; Parker, J.A.; Shafford, H.; Tearney, C. 2020 AAHA Anesthesia and Monitoring Guidelines for Dogs and Cats. *J. Am. Anim. Hosp. Assoc.* **2020**, *56*, 59–82. <https://doi.org/10.5326/jaaha-ms-7055>.
13. AVA Association of Veterinary Anaesthetists. Recommended Requirements When Performing General Anaesthesia of Dogs, Cats and Horses. 2008. Available online: <https://ava.eu.com/wp-content/uploads/2015/10/AVA-RECOMMENDED-REQUIREMENTS-ENG.pdf> (accessed on 17 June 2022).
14. Sengupta, P.; Sessler, D.I.; Maglinger, P.; Wells, S.; Vogt, A.; Durrani, J.; Wadhwa, A. Endotracheal tube cuff pressure in three hospitals, and the volume required to produce an appropriate cuff pressure. *BMC Anesthesiol.* **2004**, *4*, 8. <https://doi.org/10.1186/1471-2253-4-8>.
15. Khan, M.; Khokar, R.; Qureshi, S.; Al Zahrani, T.; Aqil, M.; Shiraz, M. Measurement of endotracheal tube cuff pressure: Instrumental versus conventional method. *Saudi J. Anaesth.* **2016**, *10*, 428–431. <https://doi.org/10.4103/1658-354x.179113>.
16. Hockey, C.A.; Van Zundert, A.A.J.; Paratz, J.D. Does Objective Measurement of Tracheal Tube Cuff Pressures Minimise Adverse Effects and Maintain Accurate Cuff Pressures? A Systematic Review and Meta-Analysis. *Anaesth. Intensiv. Care* **2016**, *44*, 560–570. <https://doi.org/10.1177/0310057x1604400503>.
17. Holland, M.; Snyder, J.R.; Steffey, E.P.; Heath, R.B. Laryngotracheal injury associated with nasotracheal intubation in the horse. *J. Am. Vet. Med. Assoc.* **1986**, *189*, 1447–1450.
18. Heath, R.B.; Steffey, E.P.; Thurmon, J.C.; Wertz, E.M.; Meagher, D.M.; Hyypä, T.; Van Slyke, G.L. Laryngotracheal lesions following routine orotracheal intubation in the horse. *Equine Vet. J.* **1989**, *21*, 434–437. <https://doi.org/10.1111/j.2042-3306.1989.tb02190.x>.
19. Abud, T.M.V.; Braz, J.R.C.; Martins, R.H.G.; Gregório, E.A.; Saldanha, J.C.; Raiza, A.C.P. The Lanz® endotracheal tube decreases tracheal injury in dogs. *J. Anaesth.* **2005**, *52*, 878–882. <https://doi.org/10.1007/bf03021785>.
20. Wylie, C.E.; Foote, A.K.; Rasotto, R.; Cameron, I.J.; Greet, T.R.C. Tracheal necrosis as a fatal complication of endotracheal intubation. *Equine Vet. Educ.* **2015**, *27*, 170–175. <https://doi.org/10.1111/eve.12276>.
21. Saulez, M.N.; Dziki, B.; Voigt, A. Traumatic perforation of the trachea in two horses caused by orotracheal rupture. *Vet. Rec. Case Rep.* **2009**, *164*, 719–722.
22. Veen, I.; de Grauw, J. Endotracheal tube obstruction due to cuff over-inflation or cuff herniation in small equids—A case series. *Equine Vet. Educ.* **2022**, *accepted*.
23. Liu, J.; Zhang, X.; Gong, W.; Li, S.; Wang, F.; Fu, S.; Zhang, M.; Hang, Y. Correlations between controlled endotracheal tube cuff pressure and postprocedural complications: a multicentre study. *Anesth. Analg.* **2010**, *111*, 1133–1137.
24. Ansari, L.; Bohlu, B.; Mahaseni, H.; Valaei, N.; Sadr-Eshkevari, P.; Rashad, A. The effect of endotracheal tube cuff pressure control on postextubation throat pain in orthognathic surgeries: A randomized double-blind controlled clinical trial. *Br. J. Oral Maxillofac. Surg.* **2014**, *52*, 140–143. <https://doi.org/10.1016/j.bjoms.2013.10.005>.
25. Bunegin, L.; Albin, M.S.; Smith, R.B. Canine Tracheal Blood Flow After Endotracheal Tube Cuff Inflation During Normotension and Hypotension. *Anesth. Analg.* **1993**, *76*, 1083–1090. <https://doi.org/10.1213/00000539-199305000-00030>.
26. Talekar, C.R.; Udy, A.; Boots, R.; Lipman, J.; Cook, D. Tracheal Cuff Pressure Monitoring in the ICU: A Literature Review and Survey of Current Practice in Queensland. *Anaesth. Intensiv. Care* **2014**, *42*, 761–770. <https://doi.org/10.1177/0310057x1404200612>.
27. Totonchi, Z.; Jalili, F.; Hashemian, S.M.; Jabardarjani, H.R. Tracheal Stenosis and Cuff Pressure: Comparison of Minimal Occlusive Volume and Palpation Techniques. *Tanaffos* **2015**, *14*, 252–256.
28. Harvie, D.A.; Darvall, J.N.; Dodd, M.; De La Cruz, A.; Tacey, M.; D’Costa, R.L.; Ward, D. The Minimal Leak Test Technique for Endotracheal Cuff Maintenance. *Anaesth. Intensiv. Care* **2016**, *44*, 599–604. <https://doi.org/10.1177/0310057x1604400512>.
29. Sanaie, S.; Rahmani, F.; Chokhachian, S.; Mahmoodpoor, A.; Panahi, J.R.; Esfanjani, R.M.; Mirzaei, M.; Soleimanpour, H. Comparison of tracheal tube cuff pressure with two technique: fixed volume and minimal leak test techniques. *J. Cardiovasc. Thorac. Res.* **2019**, *11*, 48–52. <https://doi.org/10.15171/jcvtr.2019.08>.
30. White, D.M.; Redondo, J.I.; Mair, A.R.; Martinez-Taboada, F. The effect of user experience and inflation technique on endotracheal tube cuff pressure using a feline airway simulator. *Vet. Anaesth. Analg.* **2017**, *44*, 1076–1084. <https://doi.org/10.1016/j.vaa.2016.11.006>.
31. Bird, A.R.; Bird, D.J.; McMillan, M.W. Aspects of in vivo endotracheal tube intracuff pressure in cats. *Vet. Anaesth. Analg.* **2018**, *46*, 55–63. <https://doi.org/10.1016/j.vaa.2018.06.017>.
32. Hung, W.-C.; Ko, J.C.; Weil, A.B.; Weng, H.-Y. Evaluation of Endotracheal Tube Cuff Pressure and the Use of Three Cuff Inflation Syringe Devices in Dogs. *Front. Vet. Sci.* **2020**, *7*, 39. <https://doi.org/10.3389/fvets.2020.00039>.
33. Hoffman, R.J.; Parwani, V.; Hahn, I.-H. Experienced emergency medicine physicians cannot safely inflate or estimate endotracheal tube cuff pressure using standard techniques. *Am. J. Emerg. Med.* **2006**, *24*, 139–143. <https://doi.org/10.1016/j.ajem.2005.07.016>.

34. Giusti, G.D.; Rogari, C.; Gili, A.; Nisi, F. Cuff pressure monitoring by manual palpation in intubated patients: How accurate is it? A manikin simulation study. *Aust. Crit. Care* **2016**, *30*, 234–238. <https://doi.org/10.1016/j.aucc.2016.10.001>.
35. Bulamba, F.; Kintu, A.; Ayupo, N.; Kojjo, C.; Ssemogerere, L.; Wabule, A.; Kwizera, A. Achieving the Recommended Endotracheal Tube Cuff Pressure: a Randomized Control Study Comparing Loss of Resistance Syringe to Pilot Balloon Palpation. *Anesthesiol. Res. Pract.* **2017**, *2017*, 2032748. <https://doi.org/10.1155/2017/2032748>.
36. Abubaker, J.; Ullah, S.Z.; Ahmed, S.; Memon, A.U.R.; Abubaker, Z.J.; Ansari, M.I.; Karim, M. Evaluating the Knowledge of Endotracheal Cuff Pressure Monitoring Among Critical Care Providers by Palpation of Pilot Balloon and By Endotracheal Tube Cuff Manometer. *Cureus* **2019**, *11*, e5061. <https://doi.org/10.7759/cureus.5061>.
37. Shin, C.W.; Son, W.; Jang, M.; Kim, H.; Han, H.; Cha, J.; Lee, I. Changes in endotracheal tube intracuff pressure and air leak pressure over time in anaesthetized Beagle dogs. *Vet. Anaesth. Analg.* **2018**, *45*, 737–744.
38. Marti, J.D.; Bassi, G.L.; Isetta, V.; Lazaro, M.R.; Aguilera-Xiol, E.; Comaru, T.; Battaglini, D.; Meli, A.; Ferrer, M.; Navajas, D.; et al. An in-vitro study to evaluate high-volume low-pressure endotracheal tube cuff deflation dynamics. *Minerva Anesthesiol.* **2019**, *85*, 846–853. <https://doi.org/10.23736/S0375-9393.19.13133-1>.
39. Adachi, K.; Kameyama, Y.; Andoh, K. Examination of the frequency of upward and downward fluctuations in the pressure obtained from the cuff pressure-time curve by continuous measurement of endotracheal tube cuff pressure during thyroid surgery: a case series. *JA Clin. Rep.* **2021**, *7*, 63. <https://doi.org/10.1186/s40981-021-00466-4>.
40. Asai, S.; Motoyama, A.; Matsumoto, Y.; Konami, H.; Imanaka, H.; Nishimura, M. Decrease in cuff pressure during the measurement procedure: an experimental study. *J. Intensiv. Care* **2014**, *2*, 34. <https://doi.org/10.1186/2052-0492-2-34>.
41. Sole, M.L.; Aragon, D.; Bennett, M.; Johnson, R.L. Continuous measurement of endotracheal tube cuff pressure: how difficult can it be? *AACN Adv. Crit. Care* **2008**, *19*, 235–243.
42. Krishna, S.; Ramesh, A.S.; Jatana, K.R.; Elmaraghy, C.; Merz, M.; Ruda, J.; Tobias, J.D. A technique to measure the intracuff pressure continuously: an in vivo demonstration of its accuracy. *Pediatr. Anaesth.* **2014**, *24*, 999–1004.
43. Dauvergne, J.E.; Geffray, A.-L.; Asehnoune, K.; Rozec, B.; Lakhal, K. Automatic regulation of the endotracheal tube cuff pressure with a portable elastomeric device. A randomised controlled study. *Anaesth. Crit. Care Pain Med.* **2020**, *39*, 435–441. <https://doi.org/10.1016/j.accpm.2020.04.007>.
44. Sevdi, M.S.; Demirgan, S.; Erkalp, K.; Akyol, O.; Ozcan, F.G.; Guneyli, H.C.; Tunali, M.C.; Selcan, A. Continuous endotracheal tube cuff pressure control decreases incidence of ventilator-associated pneumonia in patients with traumatic brain injury. *J. Investig. Surg.* **2022**, *35*, 525–530.