

BTS - 600 Prüfprogramm: fstCheck

Datum/Zeit : 14.09.2017 19:36:29

Version : V 1.600.395

Bediener : Digatron

Programm : fstCheck

Titel : first CheckUpTC

Schritt	Label	Prozedur	Sollwert	Limit	Aktion	Registrierung
1		SET				Multi 0.1 V 0.1 A 1 min
2		SET		> 3.0 V < 1.0 V	UNT UNT	
3		SET	ClimaOn = 1.0			
4		SET	SetTemp = 25.0			
5		UNT				
6		PAU		68 h		
!	first EIS measurement (removed)					
!	first charging and cap-test (removed)					
7		LAD	5.0 ACN20 2.6 V	30 h		
8		PAU		4 h		
9		ELA	4.0 ACN20	< 1.8 V		
10		PAU		4 h		
11		LAD	5.0 ACN20 2.6 V	24 h		
!	EIS at 80 % SoC					
12		PAU		24 h		
13		UNT				
14		ELA	2.0 ACN20	2.0 AhStep		
15		EISwDCA				
16		PAU		24 h		
17		EISwDCA				
!	EIS at 40 % SoC					
18		ELA	2.0 ACN20	4.0 AhStep		
19		EISwDCA				
!	EIS at 0 % SoC					
20		ELA	2.0 ACN20	4.0 AhStep		
21		EISwDCA				
!	recharge					
22		LAD	5.0 ACN20 2.6 V	24 h		
23		PAU		10 h		
24		STO				