

BTS - 600 Prüfprogramm: EISslave

Datum/Zeit : 20.02.2018 11:54:43

Version : V 1.600.395

Bediener : Digatron

Programm : EISslave

Titel : programm f EISmeter

Schritt	Label	Prozedur	Sollwert	Limit	Aktion	Registrierung
1		SET				EisReg 10 min
2		BEG				
3		PAU		EISstart > 0.0		
4		SET	EISready = 0.0			
5		PAU		2 sec		
6		EIS	0.1 mHz 6.0 kHz 2.0 AAcMax 4.0 VAcMax -4.0 VAcMin 1.5 mVideal 2.0 AAcStart 1.0 EisTime	EISstart < 1.0		
7		PAU		2 sec		
8		SET	EISready = 1.0			
9		ZYK	16*			
10		STO				