

Supplementary Table S1

Supplementary Table S1. Chromatographic and electrospray ionization conditions.

Chromatographic conditions		Electrospray source conditions	
Flow	0.300	Gas temperature (°C)	225
Injection volume	10 µL	Gas Flow (L/min)	8
Solvent A	13 mM ammonium acetate pH 7	Nebulizer (psi)	30
Solvent B	Acetonitrile/formic acid (99.99:0.01, v/v)	Seath Gas Heater (°C)	350
Gradient		Seath Gas Flow (L)	10
		Capillary (V)	3500
Time (min.)	%B	VCharging (eV)	1250
0.20	12.00		
0.21	70.00		
1.00	30.00		
1.01	52.00		
2.50	52.00		
4.00	100.00		
7.00	100.00		
80.00	12.00		