

Dynamic Susceptibility Contrast-MRI Quantification Software Tool: Development and Evaluation

Panagiotis Korfiatis¹, Timothy L. Kline¹, Zachary S. Kelm¹, Rickey E. Carter², Leland S. Hu³, and Bradley J. Erickson¹

¹Department of Radiology, Mayo Clinic, Rochester, Minnesota; ²Department of Health Sciences Research, Mayo Clinic, Rochester, Minnesota; and ³Department of Radiology, Mayo Clinic, Scottsdale, Arizona

Corresponding Author:

Bradley J. Erickson, MD, PhD
Department of Radiology, Mayo Clinic,
200 first Street SW,
Rochester, MN 55905;
E-mail: bje@mayo.edu

Key Words: dynamic susceptibility contrast, glioblastoma, atlas segmentation, white matter

Abbreviations: Relative cerebral blood volume (rCBV), white matter (WM), magnetic resonance imaging (MRI), arterial input function (AIF), regions of interest (ROIs), gadolinium (Gd), cerebral blood volume (CBV), Food and Drug Administration (FDA), T1-weighted (T1w), relaxation time (TR), echo time (TE), flip angle (FA), field of view (FOV), signal intensity (SI), intraclass correlation coefficient (ICC), standard deviations (SDs)

ABSTRACT

Relative cerebral blood volume (rCBV) is a magnetic resonance imaging biomarker that is used to differentiate progression from pseudoprogression in patients with glioblastoma multiforme, the most common primary brain tumor. However, calculated rCBV depends considerably on the software used. Automating all steps required for rCBV calculation is important, as user interaction can lead to increased variability and possible inaccuracies in clinical decision-making. Here, we present an automated tool for computing rCBV from dynamic susceptibility contrast-magnetic resonance imaging that includes leakage correction. The entrance and exit bolus time points are automatically calculated using wavelet-based detection. The proposed tool is compared with 3 Food and Drug Administration-approved software packages, 1 automatic and 2 requiring user interaction, on a data set of 43 patients. We also evaluate manual and automated white matter (WM) selection for normalization of the cerebral blood volume maps. Our system showed good agreement with 2 of the 3 software packages. The intraclass correlation coefficient for all comparisons between the same software operated by different people was >0.880 , except for FuncTool when operated by user 1 versus user 2. Little variability in agreement between software tools was observed when using different WM selection techniques. Our algorithm for automatic rCBV calculation with leakage correction and automated WM selection agrees well with 2 out of the 3 FDA-approved software packages.

INTRODUCTION

Relative cerebral blood volume (rCBV) is a magnetic resonance imaging (MRI) biomarker computed from dynamic susceptibility contrast (DSC) images, and has been used extensively in brain tumor imaging for differentiation of progression versus pseudoprogression (1, 2), tumor grading (3), survival prediction (4), and tumor differentiation (5).

Perfusion analysis software to compute rCBV from DSC-MRI is widely available in clinical practice. However, it is commonly treated as a “black box,” and broad-scale integration has been slowed by the need for defining optimal methodological conditions to maximize rCBV accuracy.

The rCBV for each voxel is calculated by trapezoidal integration under the $\Delta R2^*(t)$ area curve from the start to the end of the first-pass contrast bolus on a voxel basis divided by the value calculated for normal-appearing white matter (WM). The following pitfalls are included in processing DSC-MRI data: an inappropriate kinetic model, incorrect arterial input function

(AIF) detection or deconvolution, inaccurate geometric mapping of selected anatomic regions of interest (ROIs) to the perfusion maps, incorrect calibration or normalization techniques, and incorrect identification of the bolus entrance and exit time points (6). Determining the start, end, and peak of the bolus is the most critical step of the algorithm, and it will affect calculation of percent signal recovery and mean transit time, in addition to rCBV.

DSC imaging relies on the assumptions that the gadolinium (Gd)-based contrast agent passes through tissue as a bolus and it remains within vessels without leaking into the surrounding tissue (7). Disruption of the blood–brain barrier in glioblastoma multiforme tissue violates the second assumption, and thus “leakage correction” is needed (8). Leakage correction can be addressed using one of the following methods: “preloading,” where some Gd is administered before DSC imaging (9, 10), leakage modeling during postprocessing (7, 11), by shortening the flip angle (FA) (12, 13), or with dual-echo acquisitions (14).

Although most clinical sites preload, this does not eliminate leakage. Reducing the FA decreases signal, and the use of dual-echo sequences is clinically challenging, and is currently rarely used. Thus, modeling of leakage is essential for accurate rCBV calculation. The rCBV that is calculated depends considerably on the software used (6), with modeling correction necessary for correction of residual T1 errors and T2/T2*-weighted recirculation. Currently, the method published by Boxerman et al. (11) is considered the standard for DSC-MRI leakage correction.

Recently, it was reported (15, 16) that normalization of the derived perfusion metrics to the (presumably healthy) contralateral WM resulted in increased repeatability of measurements, whereas deconvolution of the AIF may reduce repeatability. In clinical practice, a cerebral blood volume (CBV) normalized to contralateral WM without AIF deconvolution is considered to result in more repeatable values (15, 16). Currently, only 1 automated method exists for performing normalization, but it requires both WM and gray matter identification (17).

Here, we present an automated DSC-MRI quantification tool that performs leakage correction and compare it with 3 Food and Drug Administration (FDA)-approved software packages, 1 automatic and 2 requiring user interaction, on a data set of 43 patients. We also evaluate manual and automated WM selection for normalization of the CBV maps.

MATERIALS AND METHODS

Data Set

This study was reviewed and approved by our institutional review board, with waiver of informed consent. Inclusion criteria included biopsy diagnosis of glioblastoma multiforme and treatment with radiation therapy and temozolomide administration according to Stupp protocol during the period from 2007 to 2013. In total, 43 patients where post-processing with all the 3 FDA-approved software packages was available were considered.

MRI Protocol

Each imaging examination was acquired using a General Electric MR scanner (GE Healthcare, Milwaukee, Wisconsin), operating at 1.5 T (N = 27) or 3 T (N = 16). The T1-weighted (T1w) postcontrast images, except for 2 cases, were acquired at an oblique axial angle using either spin echo or fast-spin echo sequences ~10 minutes after Gd injection. The T1w parameters for the 1.5 T spin echo sequence were as follows: relaxation time (TR) = 433–683 milliseconds, echo time (TE) = 20–21 milliseconds, FA = 90°, matrix = 256 × 192, field of view (FOV) = 220 × 220–250 × 250 mm, and section thickness = 4 mm, with no section gap. T1w parameters for 2-dimensional 3 T acquisitions were as follows: TR = 467–700 milliseconds, TE = 20 milliseconds, FA = 90°, matrix = 320 × 192, FOV = 220 × 220 mm, and section thickness = 4 mm, with no section gap. For both the 1.5 and 3 T scans, the DSC images were obtained using a spin echo echo-planar sequence with axial orientation and TR = 2217–2225 milliseconds, TE = 60 milliseconds, FA = 90°, matrix = 128 × 96, FOV = 240 × 240 mm, section thickness = 5 mm, and section gap = 5 mm. In total, 40 successive time points were imaged with ~2 seconds between acquisitions. For DSC imaging, a total dose of 0.1 mmol/kg gadolinium contrast was used, with 2 cc of the dose injected about 5 minutes before

preloading. The bolus injection commenced 10 seconds after acquisition started, and with transit time through the venous system, lungs, and great vessels, appearance in the brain was typically 25 seconds from start of acquisition.

Tumor Segmentation

Enhancing tumor volumes were segmented from postsurgical, postcontrast, T1w images by a semiautomated thresholding technique. A user drew a generous boundary around the tumor, and subsequently, a thresholding technique based on Otsu thresholding was applied to finalize tumor segmentation (18).

WM Selection

We studied both manual and automatic WM segmentation techniques. The automatic technique (automated WM selection) was designed to automatically select WM areas that were not affected by tumor, based on the DSC MRI data. To select WM areas, the DSC signal data were transformed to concentration images (19). All the data before the time point of bolus entrance were discarded. An unsupervised segmentation algorithm based on mean shift (20, 21) was used to segment the different classes of tissue based on the concentration–time curves. Because the resulting clusters did not always correspond to WM, a WM atlas was used to select the appropriate WM cluster. For this, we used the ICBM 152 Nonlinear Atlas Version 2009 (22). To register the T2-based atlas to the DSC perfusion scan, a symmetric diffeomorphic deformation model (that preserves anatomical topology even in cases requiring large deformations) was used, with mutual information as the similarity metric. The implementation was based on the ANTs registration package (23). Pixels with at least 95% WM probability were selected as the WM ROI.

For the manual WM ROI selection, two users used itk-SNAP (24) to select areas of WM contralateral to the tumor region (referred to as WM user 1 and WM user 2).

CBV Calculation Algorithm

The following steps were used to calculate the bolus entrance and exit time points.


- (1) The first 3 time points of each perfusion series are removed because of saturation effects.
- (2) The average signal–time curve for the brain region is calculated.
- (3) A continuous wavelet transform on the signal–time curve is performed.
- (4) The maxima and minima at each scale and link across each scale are identified. The links are used to identify the maxima and minima of the curve.
- (5) The signal minima and the two local maxima closest to it are located.

Subsequently, the baseline signal intensity (SI) is calculated. Next, the signal–time curves are converted to concentration–time curves based on equation 1 as follows:

$$\Delta 2^*((t)) = -\ln(S(t)/S_{base})/TE \quad (1)$$

where $\Delta 2^*((t))$ is the reflexivity–time curve and is a parameter related to the concentration of the Gd in a voxel, S is the dynamic SI, S_{base} is the baseline, TE is the echo time, and S(t) is the dynamic SI.

Figure 1. Proposed system processing pipeline.



CBV maps are then calculated on a voxel-wise basis using trapezoidal integration of the leakage-corrected (11) concentration-time curves between entrance and exit bolus time points recalculated for each individual pixel's corresponding concentration curve. The CBV parametric map calculation is highly affected by the selection of these points, as different choices can lead to under- or overestimation of the integration area. The rCBV map is then calculated by dividing all intensities by the




Figure 2. Bland–Altman plots for the mean relative cerebral blood volume (rCBV) values between the Food and Drug Administration (FDA)-approved and the proposed software systems for the mean rCBV measurement. The solid line represents the mean value for the data points and the slashed line represents the 2*SD.

mean intensity of the normal-appearing white matter CBV value.

FDA-Approved Software Utilization

One operator created rCBV images from the DSC-MRIs using IB Neuro ver. 1.1 (Imaging Biometrics, Elm Grove, Wisconsin), as no interaction is required beyond loading the magnetic resonance images. Three operators used FuncTool ver. 4.5.3 (GE Healthcare) and nordicICE ver. 2.3.13 (NordicNeuroLab, Bergen, Norway). An effort was made to operate FuncTool and nordicICE in a similar way. FuncTool required manual selection of the prebolus baseline and integration starting and stopping time points, whereas nordicICE required manual specification of the prebolus baseline only when its automatic selection algorithm failed (7/43 cases). We used default settings for nordicICE and IB Neuro, except that leakage correction was enabled. For FuncTool, the baseline was interpolated between the integration time points. No leakage correction option was available in this version of the FuncTool.

Statistical Methods

We compared our rCBV calculations against the 3 FDA-approved software tools' calculations using Bland–Altman analysis and the intraclass correlation coefficient (ICC). The 1-way random-effects model was used (25). The ICC reveals how strongly units in the same group resemble each other. In comparison with other correlation measures, it treats the data as groups rather than as paired observations. The analysis was performed for the tumor and the WM rCBV values.

RESULTS

The proposed and the FDA-approved software were used to create the CBV maps. Subsequently, the 3 ROIs available (1 automatic and 2 user-defined ROIs) were used to normalize the CBV maps. Subsequently, 2 first-order statistics, the mean value, and the 95th percentile, were calculated for each patient of the data set. Figure 1 illustrates the processing pipeline of the proposed system.




Figure 3. Bland–Altman plots for the 95th percentile rCBV values between the FDA-approved and the proposed software systems for the 95th percentile rCBV measurement. The solid line represents the mean value for the data points and the dashed line represents the 2*SD.




Figure 4. Bland–Altman plots for the mean rCBV values between FDA-approved software requiring user input for the mean rCBV measurement. The solid line represents the mean value for the data points and the dashed line represents the 2*SD.

The proposed system shows excellent agreement with 2 of the 3 FDA-approved software considered in this study. Figures 2 and 3 are the Bland–Altman plots for the difference in measured values between the FDA-approved and the proposed system for the mean and 95th percentile of rCBV, respectively. Based on the plots, the proposed system falls within 2 standard deviations (SDs) of the FDA-approved software systems, which is similar to the agreement between the 2 FDA cleared systems. A larger bias was observed in case of NordicICE compared with our system (0.20 and 1.34 for mean rCBV and 95th percentile rCBV, respectively).

Figures 2 and 3 reveal the difference between the mean and the 95th percentile rCBV in the sensitivity of the imaging biomarkers, as 95th percentile reflects only a very small number of pixels, and thus is more sensitive to local artifacts. Smoothing or low-pass filtering will also affect this measure.

Figures 4 and 5 correspond to Bland–Altman plots for the difference of measured values between the two packages requiring user interaction (mean and 95th rCBV, respectively). In all

cases for the mean rCBV, the bias was <0.03 . The largest 2*SD observed was 0.41 in the case of FuncTool for users 1 and 2. This range is comparable with the range observed between the proposed system and the 3 FDA-approved software. The 95th rCBV and the 2*SD are lower than those observed for the software comparison in all cases.

Figure 6 displays the Bland–Altman plots for rCBV values corresponding to the area of a tumor selected from our data set for all software tools considered in this study. Proportional bias is observed between FuncTool and the proposed method. The existence of proportional bias indicates that the methods do not agree equally through the range of measurements. This can be attributed to the smoothing observed in the CBV maps created by FuncTool (Figure 7). Figures 7 and 8 depict the CBV maps created by the 4 software tools considered in this study for 2 different subjects.

Table 1 summarizes the ICC for both mean and 95th percentile rCBV for all the comparisons considered in this study and




Figure 5. Bland–Altman plots for the 95th percentile rCBV values between FDA-approved software requiring user input for the 95th percentile rCBV measurement. The solid line represents the mean value for the data points and the dashed line represents the 2*SD.





Figure 6. Bland–Altman plots for rCBV values corresponding to a specific tumor between all FDA-approved software systems with different user setup and the proposed system. The solid line represents the mean value for the data points and the dashed line represents the 2*SD.

Figure 7. Cerebral blood volume (CBV) maps created with the software tools used in this study and the corresponding post-contrast T1 image.



all the methods for normalization used in this study. The results from the proposed system correlate better with IBNeuro and FuncTool. FuncTool results are less correlated with IBNeuro, whereas ICC is good between FuncTool and IBNeuro. The ICC for all the comparisons considered between the same software operated by different people was >0.880 with the exception of FuncTool when operated by user 1 versus user 2.

Little variability in agreement between software tools was observed when using different WM selection techniques. All software tools yielded high ICC for both the 95th and mean rCBV metrics for both automated and manual WM selection techniques.

DISCUSSION

Here, we present a method for estimating rCBV metrics from DSC-MRI with an automated WM selection step using a probabilistic atlas to further standardize rCBV calculation. This automated technique was found to be in close agreement with manual WM selection.

Our software is in good agreement with IBNeuro and FuncTool for tumor rCBV measurements of both mean and the 95th percentile values based on ICC. Mean rCBV ICC values are higher than the 95th percentile measures, most likely because the 95th percentile value is more sensitive to outliers, as it reflects a much smaller sample size.

Figure 8. CBV maps created with NordicICE (row 1) and FuncTool (row 2) produced by the 3 operators participating in this study.




Table 1. ICC for the 3 Software and Our Tool

ROI	Comparisons	Tumor-ICC (95 th percentile rCBV)	Tumor-ICC (mean rCBV)	WM-ICC (95 th percentile rCBV)	WM-ICC (mean rCBV)
Automated WM selection	NordicICE vs proposed	0.664	0.770	0.754	0.800
	IBneuro vs proposed	0.868	0.780	0.881	0.823
	FuncTool vs proposed	0.803	0.810	0.795	0.845
	NordicICE vs IBneuro	0.572	0.680	0.657	0.699
	FuncTool vs IBneuro	0.700	0.670	0.745	0.779
	NordicICE (user 1) vs NordicICE (user 2)	0.977	0.900	0.77	0.824
	NordicICE (user 1) vs NordicICE (user 3)	0.949	0.880	0.784	0.794
	FuncTool (user 1) vs FuncTool (user 2)	0.922	0.920	0.754	0.806
	FuncTool (user 1) vs FuncTool (user 3)	0.893	0.830	0.815	0.927
	NordicICE vs FuncTool	0.156	0.160	0.767	0.864
WM user 1	NordicICE vs proposed	0.57	0.515	0.952	0.908
	IBneuro vs PROPOSED	0.745	0.821	0.732	0.632
	FuncTool vs proposed	0.759	0.804	0.933	0.921
	NordicICE vs IBneuro	0.353	0.539	0.743	0.771
	FuncTool vs IBneuro	0.744	0.757	0.738	0.708
	NordicICE (user 1) vs NordicICE (user 2)	0.951	0.939	0.745	0.769
	NordicICE (user 1) vs NordicICE (user 3)	0.958	0.917	0.71	0.83
	FuncTool (user 1) vs FuncTool (user 2)	0.932	0.834	0.79	0.807
	FuncTool (user 1) vs FuncTool (user 3)	0.895	0.894	0.985	0.988
	NordicICE vs FuncTool	0.17	0.305	0.936	0.928
WM user 2	NordicICE vs proposed	0.569	0.606	0.823	0.742
	IBneuro vs proposed	0.754	0.814	0.773	0.605
	FuncTool vs proposed	0.782	0.822	0.787	0.747
	NordicICE vs IBneuro	0.401	0.674	0.767	0.645
	FuncTool vs IBneuro	0.743	0.804	0.685	0.624
	NordicICE (user 1) vs NordicICE (user 2)	0.949	0.941	0.846	0.852
	NordicICE (user 1) vs NordicICE (user 3)	0.944	0.912	0.716	0.836
	FuncTool (user 1) vs FuncTool (user 2)	0.724	0.639	0.877	0.854
	FuncTool (user 1) vs FuncTool (user 3)	0.717	0.697	0.99	0.972
	NordicICE vs FuncTool	0.018	0.161	0.863	0.787

Abbreviations: ROI, regions of interest; ICC, intraclass correlation coefficient; WM, white matter; rCBV, relative cerebral blood volume.

User interaction can lead to increased variation in the rCBV measurements (Figures 4 and 5), which is more pronounced in the case of NordicICE.

We compared both the rCBV metrics originating from tumor regions and from WM ROIs, both manually and by automatic selection. The ICC for measures extracted from the tumor ROI for each software comparison for the different WM selection methods was similar, revealing good agreement in the majority of the comparisons. This was not observed when using the ROI selections made by user 2.

Calculation of rCBV for the tumor region is expected to be more complicated because of leakage effects. However, based on our analysis, it seems that WM rCBV values vary by software package. Selection of the normalizing WM ROI is crucial, as it can lead to further increases in the variation of rCBV measurements.

Further investigation is needed with respect to the variation introduced because of the software used for rCBV calculation and

the clinical impact. Recently, Kelm et al. (6) showed that the software used for rCBV calculation can result in very different values, and that across a range of clinically relevant thresholds, clinical decisions will be different in a large fraction of patients. Hu et al. (7), recently, compared IBNeuro and NordicICE modeling methods in a cohort of 52 patients with glioma. They found that IBNeuro showed significantly better T1w leakage correction compared with NordicICE. Similarly, rCBV, as measured on IBNeuro, showed stronger correlation with image-guided microvessel quantification and higher accuracy in predicting tumor recurrence from pseudoprogression/radiation necrosis, based on validation by surgical histologic tumor quantification. Orsingher et al. (26) also found significant differences in rCBV values when comparing General Electric and NordicICE software packages. The findings of both these studies coincide with the findings of the current study.

One of the limiting factors of this study is the fact that proprietary commercial packages were used and thus many details of the algorithmic implementation were unavailable for explaining the differences observed. The variation in measurements seen is likely due to differences in how the DSC-MRI intensities are modeled. Other limitations of this study include the use of spin echo technique with less than a single-dose bolus injection and low amounts of preload dose. Gradient-echo T2*-weighted DSC represents the more commonly used and widely published method for DSC. Although spin echo T2-weighted DSC offers a higher signal-to-noise ratio and fewer susceptibility artifacts, double or triple doses of contrast bolus may be required to achieve sufficient bolus response in every case (7).

In addition, the images used for this work were obtained on just one vendor platform (a limitation imposed by one of the

software packages we compared). However, we expect that including more scanner vendor platforms would likely increase the variance between software.

The proposed algorithm and the analysis was performed using Python (Python, Python Software Foundation: v.2.7.4, numpy: v.1.8.1, scipy: v.0.13.3, matplotlib: v1.5.0).

CONCLUSION

We describe an automatic algorithm for rCBV calculation that includes leakage correction and an automated step for WM selection. The algorithm was evaluated against 3 FDA-approved software packages with respect to tumor and WM ROIs. The proposed system has good agreement with 2 out of 3 software tools considered.

ACKNOWLEDGMENTS

PK and BJE are supported by the National Cancer Institute (NCI) under Grant/Award Number CA160045. TLK is supported by the National Institute of Diabetes and Digestive

and Kidney Diseases (NIDDK) under Grant/Award Number P30 DK090728 and the PKD Foundation under Grant Number 206g16a.

REFERENCES

- Boxerman JL, Ellingson BM, Jeyapalan S, Elinzano H, Harris RJ, Rogg JM, Pope WB, Safran H. Longitudinal DSC-MRI for distinguishing tumor recurrence from pseudoprogression in patients with a high-grade glioma. *Am J Clin Oncol*. 2014. [Epub ahead of print]. doi: 10.1097/COC.0000000000000156
- Young RJ, Gupta A, Shah AD, Graber JJ, Chan TA, Zhang Z, Shi W, Beal K, Omuro AM. MRI perfusion in determining pseudoprogression in patients with glioblastoma. *Clin Imaging*. 2013;37(1):41–49.
- Law M, Yang S, Babb JS, Knopp EA, Golfinos JG, Zagzag D, Johnson G. Comparison of cerebral blood volume and vascular permeability from dynamic susceptibility contrast-enhanced perfusion MR imaging with glioma grade. *AJNR Am J Neuroradiol*. 2004;25(5):746–755.
- Law M, Young RJ, Babb JS, Peckerelli N, Chheang S, Gruber ML, Miller DC, Golfinos JG, Zagzag D, Johnson G. Gliomas: predicting time to progression or survival with cerebral blood volume measurements at dynamic susceptibility-weighted contrast-enhanced perfusion MR imaging. *Radiology*. 2008;247(2):490–498.
- Weber MA, Zoubaa S, Schlieter M, Jüttler E, Huttner HB, Geletnek K, Ittrich C, Lichy MP, Kroll A, Debus J, Giesel FL, Hartmann M, Essig M. Diagnostic performance of spectroscopic and perfusion MRI for distinction of brain tumors. *Neurology*. 2006;66(12):1899–1906.
- Kelm ZS, Korfiatis PD, Lingineni RK, Daniels JR, Buckner JC, Lachance DH, Parney IF, Carter RE, Erickson BJ. Variability and accuracy of different software packages for dynamic susceptibility contrast magnetic resonance imaging for distinguishing glioblastoma progression from pseudoprogression. *J Med Imaging (Bellingham)*. 2015;2(2):026001.
- Hu LS, Kelm Z, Korfiatis P, Dueck AC, Elrod C, Ellingson BM, Kaufmann TJ, Eschbacher JM, Karis JP, Smith K, Nakaji P, Brinkman D, Pafundi D, Baxter LC, Erickson BJ. Impact of software modeling on the accuracy of perfusion MRI in glioma. *AJNR Am J Neuroradiol*. 2015;36(12):2242–2249.
- Prah MA, Stufflebeam SM, Paulson ES, Kalpathy-Cramer J, Gerstner ER, Batchelor TT, Barboriak DP, Rosen BR, Schmainda KM. Repeatability of standardized and normalized relative CBV in patients with newly diagnosed glioblastoma. *AJNR Am J Neuroradiol*. 2015;36(9):1654–1661.
- Hu LS, Baxter LC, Pinnaduwa DS, Paine TL, Karis JP, Feuerstein BG, Schmainda KM, Dueck AC, Debbins J, Smith KA, Nakaji P, Eschbacher JM, Coons SW, Heiserman JE. Optimized preload leakage-correction methods to improve the diagnostic accuracy of dynamic susceptibility-weighted contrast-enhanced perfusion MR imaging in posttreatment gliomas. *AJNR Am J Neuroradiol*. 2010;31(1):40–48.
- Boxerman JL, Prah DE, Paulson ES, Machan JT, Bedekar D, Schmainda KM. The Role of preload and leakage correction in Gd-based cerebral blood volume estimation determined by comparison with MION as a criterion standard. *AJNR Am J Neuroradiol*. 2012;33(6):1081–1087.
- Boxerman JL, Schmainda KM, Weisskoff RM. Relative cerebral blood volume maps corrected for contrast agent extravasation significantly correlate with glioma tumor grade, whereas uncorrected maps do not. *Am J Neuroradiol*. 2006;27(4):859–867.
- Essock-Burns E, Phillips JJ, Molinaro AM, Lupo JM, Cha S, Chang SM, Nelson SJ. Comparison of DSC-MRI post-processing techniques in predicting microvascular histopathology in patients newly diagnosed with GBM. *J Magn Reson Imaging*. 2013;38(2):388–400.
- Johnson G, Wetzel SG, Cha S, Babb J, Tofts PS. Measuring blood volume and vascular transfer constant from dynamic, T(2)*-weighted contrast-enhanced MRI. *Magn Reson Med*. 2004;51(5):961–968.
- Paulson ES, Schmainda KM. Comparison of dynamic susceptibility-weighted contrast-enhanced MR methods: recommendations for measuring relative cerebral blood volume in brain tumors. *Radiology*. 2008;249(2):601–613.
- Jafari-Khouzani K, Emblem KE, Kalpathy-Cramer J, Bjørnerud A, Vangel MG, Gerstner ER, Schmainda KM, Paynabar K, Wu O, Wen PY, Batchelor T, Rosen B, Stufflebeam SM. Repeatability of cerebral perfusion using dynamic susceptibility contrast MRI in glioblastoma patients. *Transl Oncol*. 2015;8(3):137–146.
- Meijs M, Christensen S, Lansberg MG, Albers GW, Calamante F. Analysis of perfusion MRI in stroke: to deconvolve, or not to deconvolve. *Magn Reson Med*. 2016;76(4):1282–1290.
- Emblem KE, Bjørnerud A. An automatic procedure for normalization of cerebral blood volume maps in dynamic susceptibility contrast-based glioma imaging. *AJNR Am J Neuroradiol*. 2009;30(10):1929–1932.
- Otsu N. A threshold selection method from gray-level histograms. *IEEE Trans Syst Man Cybern Syst*. 1979;9(1):62–66.
- Willats L, Calamante F. The 39 steps: evading error and deciphering the secrets for accurate dynamic susceptibility contrast MRI. *NMR Biomed*. 2013;26(8):913–931.
- Comaniciu D, Meer P. Mean shift: a robust approach toward feature space analysis. *IEEE Trans Pattern Anal Mach Intell*. 2002;24(5):603–619.
- Mayer A, Greenspan H. An adaptive mean-shift framework for MRI brain segmentation. *IEEE Trans Med Imaging*. 2009;28(8):1238–1250.
- Fonov V, Evans AC, Botteron K, Almli CR, McKinstry RC, Collins DL; Brain Development Cooperative Group. Unbiased average age-appropriate atlases for pediatric studies. *Neuroimage*. 2011;54(1):313–327.
- Avants BB, Tustison NJ, Song G, Cook PA, Klein A, Gee JC. A reproducible evaluation of ANTs similarity metric performance in brain image registration. *Neuroimage*. 2011;54(3):2033–2044.
- Yushkevich PA, Piven J, Hazlett HC, Gimpel Smith R, Ho S, Gee JC, Gerig G. User-guided 3D active contour segmentation of anatomical structures: significantly improved efficiency and reliability. *Neuroimage*. 2006;31(3):1116–1128.
- Shrout PE, Fleiss JL. Intraclass correlations: uses in assessing rater reliability. *Psychol Bull*. 1979;86(2):420–428.
- Orsingher L, Piccinini S, Crisi G. Differences in dynamic susceptibility contrast MR perfusion maps generated by different methods implemented in commercial software. *J Comput Assist Tomogr*. 2014;38(5):647–654.