

*Abstract*

# Analysis of SPR Sensors in D-Shaped POF Realized by Hand and Mechanical Polishing <sup>†</sup>

Letizia De Maria <sup>1,\*</sup>, Nunzio Cennamo <sup>2</sup>, Francesco Mattiello <sup>2</sup>, Cristina Chemelli <sup>1</sup>, Maria Pesavento <sup>3</sup>, Simone Marchetti <sup>3</sup> and Luigi Zeni <sup>2</sup>

<sup>1</sup> RSE Research on Energetic System, via Rubattino 54, 20134 Milan, Italy; cristina.chemelli@rse-web.it

<sup>2</sup> Department of Industrial and Information Engineering, University of Campania Luigi Vanvitelli, via Roma 29, 81031 Aversa (CE), Italy; nunzio.cennamo@unicampania.it (N.C.); f.mattiello@unicampania.it (F.M.); luigi.zeni@unicampania.it (L.Z.)

<sup>3</sup> Department of Chemistry, University of Pavia, via Taramelli 12, 27100 Pavia, Italy; maria.pesavento@unipv.it (M.P.); simone.marchetti@unipv.it (S.M.)

\* Correspondence: letizia.demaria@rse-web.it

<sup>†</sup> Presented at the 5th International Symposium on Sensor Science (I3S 2017), Barcelona, Spain, 27–29 September 2017.

Published: 29 November 2017

Optical fiber sensors based on Surface Plasmon Resonance are today widely proposed for applications in different areas of bio-chemical and chemical sensing. Among them, SPR sensors based on plastic optical fibers (POFs) can offer advantages due to their low cost, flexibility, robustness and simplicity of fabrication. Recently, authors reported several bio-chemical and chemical applications based on an SPR sensor platform in D-shaped POFs. The adopted D-shaped fabrication procedure is based on the insertion of a portion of the POF fiber (about 10 mm long) in a resin support and on a successive hand grinding of the fiber surface. This procedure guarantees an easy and low-cost effective way for the removal of the cladding layer in the POF sensing region and for the reduction of the exposed core. At the end, a thin gold film (60 nm thick) can be deposited on the flat hand grinded D-shaped POF core, for exciting SPR resonance at the metal/external medium interface. Despite the straightforwardness of the hand polishing method, described above, in this work we want to compare performances of a hand polished SPR D-shaped POF platform with the ones achieved by a mechanical polishing procedure. Actually, the performances are influenced by small variations in the morphology of the D-shaped region (roughness and total depth) resulting from the manual process used for the preparation of the sensors.

**Acknowledgements:** This work has been financed by the Research Fund for the Italian Electrical System under the Contract Agreement between RSE and the Ministry of Economic Development-General Directorate for Energy and Mining Resources stipulated on 29 July 2009 in compliance with the Decree of 19 March 2009.



© 2017 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).