

Article

Development of a Variable-Temperature Mobile NMR Instrument for Applications in Food Science, Polymer Science and Geology

David Pickup^{1,2}  and J. Beau W. Webber^{3,*} 

¹ School of Engineering, Mathematics and Physics, University of Kent, Canterbury CT2 7NH, UK; dmp@kent.ac.uk or dave.pickup@krimsontechnology.com

² Krimson Technology, Blean Common, Canterbury CT2 9JJ, UK

³ Lab-Tools Ltd. (Nano-Science), Marlowe Innovation Centre, Marlowe Way, Ramsgate CT12 6FA, UK

* Correspondence: beau@lab-tools.co.uk; Tel.: +44-(0)-7805-437-241

Abstract

This article describes the development of a compact and affordable variable-temperature NMR instrument designed primarily to measure dynamic molecular motion in solids and liquids. The instrument consists of Lab-Tools' Mk4 palm-top time-domain NMR spectrometer fitted with a Peltier-cooled variable-temperature probe inside a shimmed Halbach magnet. Measurement of NMR relaxation times T_1 , T_2 , and $T_1\rho$ is possible over the temperature range -20°C to 70°C with cooling and heating rates, and data acquisition is controlled from an integrated mini-PC. The overall footprint of the instrument is roughly that of a shoe box, making both in-the-field and bench-top measurements possible. Applications of this instrument include measuring pore-size distribution in porous rocks, the viscosity of oils and tars trapped in porous rock, the properties of polymers, and the viscosity of the liquid components of foods (e.g. fruits, vegetables and seeds). Results of test measurements for calibrated oils and olive oil are presented together with measurements of molecular mobility in a solid polymer.

Keywords: mobile nmr; materials science; foodstuff; oil properties; porous rock; seeds; nmr relaxation; time-domain nmr; td nmr

1. Introduction

Time-domain NMR (TD-NMR) provides direct information on physical motion in samples, allowing properties such as mobility within a sample, dynamics, viscosity and rigidity to be probed [1–13].

Most NMR data is first measured in the time-domain. For chemical spectra this data is then Fourier-transformed to give spectral chemical data; however for study of the above physical properties, it is often more advantageous to remain in the time domain.

A simple explanation of this sensitivity to motion in a sample is that the magnetic field experienced by a nucleus is modified by the magnetic moments of adjacent nuclei, changing the rate at which it precesses. In a rigid sample, this perturbation of the external field has a long time to act, and causes the different nuclei in the sample to precess at different rates and rapidly fall out of step with one another. This results in a net rotating magnetic field that decays quickly, possibly on the order of tens of μs . In a sample with a significant degree of molecular motion, for example at higher temperatures, the perturbations to the



Academic Editor: Marcello Locatelli

Received: 7 May 2026

Revised: 8 June 2026

Accepted: 10 June 2026

Published: 15 June 2026

Copyright: © 2026 by the authors.

Licensee MDPI, Basel, Switzerland.

This article is an open access article distributed under the terms and conditions of the [Creative Commons Attribution \(CC BY\)](https://creativecommons.org/licenses/by/4.0/) license.

fixed external field start averaging out over an NMR rotational period. This is known as “Motional Narrowing” and results in a longer T_2 decay time, extending up to seconds [1].

Many of the NMR material science measurements that make use of the effect described above become significantly more informative if carried out as a function of temperature. In particular, the temperature dependence of the spin–lattice (T_1) and spin-lock ($T_{1\rho}$) relaxation times reveals characteristic minima that correspond to specific correlation times of molecular motion. Concurrently, the spin–spin (T_2) relaxation time provides a probe of bulk mobility, eventually reaching a rigid-lattice limit (or ‘floor’) at low temperatures as stochastic motions are frozen out. Collectively, these parameters offer a comprehensive mapping of the motional dynamics within the sample [1].

Furthermore, NMR cryoporometry (NMRC), a theory which is discussed elsewhere [10,14–17], is a powerful variable-temperature technique for the measurement of pore-size distributions in the range 1 nm to 1 μm , and total porosities. NMRC uses the isobaric (constant pressure) variant of the Gibbs equation, the Gibbs–Thomson equation, which describes the melting point depression of liquids confined in pores [14]. Pore-size distributions can also be measured by BET (Brunauer–Emmett–Teller) gas adsorption experiments which use the isothermal variant of the Gibbs equation, the Kelvin equation. A key advantage of NMRC is its applicability to ‘wet’ or hydrated systems, such as wet clays, where vacuum-based gas adsorption methods fail.

Here we discuss the development of a compact and affordable bench-top variable-temperature TD-NMR instrument designed primarily to measure dynamic molecular motions in solids and liquids. We have recently demonstrated that measurements of $T_{1\rho}$ can be used to determine the viscosity of liquid components in situ [18]. Here we present variable-temperature measurements on the dynamics of a solid polymer, and apply the viscosity measurements to two liquid oils of significantly different viscosities.

2. Materials and Methods

2.1. Instrumentation

The instrument, shown in Figure 1, consists of a Lab-Tools Ltd. (Ramsgate, UK), MK4 NMR spectrometer, a thermo-electrically cooled variable-temperature probe, and a Lab-Tools Ltd. (Ramsgate, UK) shimmed 0.5 T Halbach magnet. The NMR spectrometer and data acquisition are controlled via a short USB cable by a GMKtec mini-PC NucBox G5 running Windows 11. The thermo-electric cooling is powered by a M5Stack programmable power supply (PPS), which is controlled by an ESP32 microcontroller that, in turn, receives commands from the mini-PC. The direction of the Peltier current can be reversed so that the sample can be heated as well as cooled. This is achieved by four semi-conductors in an H-bridge circuit, at the output of the PPS. Temperature is monitored by a T-type (Cu/Co) thermocouple. The programmable power-supply, microcontroller, H-bridge, and thermocouple amplifier are all housed in a single control box. As can be seen in Figure 1, the NMR spectrometer, mini-PC, variable-temperature probe, and temperature control box all slot into recesses in a 3D-printed base. This ensures that the whole system has a compact footprint and keeps all the components in their respective positions during transportation.

2.1.1. MK4 0 MHz to 50 MHz Palm-Top NMR Spectrometer

The Lab-Tools MK4 NMR spectrometer is a compact low-cost palm-top instrument [19]. It is primarily a time-domain NMR relaxation spectrometer with digital radio-frequency (RF) processing. Paired with 0.5 T Halbach magnet, the ^1H frequency is about 21 MHz. Figure 2 illustrates the compact size of the MK4 spectrometer together with an NMR probe and a 0.5 T Halbach magnet. The ^1H NMR $\pi/2$ RF pulse width was 7.5 μs for these measurements.




Figure 1. Variable-temperature mobile NMR instrument: (a) front view and (b) back view.

Instrument control and data acquisition is performed over Ethernet using a GMKtec NucBox G5, Intel 12th Gen Alder Lake N97 (3.3 GHz) Windows 11 mini-PC, with the data-array processing, graphical user interface (GUI), and communications all programmed in MicroApl's AplX [20,21].




Figure 2. (a) Lab-Tools MK4 NMR Spectrometer and probe, and (b) 0.5 T Halbach magnet.

2.1.2. Variable-Temperature Probe

Thermo-electric cooling and heating are achieved by passing a current through a dual-layer 51.6 W 40 mm × 40 mm Peltier element in contact with a copper plate that is part of the cold finger leading to the NMR sample. A thermocouple sensor is mounted on the cold finger close to the sample to monitor the sample's temperature. This probe design is derived from my earlier patented design in Section 5. Waste heat from the Peltier element is removed by a 90 mm × 89 mm CPU cooler with a thermal power design (TDP) of 155 W. Temperature control is afforded by a proportional–integral–derivative controller (PID) varying the current through the Peltier element. Using this approach, sample temperatures from −20 °C to 70 °C can be attained.

Samples in 5 mm NMR tubes inserted into TX and RX coils situated in the brass can be attached to the cold finger. By adjusting the variable capacitor on the MK4 spectrometer, the probe has a tuning frequency range of 20.4 MHz to 21.2 MHz allowing both ^1H and ^{19}F nuclei to be routinely studied. By adding an internal capacitor of the correct size to the spectrometer, ^7Li , ^{23}Na , ^{11}B , and ^{31}P become accessible.

2.1.3. Temperature Control Box

The temperature of the variable-temperature probe is controlled by an ESP32 micro-controller (MCU). This MCU monitors the sample temperature via a T-type thermocouple in combination with a single IC that amplifies the signal and converts the readings to serial data. This IC has a temperature resolution of less than 0.01 °C and an error of $\pm 0.3\%$. The MCU receives the desired target temperature from the mini-PC and uses a PID to determine how much current to pass through the Peltier element to reach this set-point. The M5Stack programmable power supply (PPS) receives commands from the MCU and supplies current to the Peltier accordingly. An H-bridge, again controlled by the MCU, allows the direction of the current through the Peltier element to be reversed, providing cooling or heating of the sample as required.

The fan of the CPU cooler helps to remove waste heat. The speed of this fan is controlled from the mini-PC via the MCU and can be varied from 0 to 6000 RPM in 255 increments.

The MCU, PPS, thermocouple IC and H-bridge are all housed in a custom 3D-printed box measuring 110 mm × 70 mm × 70 mm. Panel and PCB-mounted sockets allow this box to be readily unplugged from the other components if required. This control box is also fitted with a 0.42" OLED display which shows the PPS output voltage and current.

2.1.4. Comparison with Other Similar Time-Domain Benchtop NMR Spectrometers

Figures 5 and 6 in a prior publication [22] contain tabular comparison of time-domain NMR spectrometers.

Here we provide comparative data for the three spectrometers currently available from Lab-Tools Ltd.: Table 1.

2.1.5. Instrument Control and Use

The MK4 NMR spectrometer is controlled by the mini-PC and operated via a graphical user interface (GUI) [23]. This GUI allows parameters such as NMR Pulse Sequence, π and $\pi/2$ pulse widths, τ pulse separation and number of repetitions to be set. An example GUI configuration is shown in Figure 3.

Beyond setting the parameters for data acquisition, the GUI allows users to average, plot and process data [23]. An example of the latter would be the single-component fitting of a Free Induction Decay (FID) with a polynomial, exponential or Gaussian function to determine T_2 .

Table 1. Summary table of properties for these three time-domain NMR spectrometers.

Name	Maker	NMR Frequency	Time-Domain	Frequency-Domain	Rigid Solids	Solid Polymers	Liquids	T1Rho	Special Features	Size/Weight
MK4	Lab-Tools accessed on 5 May 2026. www.labtoolsnanoscience.co.uk/ (accessed on 24 February 2026)	Broadband 3 to 50 MHz	Yes ¹ H, ¹⁹ F, ¹¹ B, ²³ Na, ³¹ P, ⁷ Li	¹⁹ F only	Yes	Yes	Yes	Yes	Fast-recovery Large bore magnets for V-T probes, etc.	Palm top, Mobile. Spectrometer: 1 kg Magnets: <1 kg 0.3 T, 1.5 kg 0.5 T, 7 kg 0.33 T
MK4-VT	Lab-Tools www.labtoolsnanoscience.co.uk/ (accessed on 24 February 2026)	Broadband 3 to 50 MHz	Yes ¹ H, ¹⁹ F, ¹¹ B, ²³ Na, ³¹ P, ⁷ Li	¹⁹ F only	Yes	Yes	Yes	Yes	V-T −20 °C to 70 °C Includes Win 11 computer	Shoebox, Mobile 4 kg inc Magnet: 0.5 T
MK4-CryoP8	Lab-Tools www.labtoolsnanoscience.co.uk/ (accessed on 24 February 2026)	Broadband 3 to 50 MHz	Yes ¹ H, ¹⁹ F, ¹¹ B, ²³ Na, ³¹ P, ⁷ Li	¹⁹ F only	Yes	Yes	Yes	Yes	V-T −60 °C to 90 °C NMR Cryoporometric pore sizes. Large bore Magnets for V-T probes, etc.	Benchtop 15 kg plus Recirculating chiller Magnets 0.33 T or 0.5 T




Figure 3. A typical graphical user interface front panel for a CPMG NMR pulse sequence.

As well as displaying data, instrument status and acquisition parameters on a monitor attached to the mini-PC, the instrument is also fitted with a 7" IPS-LCD screen to display data from any measurements currently in progress. The presence of this IPS-LCD screen and the OLED display on the temperature control box are particularly useful for checking the status of the instrument if it is being operated remotely over the internet and for trouble-shooting.

2.2. Performance Testing and Example Data Acquisition

2.2.1. Measuring T_2 to Determine the Phase Composition of Nylon 66

The T_2 NMR relaxation times of a solid Nylon 66 sample were measured over the temperature range 10 °C to 70 °C, at about 5 °C intervals, in order to follow the evolution of the phase composition with temperature [24].

For the solid Nylon 66 sample, capturing the full signal decay required a Carr–Purcell–Meiboom–Gill (CPMG) NMR pulse sequence over 4 ms, with a Tau setting of 25 μ s, resulting in a 50 μ s echo spacing. An averaged trace for the fitted echo peaks for a sample at about 61 °C can be seen in Figure 4 (dots). Fitting this trace with exponentials using the MK4 built-in two-component fitting routine gave a reasonable fit, but neither signal-to-noise nor fitting precision explained the fit imperfections. Thus a three-component off-line routine was evaluated; Figure 4 shows the automatically selected optimum fit of two short Gaussians and one longer exponential. This three-component fit was very stable, and will be incorporated as a built-in fitting routine in the future.

2.2.2. Measuring $T_1\rho$ to Determine the Viscosity of Oils

Spin-lock relaxation times ($T_1\rho$) of 4 mineral oils of known viscosity (Brookfield Engineering) were measured at room temperature to verify that we could correctly determine their viscosities [18]. $T_1\rho$ relaxation times of one of these oils (Brookfield Engineering No. B1060) and a sample of Greek olive oil were then measured at temperatures of 10 °C to 70 °C to monitor how their viscosities changed as they were heated.




Figure 4. Representative three-component fit of a CPMG echo envelope for solid Nylon 66 sample. The multi-component decay is resolved into three distinct T_2 relaxation times: a short Gaussian, an intermediate Gaussian and a longer exponential (all thin red lines). Their sum is plotted as a thick red line.

To facilitate the measurement of $T_{1\rho}$ relaxation data with these rapidly decaying FID data, a 2D NMR routine was written in the concise array language APLX. This consisted of a full amplitude 90° pulse (RF pulse width $7.5 \mu\text{s}$), a reduced amplitude long spin-lock pulse, followed by a CPMG sequence. For a 10 kHz spin-lock frequency, 30% of full RF power was used. The Tau pulse separation for the CPMG train was $500 \mu\text{s}$. Spin-lock pulse lengths were limited to about 700 ms maximum due to RF sample heating effects to maintain the sample temperatures at 25°C , and repetition times were increased from a minimum of about 2 s (governed by T_1 of the samples) to 20 s, to also limit sample heating effects. Averaging repetitions per measurement were 30 averages for bulk samples, and 100 averages for confined pore samples. For the variable-temperature measurements, the heating rate was slow at 0.1°C , per minute, with then an equilibration time of 10 min.

An example of a single trace capture can be seen in Figure 5. The initial amplitude of the FID following the long spin-lock pulse was determined using the polynomial fitted and algebraically solved CPMG echo peaks, extrapolated back to the end of the spin-lock pulse, as shown in Figure 6. The 2D routine implemented here provided a significant improvement to the precision and signal-to-noise ratio of these measurements.




Figure 5. The NMR receiver output (with pulse-suppression switched off) for a 2D $T_{1\rho}$ -CPMG NMR sequence.

The calibration curve shown in Figure 7 was used to calculate the viscosities of the oils from the measured $T_{1\rho}$ relaxation times [18].




Figure 6. The polynomial fitted NMR echoes for a 2D $T_1\rho$ CPMG NMR sequence, following the ending of a 5 ms spin-lock pulse.




Figure 7. A plot of $1/\text{viscosity}$ against $T_1\rho$ relaxation time for 10 samples of known viscosity. $T_1\rho$ was measured at a spin-lock frequency of 10 kHz and at or close to 25 °C. These 10 measurements were fitted with a parabola (blue line) with $a_0 = -0.59791291$, $b_0 = 0.98323349$, and $c_0 = -0.18025699$. Blue circles: four Brookfield Engineering oils of calibrated viscosity (Table 2); Cyan circles: five sucrose solutions of known viscosity; and Mauve circle: a dodecane sample of known viscosity. Red circles and red line: four slightly different Brookfield oils previously measured by fast-field-cycling NMR at 40 °C are also plotted [25]. Reproduced from Ref. [18].

Table 2. The $T_1\rho$ relaxation times of a Brookfield Engineering oil of calibrated viscosity at 25 °C (1% quoted accuracy), were measured using the Lab-Tools' MK4 TD NMR spectrometer at a temperature of 25 °C $\pm$ 0.2 °C. The final column uses the calibration curve shown in Figure 7 to evaluate viscosity from the measured $T_1\rho$ relaxation times, as a check on precision.

Oil	Brookfield No.	Viscosity (mPa s) (Brookfield Data)	Measured $T_1\rho$ (ms)	Viscosity from $T_1\rho$ (mPa s)
1	B1060	1.042	15.06	0.97285
2	B10200	10.440	3.726674	11.1973
3	B73000	66.410	1.436256	70.7513
4	B360000	457.10	0.601422	431.182

3. Results and Discussion

3.1. Phase Composition of Nylon 66

The T_2 NMR relaxation times of a solid Nylon 66 sample were measured over the temperature range 10 °C to 70 °C to determine the temperature dependence of the phase composition.

As mentioned in Section 2.2.1, it was necessary to use three components to fit the trace shown in Figure 4. Litvinov and Penning [24] also used a three-component fit in their NMR relaxation study of nylon fibres. Following a similar approach to that used in their work, we used the three distinct T_2 relaxation times obtained from the fitting process to calculate the amounts of the various phase components present at each temperature, where the shortest, intermediate and longest T_2 times correspond to rigid, semi-rigid and soft components, respectively [26]. The results of these calculations are plotted in Figure 8.




Figure 8. Temperature dependence of the phase composition of a solid sample of Nylon 66 determined from T_2 relaxation times.

In accordance with previous studies [24,27], the latter using DSC and X-ray techniques, the rigid, semi-rigid and soft components are assigned to a crystalline phase, a semi-rigid amorphous phase, and a soft amorphous phase, respectively. The results presented show an increase in rotational and translational chain mobility with temperature as the crystalline phase transitions toward more disordered states. This behaviour is facilitated by the short statistical segment of the Nylon 66 chain, approximately 6 backbone bonds [28], which allows for rapid relaxation of conformational constraints outside the crystal lattice. Consequently, the three-phase model provides an accurate description of the temperature-dependent phase composition in Nylon 66.

The results presented here demonstrate that variable-temperature NMR T_2 relaxation analysis using our compact instrumentation is not only a valuable tool for determining the crystallinity of Nylon 66, but also provides a detailed description of the polymer's structure, specifically the relative amounts and molecular mobility of the three distinct phases detected.

3.2. Viscosity of Oils

Table 2 shows the measured spin-lock $T_{1\rho}$ relaxation times and the viscosities calculated from these times [18] for four oils of known viscosity. The viscosities determined here from $T_{1\rho}$ relaxation using our variable-temperature instrument agree well with the known values, with all the experimental results within 8% of their expected values.

Once we had verified that we could measure viscosity to a satisfactory level of accuracy we proceeded with variable-temperature measurements. Figure 9 shows the temperature dependence of the $T_{1\rho}$ relaxation times of Brookfield Oil No. B1060 and Greek olive oil. As expected, for both oils, $T_{1\rho}$ increases with temperature as the molecular mobility increases and the efficiency of the relaxation processes decrease.




Figure 9. An NMR spin-lock measurement of the $T_{1\rho}$ relaxation time of two oils: Red: Brookfield Oil No. B1060 and Blue: Greek olive oil.

Figure 10 shows the temperature dependence of the viscosity of Brookfield Oil No. B1060 and Greek olive oil calculated from the data presented in Figure 9. As expected, the viscosities of both oils decrease with temperature as the molecules have greater thermal energy and are more easily able to overcome the attractive intermolecular forces that pull them together. Figure 11 shows an Arrhenius plot of the viscosity–temperature data. The data points in Figure 11 were fitted with a modified Vogel–Fulcher–Tammann (VFT) equation which describes the temperature-dependent viscosity of oil-based systems [29]:

$$\eta(T) = \eta_{\infty} \exp\{D/[(T/T_0) - 1]\} \quad (1)$$

where η_{∞} is the viscosity limit at high temperature, D is the fragility and T_0 the Kauzmann temperature [30].

The excellent fit of the data in Figure 11 confirms that the temperature-dependent viscosity of both oils is characteristic of complex organic liquids. The fitting process yielded Kauzmann temperatures of $199 \text{ K} \pm 10 \text{ K}$ and $85 \text{ K} \pm 15 \text{ K}$ for Brookfield Oil No. B1060 and Greek olive oil, respectively. Together with the corresponding fragility indices of 2.33 ± 0.5 and 25.2 ± 7.6 , these Kauzmann temperatures highlight clear differences in the molecular mobility of the two liquids [30].




Figure 10. Temperature dependence of the viscosity of the two oils, from NMR $T_{1\rho}$.




Figure 11. Arrhenius plots of the viscosity of the two oils. The lines of best fit are modified Vogel–Fulcher–Tammann functions [29].

The viscosity data presented here demonstrate that our variable-temperature instrument is well-suited to measuring the temperature dependence of liquid viscosity up to 70 °C. Furthermore, the instrument is capable of in situ measurements within solid matrices, such as oils in porous rocks, and can resolve the individual viscosities of multi-component systems, such as the distinct oil and water phases in seeds.

3.3. Other Potential Applications

The cost-effective, compact, and user-friendly bench-top variable-temperature NMR system described here has numerous potential applications in food science, ranging from research and development to quality and process control throughout the supply chain. NMR relaxation is inherently quantitative and highly sensitive to differences in molecular

mobility; consequently, the analysis of relaxation decays provides critical information on diffusion, exchange processes, and compartment sizes. Within this framework, bulk components such as water, lipids, and carbohydrates are readily detectable [31].

Furthermore, the instrument described here is sufficiently compact and has low enough power requirements to be truly transportable and well-suited for field studies [32]. Field applications of TD-NMR are diverse, ranging from quantifying oil–water ratios and liquid characteristics in porous rocks and sol–gel silicas [33] to determining the oil, water, and sugar content of agricultural seeds. Mobile TD-NMR has also proven effective in monitoring the aging of rubbers and polymers, as well as tracking environmental markers such as water purity in marshes, soil moisture, and ice deposition. It has even been successfully used to monitor the growth rates of intact plants.

4. Conclusions

This paper describes the development and application of a compact and affordable variable-temperature time-domain NMR instrument, designed primarily to measure dynamic molecular motions in solids and liquids, and thus the physical properties that depend on these.

Here it has been used to study a solid polymer, Nylon 66, to obtain the quantities of the three distinct phases present: crystalline, rigid amorphous, and soft amorphous, over the temperature range 10 °C to 70 °C. This technique may also be used to study the aging of rubbers and polymers in service. Furthermore such measurement can be carried out in relatively inhomogeneous magnetic fields, reducing the cost of the magnet required.

We have also used the instrument to determine the viscosities of two liquids, again over the temperature range 10 °C to 70 °C, by measuring $T_1\rho$. This instrument is specifically designed for studying the properties of food stuffs, and the design temperature range is well-suited to milk products such as yoghurts and the maturation of cheeses, particularly by NMR $T_1\rho$. Fruits and seeds may also be studied, as well as the viscosity of seed oils, by $T_1\rho$ [18].

An additional application for this instrument is measuring pore dimensions based on the loss of proton spin magnetisation by contact with a pore wall. This method works well for micron-sized pores [6] and is often used for studying sandstone porosities.

While the current instrumentation is not suitable for thermodynamic NMRC measurements of very small or large pore sizes, it is suitable for measuring pores in the mesoporous range (10 nm to 100 nm). This range is highly relevant for the characterisation of many geological formations and reservoir rocks.

5. Patents

Webber, J.B.W. Nuclear Magnetic Resonance Probes. U.S. Patent 9,810,750 B2, 7th November 2017.

Author Contributions: Conceptualization, J.B.W.W.; methodology, J.B.W.W. and D.P.; software, J.B.W.W.; firmware D.P.; validation, J.B.W.W. and D.P.; investigation, J.B.W.W.; resources, J.B.W.W.; data curation, J.B.W.W. and D.P.; writing—original draft preparation, J.B.W.W. and D.P.; writing—review and editing, J.B.W.W. and D.P.; visualization, J.B.W.W. and D.P. All authors have read and agreed to the published version of the manuscript.

Funding: This research received no external funding.

Data Availability Statement: Data is available on request.

Conflicts of Interest: J.B.W. Webber is employed by the company Lab-Tools Ltd. The remaining authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

References

1. Abragam, A. *The Principles of Nuclear Magnetism*; Clarendon Press: Oxford, UK, 1961.
2. Besghini, D.; Mauri, M.; Simonutti, R. Time Domain NMR in Polymer Science: From the Laboratory to the Industry. *Appl. Sci.* **2019**, *9*, 1801. [\[CrossRef\]](#)
3. Vaca Chávez, F.; Saalwächter, K. Time-Domain NMR Observation of Entangled Polymer Dynamics: Universal Behavior of Flexible Homopolymers and Applicability of the Tube Model. *Macromolecules* **2011**, *44*, 1549–1559. [\[CrossRef\]](#)
4. Todt, H.; Guthausen, G.; Burk, W.; Schmalbein, D.; Kamłowski, A. Water/moisture and fat analysis by time-domain NMR. *Food Chem.* **2006**, *96*, 436–440. [\[CrossRef\]](#)
5. Brownstein, K.R.; Tarr, C.E. Importance of classical diffusion in NMR studies of water in biological cells. *Phys. Rev. A* **1979**, *19*, 2446–2453. [\[CrossRef\]](#)
6. Gladden, L.F. Nuclear-Magnetic-Resonance Studies of Porous-Media. *Chem. Eng. Res. Des.* **1993**, *71*, 657–674.
7. Callaghan, P.; Godefroy, S.; Ryland, B. Diffusion–relaxation correlation in simple pore structures. *J. Magn. Reson.* **2003**, *162*, 320–327. [\[CrossRef\]](#)
8. Fantazzini, P.; Brown, R.J.S.; Borgia, G.C. Bone tissue and porous media: Common features and differences studied by NMR relaxation. *Magn. Reson. Imaging* **2003**, *21*, 227–234. [\[CrossRef\]](#)
9. Kimmich, R.; Fatkullin, N.; Mattea, C.; Fischer, E. Polymer chain dynamics under nanoscopic confinements. *Magn. Reson. Imaging* **2005**, *23*, 191–196. [\[CrossRef\]](#)
10. Hansen, E.W.; Fonnum, G.; Weng, E. Pore Morphology of Porous Polymer Particles Probed by NMR Relaxometry and NMR Cryoporometry. *J. Phys. Chem. B* **2005**, *109*, 24295–24303. [\[CrossRef\]](#) [\[PubMed\]](#)
11. McDonald, P.J.; Mitchell, J.; Mulheron, M.; Aptaker, P.S.; Korb, J.P.; Monteilhet, L. Two-dimensional correlation relaxometry studies of cement pastes performed using a new one-sided NMR magnet. *Cem. Concr. Res.* **2007**, *37*, 303–309. [\[CrossRef\]](#)
12. D’Agostino, C.; Mitchell, J.; Mantle, M.D.; Gladden, L.F. Interpretation of NMR Relaxation as a Tool for Characterising the Adsorption Strength of Liquids inside Porous Materials. *Chem.–Eur. J.* **2014**, *20*, 13009–13015. [\[CrossRef\]](#)
13. Webber, J.B.W. Biological, Medical and Nano Structured materials—NMR done Simply. *Arch. Biomed. Eng. Biotechnol. (ABEB)* **2019**, *1*. [\[CrossRef\]](#)
14. Mitchell, J.; Webber, J.B.W.; Strange, J. Nuclear magnetic resonance cryoporometry. *Phys. Rep.* **2008**, *461*, 1–36. [\[CrossRef\]](#)
15. Petrov, O.V.; Furó, I. NMR cryoporometry: Principles, applications and potential. *Prog. Nucl. Magn. Reson. Spectrosc.* **2009**, *54*, 97–122. [\[CrossRef\]](#)
16. Rottreau, T.J.; Parkes, G.E.; Schirru, M.; Harries, J.L.; Granollers Mesa, M.; Topham, P.D.; Evans, R. NMR cryoporometry of polymers: Cross-linking, porosity and the importance of probe liquid. *Colloids Surf. A Physicochem. Eng. Asp.* **2019**, *575*, 256–263. [\[CrossRef\]](#)
17. Webber, J.B.W. Some Applications of a Field Programmable Gate Array Based Time-Domain Spectrometer for NMR Relaxation and NMR Cryoporometry. *Appl. Sci.* **2020**, *10*, 2714. [\[CrossRef\]](#)
18. Webber, J.B.W.; Singer, P.M.; Pickup, D.M. Quantified Measurements of Viscosity in The Bulk and In Pores, Using NMR Spin-lattice Relaxation in The Rotating Frame. *Quant. NMR J.* **2026**, *1*, 19–28. Available online: <https://qnmjournal.com/index.php/qNMR/article/view/8/8> (accessed on 24 February 2026).
19. Webber, J.B.W. Updates on an Even More Compact Precision NMR Spectrometer and a Wider Range V-T Probe, for General Purpose NMR and for NMR Cryoporometric Nano- to Micro-Pore Measurements. *Micro* **2024**, *4*, 509–529. [\[CrossRef\]](#)
20. Iverson, K. *A Programming Language*; Wiley: New York, NY, USA, 1962.
21. APLX Archive. 2026. Available online: <https://www.dyalog.com/aplx.htm> (accessed on 24 February 2026).
22. Webber, J.B.W. A review of the use of simple time-domain NMR/MRI for material-science: Time-domain benchtop and mobile NMR/MRI for liquid and solid material science, and nano- to micro-pore-size measurement. *SN Appl. Sci.* **2021**, *3*, 809. [\[CrossRef\]](#)
23. Webber, J.B.W.; Demin, P. Digitally Based Precision Time-Domain Spectrometer for NMR Relaxation and NMR Cryoporometry. *Micro* **2023**, *3*, 404–433. [\[CrossRef\]](#)
24. Litvinov, V.M.; Penning, J.P. Phase Composition and Molecular Mobility in Nylon 6 Fibers as Studied by Proton NMR Transverse Magnetization Relaxation. *Macromol. Chem. Phys.* **2004**, *205*, 1721–1734. [\[CrossRef\]](#)
25. Singer, P.M.; Valiya Parambathu, A.; Wang, X.; Asthagiri, D.; Chapman, W.G.; Hirasaki, G.J.; Fleury, M. Elucidating the ¹H NMR Relaxation Mechanism in Polydisperse Polymers and Bitumen Using Measurements, MD Simulations, and Models. *J. Phys. Chem. B* **2020**, *124*, 4222–4233. [\[CrossRef\]](#) [\[PubMed\]](#)
26. Bergmann, K.; Schmiedberger, H. A rapid method for the NMR-measurement of two-phase polymers. *Colloid Polym. Sci.* **1980**, *258*, 24–26. [\[CrossRef\]](#)
27. Qiu, W.; Habenschuss, A.; Wunderlich, B. The phase structures of nylon 6.6 as studied by temperature-modulated calorimetry and their link to X-ray structure and molecular motion. *Polymer* **2007**, *48*, 1641–1650. [\[CrossRef\]](#)
28. Heymans, N. A Novel Look at Models for Polymer Entanglement. *Macromolecules* **2000**, *33*, 4226–4234. [\[CrossRef\]](#)
29. Sirota, E.B. The Viscosity of Oils. *Energy Fuels* **2025**, *39*, 14511–14533. [\[CrossRef\]](#)

30. Kauzmann, W. The Nature of the Glassy State and the Behavior of Liquids at Low Temperatures. *Chem. Rev.* **1948**, *43*, 219–256. [[CrossRef](#)]
31. van Duynhoven, J.; Voda, A.; Witek, M.; Van As, H. Chapter 3—Time-Domain NMR Applied to Food Products. In *Annual Reports on NMR Spectroscopy*; Academic Press: Cambridge, MA, USA, 2010; Volume 69, pp. 145–197. [[CrossRef](#)]
32. Webber, J.B.W.; Pickup, D.M. Compact Mobile NMR Spectrometers and Nano-pore NMR Cryoporometers. In Proceedings of the Frontiers of Magnetic Resonance, Southampton, UK, 31 March–1 April 2025.
33. Webber, J.B. Studies of nano-structured liquids in confined geometries and at surfaces. *Prog. Nucl. Magn. Reson. Spectrosc.* **2010**, *56*, 78–93. [[CrossRef](#)]

Disclaimer/Publisher’s Note: The statements, opinions and data contained in all publications are solely those of the individual author(s) and contributor(s) and not of MDPI and/or the editor(s). MDPI and/or the editor(s) disclaim responsibility for any injury to people or property resulting from any ideas, methods, instructions or products referred to in the content.