Abstract

Impedance Characterization of Particles One by One Using a Nanosensor Electronic Platform [†]

Diana Isabel Sandoval Bojorquez ^{1,*} , Eduardo Sergio Oliveros Mata ², Julian Schütt ², Michael Bachmann ¹ and Larysa Baraban ¹

¹ Helmholtz-Zentrum Dresden-Rossendorf, Institute of Radiopharmaceutical Cancer Research, 01328 Dresden, Germany; m.bachmann@hzdr.de (M.B.); larysa.baraban@tu-dresden.de (L.B.)

² Helmholtz-Zentrum Dresden-Rossendorf, Institute of Ion Beam Physics and Materials Research, 01328 Dresden, Germany; e.oliveros-mata@hzdr.de (E.S.O.M.); j.schuetz@hzdr.de (J.S.)

* Correspondence: d.sandoval-bojorquez@hzdr.de

[†] Presented at the 8th International Symposium on Sensor Science, 17–28 May 2021;

Available online: <https://i3s2021dresden.sciforum.net/>.

Citation: Sandoval Bojorquez, D.I.; Oliveros Mata, E.S.; Schütt, J.; Bachmann, M.; Baraban, L. Impedance Characterization of Particles One by One Using a Nanosensor Electronic Platform. *Eng. Proc.* **2021**, *6*, 24. <https://doi.org/10.3390/I3S2021Dresden-10110>

Academic Editors: Gianaurelio Cuniberti and Larysa Baraban

Published: 17 May 2021

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Copyright: © 2021 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

Abstract: Impedance cytometry represents a technique that allows the electronic characterization of colloids and living cells in a highly miniaturized way. In contrast with impedance spectroscopy, the measurements are performed at a fixed frequency, providing real-time monitoring of the species traveling over the sensor. By measuring the electrical properties of particles in suspension, the dielectric characteristics (electric conductivity and capacitance) of both cells and particles can be readily determined. During the last years, this technique has been broadly investigated; however, it is still not trivial to differentiate particles of similar size based on their dielectric characteristics. A way to increase the discrimination abilities of this technique could be the integration of nanostructures into the impedance platforms. In this work, we present the impedance cytometry study of particles using microfluidic channels aligned over interdigitated gold nanowire structures as our impedimetric sensor. The characterization of particles of different sizes and their comparison with particles of different compositions will provide an understanding of the correlation between the electrical signal and the characteristics of each particle. This approach is an attractive element for label-free detection platforms that can be integrated into lab-on-a-chip systems and further implemented for single-cell analysis.

Keywords: impedance cytometry; nanosensor; gold nanowires; label-free detection

Supplementary Materials: The following are available online at <https://www.mdpi.com/article/10.3390/I3S2021Dresden-10110/s1>.

Conflicts of Interest: The authors declare no conflict of interest.