

**Table S1.** Comparison of testing methods for the determination of susceptibility of challenge as well as ESBL-negative and carbapenem-susceptible isolates to ceftazidime/avibactam according to the current EUCAST breakpoints.

Organism (No of isolates)	Assay	Resistance rate	MIC (mg/L)		
			Range	MIC <sub>50</sub>	MIC <sub>90</sub>
<b><i>Enterobacteriales</i> (n = 83)</b>					
<i>K. pneumoniae</i> (n = 57)		23%	≤0.125 - >16	1	>16
Challenge (n = 44)		30%	≤0.125 - >16	1	>16
ESBL-negative, carbapenem-susceptible (n = 13)		36%	≤0.125 - >16	1	>16
<i>E. coli</i> (n = 19)		0%	≤0.125 - 0.5	≤0.125	0.5
Challenge (n = 11)	<b>BMD</b>	5%	≤0.125 - >16	≤0.125	0.25
ESBL-negative, carbapenem-susceptible (n = 8)		9%	≤0.125 - >16	≤0.125	0.25
<i>E. cloacae</i> (n = 7)		0%	≤0.125 - ≤0.125	≤0.125	≤0.125
Challenge (n = 3)		29%	≤0.125 - >16	≤0.125	>16
ESBL-negative, carbapenem-susceptible (n = 4)		67%	0.25 - >16	>16	>16
<i>E. cloacae</i> (n = 7)		0%	≤0.125 - ≤0.125	≤0.125	≤0.125
<b><i>Enterobacteriales</i> (n = 83)</b>		22%	≤0.125 - >8	0.5	>8
<i>K. pneumoniae</i> (n = 57)		28%	≤0.125 - >8	1	>8
Challenge (n = 44)		34%	≤0.125 - >8	1	>8
ESBL-negative, carbapenem-susceptible (n = 13)		0%	≤0.125 - 0.5	≤0.125	0.5
<i>E. coli</i> (n = 19)	<b>Vitek 2</b>	5%	≤0.125 - >8	≤0.125	0.25
Challenge (n = 11)		9%	≤0.125 - >8	≤0.125	≤0.125
ESBL-negative, carbapenem-susceptible (n = 8)		0%	≤0.125 - 0.25	≤0.125	0.25
<i>E. cloacae</i> (n = 7)		29%	≤0.125 - >8	0.25	>8
Challenge (n = 3)		67%	0.5 - >8	>8	>8
ESBL-negative, carbapenem-susceptible (n = 4)		0%	≤0.125 - 0.25	0.25	0.25
<b><i>Enterobacteriales</i> (n = 83)</b>		23%	0.03 - >256	1	>256
<i>K. pneumoniae</i> (n = 57)		30%	0.125 - >256	1	>256
Challenge (n = 44)		36%	0.125 - >256	2	>256
ESBL-negative, carbapenem-susceptible (n = 13)		0%	0.125 - 1	0.125	1
<i>E. coli</i> (n = 19)	<b>Etest</b>	5%	0.03 - >256	0.125	0.25
Challenge (n = 11)		9%	0.06 - >256	0.25	0.5
ESBL-negative, carbapenem-susceptible (n = 8)		0%	0.03 - 0.125	0.06	0.125
<i>E. cloacae</i> (n = 7)		29%	0.25 - >256	0.25	>256
Challenge (n = 3)		67%	1 - >256	>256	>256
ESBL-negative, carbapenem-susceptible (n = 4)		0%	0.25 - 0.25	0.25	0.25
<b><i>P. aeruginosa</i> (n = 17)</b>		29%	1 - >16	2	>16
Challenge (n = 11)	<b>BMD</b>	45%	2 - >16	8	>16
ESBL-negative, carbapenem-susceptible (n = 6)		0%	1 - 4	2	4
<b><i>P. aeruginosa</i> (n = 17)</b>		29%	1 - >8	2	>8
Challenge (n = 11)	<b>Vitek 2</b>	45%	2 - >8	8	>8
ESBL-negative, carbapenem-susceptible (n = 6)		0%	1 - 2	2	2
<b><i>P. aeruginosa</i> (n = 17)</b>		24%	1 - 256	2	128
Challenge (n = 11)	<b>Etest</b>	36%	2 - 256	4	128
ESBL-negative, carbapenem-susceptible (n = 6)		0%	1 - 2	2	2

BMD: broth microdilution

The European Committee on Antimicrobial Susceptibility Testing (EUCAST). Breakpoint tables for interpretation of MICs and zone diameters, Version 12.0, 2022.

**Table S2.** Comparison of testing methods for the determination of susceptibility of challenge as well as ESBL-negative and carbapenem-susceptible isolates to ceftolozane/tazobactam according to the current EUCAST breakpoints.

Organism (No of isolates)	Assay	Resistance rate	MIC (mg/L)		
			Range	MIC <sub>50</sub>	MIC <sub>90</sub>
<b><i>Enterobacteriales</i> (n = 83)</b>					
<i>K. pneumoniae</i> (n = 57)		48%	≤0.25 - >32	1	>32
Challenge (n = 44)		69%	≤0.25 - >32	>32	>32
ESBL-negative, carbapenem-susceptible (n = 13)		84%	≤0.25 - >32	>32	>32
<i>E. coli</i> (n = 19)		0%	≤0.25 - 0.5	≤0.25	0.5
Challenge (n = 11)	<b>BMD</b>	5%	≤0.25 - >32	≤0.25	0.5
ESBL-negative, carbapenem-susceptible (n = 8)		9%	≤0.25 - >32	≤0.25	1
<i>E. cloacae</i> (n = 7)		0%	≤0.25 - ≤0.25	≤0.25	≤0.25
Challenge (n = 3)		29%	≤0.25 - >32	≤0.25	>32
ESBL-negative, carbapenem-susceptible (n = 4)		67%	0.5 - >32	>32	>32
<i>E. cloacae</i> (n = 7)		0%	≤0.25 - ≤0.25	≤0.25	≤0.25
<b><i>Enterobacteriales</i> (n = 83)</b>					
<i>K. pneumoniae</i> (n = 57)		48%	≤0.25 - >16	0.5	>16
Challenge (n = 44)		67%	≤0.25 - >16	>16	>16
ESBL-negative, carbapenem-susceptible (n = 13)		82%	≤0.25 - >16	>16	>16
<i>E. coli</i> (n = 19)		0%	≤0.25 - 0.5	≤0.25	0.5
Challenge (n = 11)	<b>Vitek 2</b>	5%	≤0.25 - >16	≤0.25	0.5
ESBL-negative, carbapenem-susceptible (n = 8)		9%	≤0.25 - >16	≤0.25	1
<i>E. cloacae</i> (n = 7)		0%	≤0.25 - ≤0.25	≤0.25	≤0.25
Challenge (n = 3)		29%	≤0.25 - >16	≤0.25	16
ESBL-negative, carbapenem-susceptible (n = 4)		67%	1 - >16	16	>16
<i>E. cloacae</i> (n = 7)		0%	≤0.25 - ≤0.25	≤0.25	≤0.25
<b><i>Enterobacteriales</i> (n = 83)</b>					
<i>K. pneumoniae</i> (n = 57)		47%	0.06 - >256	2	>256
Challenge (n = 44)		67%	0.125 - >256	64	>256
ESBL-negative, carbapenem-susceptible (n = 13)		82%	0.25 - >256	>256	>256
<i>E. coli</i> (n = 19)		0%	0.125 - 1	0.125	1
Challenge (n = 11)	<b>Etest</b>	5%	0.06 - >256	0.125	1
ESBL-negative, carbapenem-susceptible (n = 8)		9%	0.125 - >256	0.25	2
<i>E. cloacae</i> (n = 7)		0%	0.06 - 0.125	0.125	0.125
Challenge (n = 3)		29%	0.25 - >256	0.5	>256
ESBL-negative, carbapenem-susceptible (n = 4)		67%	0.5 - >256	>256	>256
<i>E. cloacae</i> (n = 7)		0%	0.25 - 0.5	0.25	0.5
<b><i>P. aeruginosa</i> (n = 17)</b>					
Challenge (n = 11)	<b>BMD</b>	29%	0.5 - >32	0.5	>32
ESBL-negative, carbapenem-susceptible (n = 6)		45%	0.5 - >32	1	>32
		0%	0.5 - 0.5	0.5	0.5
<b><i>P. aeruginosa</i> (n = 17)</b>					
Challenge (n = 11)	<b>Vitek 2</b>	29%	0.5 - >16	0.5	>16
ESBL-negative, carbapenem-susceptible (n = 6)		45%	0.5 - >16	1	>16
		0%	0.5 - 0.5	0.5	0.5
<b><i>P. aeruginosa</i> (n = 17)</b>					
Challenge (n = 11)	<b>Etest</b>	29%	0.5 - >256	1	>256
ESBL-negative, carbapenem-susceptible (n = 6)		45%	0.5 - >256	1	>256
		0%	0.5 - 0.5	0.5	0.5

BMD: broth microdilution

The European Committee on Antimicrobial Susceptibility Testing (EUCAST). Breakpoint tables for interpretation of MICs and zone diameters, Version 12.0, 2022.

**Table S3.** Method comparison agreement, error categories and acceptance criteria as per the ISO standard 20776-2.

Agreement or error category	Acceptance criteria
Category Agreement (CA) Percentage of isolates (%) producing the same interpretive category result (susceptible or resistant) as compared to the reference broth microdilution (BMD) method	$\geq 90\%$ CA
Essential Agreement (EA) Percentage of isolates (%) producing MICs that are within $\pm$ one serial 2-fold dilution of the BMD method	$\geq 90\%$ EA
Major Error (ME) ME occurred when VITEK 2 AST-GN or Etest categorized a given isolate as resistant and BMD as susceptible (false resistance)	$\leq 3\%$ ME
Very Major Error (VME) VME occurred when VITEK 2 AST-GN or Etest categorized an isolate as susceptible and BMD as resistant (false susceptibility)	$\leq 3\%$ VME

ISO - ISO 20776-2:2021 - Clinical laboratory testing and *in vitro* diagnostic test systems - Susceptibility testing of infectious agents and evaluation of performance of antimicrobial susceptibility test devices - Part 2: Evaluation of performance of antimicrobial susceptibility test devices against reference broth micro-dilution.