

Table S1. Analytical information for serum calprotectin assay (GCAL®, Gentian AS) on an Abbott Architect instrument (The Hospital for Sick Children, Toronto, ON, Canada).

Characteristic	Information
Limit of Quantification (manufacturer defined)	0.30 mg/L
Limit of Detection (manufacturer defined)	0.15 mg/L
Analytical Measuring Range (manufacturer defined)	0.43-20.66 mg/L
Precision	Level 1 [mean (CV)]: 1.04 mg/L (7.5%) Level 2 [mean (CV)]: 10.1 mg/L (1.5%)
Linearity	Range Evaluated: 0–19 mg/L Slope: 0.97 Intercept: 0.12 mg/L R ² : 0.999
Method Comparison	Range Evaluated: 0–19 mg/L Slope: 0.9502 Intercept: -0.08 R ² : 0.997 Average Bias: -0.18 mg/L