

Figure S1




Figure S1: The oxygen reduction current from somatic cells in the milk cell chip filtered using nanofibers and without centrifugation. Approximately 0.27 g of nanofibers was inserted into a 10 mL disposable syringe and 5 mL milk (6.5×10^6 cells/mL) was filtered through a syringe. SECM-SCC was performed in the same manner as in section 2.5.