

Figure S1. Influences of (A) pH, (B) temperature, (C) aptamer concentration, (D) TMB concentration, (E) H₂O₂ concentration, (F) MoS₂-MIL-101(Fe) concentration for the MoS₂-MIL-101(Fe) hybrid nanozyme-based aptasensor.




Figure S2. Comparison of absorbances between measurement at first-day synthesis and measurement after 30-day storage at 4 °C with the same concentration of exosomes. Each column exhibited the average value of three independent data with an error bar presented.

Table S1. Detection of HGC-7901-derived exosomes in human serum.

Sample Number	Added (Particles/μL)	Found (Particles/μL)	Recovery (%)	RSD (%) (n = 3)
1	1.60×10^4	1.54×10^4	103	0.1
2	4.80×10^4	4.65×10^4	97	4.4
3	8.00×10^4	7.86×10^4	98	2.3
4	1.60×10^5	1.65×10^5	103	5.0
5	9.60×10^5	9.12×10^5	95	2.2