



Supplementary Figure S1 bioprinting evaluation of cell-containing GelMA: Bioprinting outcomes with insufficiently adjusted parameters. Left too low pressure and temperature, middle optimized parameters, right too high pressure. The phenomenon can be counteracted by adjusting the parameters in the scala below.