

Supplementary Information




Figure S1. The sequencing results of the positive control highlight a crucial distinction between the extraction protocols: the automated ones struggle to consistently detect Gram-positive pathogens with resilient cell walls, most likely due to the significantly reduced bead-beating intensity. On the y-axis: relative prevalence.