

Table S1. ICP-OES parameters for liquid reference measurements.

| | | |
|---------------------------------------|---|--|
| RF-Power | | 1300 W |
| Radial observation height | | 14 mm |
| Cooling gas flow (Ar) | | 12 L/min |
| Nebulizer gas flow (Ar) | | 0.65 L/min |
| Auxiliary gas flow (Ar) | | 0.4 L/min |
| Integration time per replicate | | 8 s |
| Replicates per sample | | 3 |
| Purge pump flow rate | | 1.6 mL/min |
| Analysis pump flow rate | | 0.8 mL/min |
| Measured Elements | Emission line used for quantification [nm] | Emission line used for quality control [nm] |
| B | 249.773 (I) | 208.893 (I) |
| Si | 250.690 (I) | 288.158(I) |
| Cr | 283.563(II) | 284.325(II) |

The liquid LOD was calculated by $\frac{3 \cdot \sigma}{k}$

Whereas k is the slope of the regression line, and σ is the standard derivation of the blank.




Figure S1. Particle size distribution measurement of the particles generated by online-LASIL