

You have not supplied any structure factors. As a result the full set of tests cannot be run.

No syntax errors found. CIF dictionary Interpreting this report

Bond precision:	C-C = 0.0075 A	Wavelength=0.71073	
Cell:	a=7.426 (6)	b=4.668 (4)	c=22.114 (19)
	alpha=90	beta=93.448 (11)	gamma=90
Temperature:	296 K		

```
Correction method= # Reported T Limits: Tmin=0.493 Tmax=0.746
AbsCorr = MULTI-SCAN
```

```
R(reflections)= 0.0788( 780)      wR2(reflections)=
S = 1.157                        0.1639( 1302)
Npar= 92
```

---

The following ALERTS were generated. Each ALERT has the format

**test-name\_ALERT\_alert-type\_alert-level.**

Click on the hyperlinks for more details of the test.

---



### Alert level C

PLAT029_ALERT_3_C	_diffn_measured_fraction_theta_full value Low	0.961	Why?
PLAT340_ALERT_3_C	Low Bond Precision on C-C Bonds .....	0.0075	Ang.

---



### Alert level G

PLAT007_ALERT_5_G	Number of Unrefined Donor-H Atoms .....	1	Report
	H003		
PLAT333_ALERT_2_G	Large Aver C6-Ring C-C Dist C004 -C006_a	1.42	Ang.
PLAT335_ALERT_2_G	Check Large C6 Ring C-C Range C004 -C006_a	0.16	Ang.
PLAT720_ALERT_4_G	Number of Unusual/Non-Standard Labels .....	20	Note
	C101 O002 N003 H003 C004 C005 C006 C007		
	H00A H00B C008 H00C H00D C009 H00E H00F		
	C00A H00G H00H H00I		
PLAT933_ALERT_2_G	Number of HKL-OMIT Records in Embedded .res File	4	Note
	-8 1 8, -8 1 11, -7 0 15, -7 0 17,		
PLAT941_ALERT_3_G	Average HKL Measurement Multiplicity .....	3.9	Low
PLAT967_ALERT_5_G	Note: Two-Theta Cutoff Value in Embedded .res ..	50.0	Degree

---

- 0 **ALERT level A** = Most likely a serious problem - resolve or explain  
0 **ALERT level B** = A potentially serious problem, consider carefully  
2 **ALERT level C** = Check. Ensure it is not caused by an omission or oversight  
7 **ALERT level G** = General information/check it is not something unexpected

- 0 ALERT type 1 CIF construction/syntax error, inconsistent or missing data  
3 ALERT type 2 Indicator that the structure model may be wrong or deficient  
3 ALERT type 3 Indicator that the structure quality may be low  
1 ALERT type 4 Improvement, methodology, query or suggestion  
2 ALERT type 5 Informative message, check
- 
-

It is advisable to attempt to resolve as many as possible of the alerts in all categories. Often the minor alerts point to easily fixed oversights, errors and omissions in your CIF or refinement strategy, so attention to these fine details can be worthwhile. In order to resolve some of the more serious problems it may be necessary to carry out additional measurements or structure refinements. However, the purpose of your study may justify the reported deviations and the more serious of these should normally be commented upon in the discussion or experimental section of a paper or in the "special\_details" fields of the CIF. checkCIF was carefully designed to identify outliers and unusual parameters, but every test has its limitations and alerts that are not important in a particular case may appear. Conversely, the absence of alerts does not guarantee there are no aspects of the results needing attention. It is up to the individual to critically assess their own results and, if necessary, seek expert advice.

### **Publication of your CIF in IUCr journals**

A basic structural check has been run on your CIF. These basic checks will be run on all CIFs submitted for publication in IUCr journals (*Acta Crystallographica*, *Journal of Applied Crystallography*, *Journal of Synchrotron Radiation*); however, if you intend to submit to *Acta Crystallographica Section C* or *E* or *IUCrData*, you should make sure that full publication checks are run on the final version of your CIF prior to submission.

### **Publication of your CIF in other journals**

Please refer to the *Notes for Authors* of the relevant journal for any special instructions relating to CIF submission.

