

Highlights:

1. Aptamer-based sensor was developed for rapid on-site detection of ofloxacin.
2. Limit of detection of sensor was 0.61 nM of ofloxacin.
3. A fluorescent sensor designed for ofloxacin;
4. Direct detection of ofloxacin without complex pre-treatment.