

Supplementary Table S1: Repeatability performances, estimated using 6 samples

Sample number	Experimental set-up	Mean value (kAU/L)	Repeatability precision (%)
1	Five repetition, randomly pipetted in one plate	1.61	5.5
2	Five repetition, randomly pipetted in one plate	3.77	3.9
3	Five repetition, randomly pipetted in one plate	5.96	5.99
4	Five repetition, randomly pipetted in one plate	7.42	23.6
5	Five repetition, randomly pipetted in one plate	10.1	15.90
6	Twenty-five repetitions in different plates (5 each plate)	2.01	20.9